

REQUEST FOR QUOTE - Rear Slide-Out Pan Bedside Commode (Rev B)

DATE

PROJECT

Retrofit module for a standard folding aluminium bedside commode that lets a caregiver empty the waste pan while the user remains seated. The round pan slides 20" (508 mm) straight out the BACK of the chair on two ball-bearing drawer slides, lifts out for emptying, and slides back to a latched position under the seat.

SCOPE OF QUOTE

Please quote the following, priced separately:

- A. Prototype: 1 complete working unit (module fitted to a base chair) for fit and function testing.
- B. Pilot run: 10 units.
- C. Production: 100 and 500 units (unit price, tooling, lead time).

For each, state whether you supply the base chair or fit the module to a chair we supply.

PARTS TO BE MADE / SOURCED (SEE SHEET 4 BOM FOR FULL LIST)

- SP-01 Carrier plate 375 x 400 x 6 mm, 320 mm round cutout, side flanges, rear D-handle (Sheet 2)
- SP-02 Round pan, 310 mm rim / 340 mm flange / 250 mm base, 152 mm deep, 7 qt working (Sheet 3)
- SP-03 Snap-on splash lid for pan
- SP-04 Tube clamp bracket, 1" saddle with foot, rubber-lined, qty 4 (Sheet 4)
- SP-05 Padded backrest 460 x 200 x 45 mm, mounts in FRONT of the rear bar
- SP-06 Angle rail 40 x 40 x 3 mm x 508 mm, 6061-T6, qty 2 (Sheet 4)
- Bought-in: 2x 20" full-extension slides (100 lb/pair, corrosion resistant), D-handle, magnetic catch, bump stops, stainless fasteners.

KEY REQUIREMENTS

- Retrofit: clamps to the existing frame, no welding or drilling of the chair.
- Pan and carrier concentric with the 440 mm round seat; pan must lift freely out of the carrier at full extension.
- Design load 300 lb (136 kg) seated user; 15 kg in pan at full extension; no tipping with pan extended.
- Pan, lid and carrier smooth, non-porous, dishwasher/disinfectant safe (PP or 304 SS).
- Slides and hardware stainless or zinc-plated with sealed bearings; washable.
- Tolerances per drawings (+/-1.0 mm untoleranced); all edges deburred.

SUPPLIER TO CONFIRM / ADVISE

1. Material recommendation for carrier (HDPE vs 304 SS) and pan (moulded PP vs drawn SS) at each volume.
2. Engineering review of slide/rail loading and tip stability with the pan extended; propose changes if needed.
3. Whether the design fits your standard commode frame, or dimensions that must change.
4. Lead time for prototype, pilot and production; tooling costs and ownership; MOQ.

US MARKET / REGULATORY REQUIREMENTS (PRODUCT SHIPS TO THE UNITED STATES)

- FDA: commode is a Class I medical device (21 CFR 890.3100/890.3110, 510(k)-exempt). Manufacturer must hold FDA Establishment Registration and Device Listing and follow 21 CFR 820 QMS (or ISO 13485). Provide evidence.
- Testing: static load and stability per ISO 17966 (assistive products for personal hygiene) and/or RESNA; supply test reports for 300 lb (136 kg) rated user load with pan extended.
- Materials: skin-contact parts biocompatible (ISO 10993-5/-10); pan/lid plastics food-grade PP. Vinyl, foam and plastics to meet California Prop 65 and CPSIA/16 CFR limits for lead, cadmium and phthalates; RoHS declaration.
- Corrosion: salt-spray / cleaning-agent resistance test on slides, rails, clamps and fasteners (e.g. ASTM B117).
- Labeling: permanent label with model, rated load, manufacturer, country of origin (19 CFR 134 'Made in China'), English instructions for use and cleaning; UDI/labeling per 21 CFR 801 as applicable.
- Quality: ISO 9001 certificate; pre-shipment inspection (AQL) report; packaging drop test (ISTA/ASTM D4169).
- Supplier to state the HTS code and confirm any US import tariff applicable to their quote.

FILES PROVIDED

README_manufacturer_brief.md (start here) | sheet1-4 PDF + DXF drawings and BOM
commode_slide_pan_closed / _extended .STEP + .STL (full assembly) | part_*.step (individual parts)
render_01-06.png (how it works) | model.py / drawings.py / render.py (parametric source)

NOTE

This is a concept design based on standard commode dimensions and has not been physically tested. Fit to the actual base chair and all structural/safety validation are to be performed by the supplier.

Please reply with your quotation and any design comments. Thank you.