
Evaluation protocol determines reported accuracy
(a) Predictive R² collapses under leakage-free evaluation
R² (coefficient of determination)
0.3
0.2
0.1
0.0
−0.1
−0.2
+0.248
-0.132
-0.086
-0.076
Random window split (as submitted)
Event-grouped
Subject-grouped (LOSO)
LOSO + event-level
(b) Error rises under leakage-free evaluation
MARD (%)
70
60
50
40
30
20
10
0
Clinical target (15%)
38.6
57.8
42.1
41.3
Random window split (as submitted)
Event-grouped
Subject-grouped (LOSO)
LOSO + event-level