
Evaluation protocol determines reported accuracy
(a) Predictive R² collapses under leakage-free evaluation
R² (coefficient of determination)
0.4
0.3
0.2
0.1
0.0
−0.1
−0.2
+0.392
+0.391
-0.097
-0.061
Random window split (as submitted)
Event-grouped
Subject-grouped (LOSO)
LOSO + event-level
(b) Error rises under leakage-free evaluation
Clinical target (15%)
MARD (%)
35
30
25
20
15
10
5
0
21.6
21.3
30.3
29.7
Random window split (as submitted)
Event-grouped
Subject-grouped (LOSO)
LOSO + event-level