MMS / EQUIPMENT BUYING GUIDE

Anesthesia machine buying guide

Choose around the interruption that stops your operating list: unreliable oxygen, unreliable electricity, missing ventilation capability or an unsustainable repair arrangement.

START HERE

Evaluate Gradian UAM where interrupted utilities shape the service, with its concentrator and ventilation maintenance funded. Compare Carestation 620 where dependable gases, power and GE support already exist.

PAY MORE WHEN

Replace a supportable Aestiva when the new package removes an actual capability or continuity gap. Integra SP VSO2 fits repeated travel only after the complete ventilation, monitoring and gas kit is specified.

WATCH THE BUDGET

UAM's published six-hour backup covers the ventilator and vital-sign monitor. Its concentrator needs mains power; room-air fallback does not preserve oxygen-enriched delivery. Fund backup oxygen separately.

CHECK THE SUPPORT

Require the exact power-loss behavior and applicable corrective-action record. Certain Carestation power boards have a documented battery-transfer correction; the model name alone cannot establish the offered unit's status.

THREE QUESTIONS FOR THE SUPPLIER

1. Which vaporizers, ventilation and monitoring functions are included?

2. What operates after mains loss, and for how long?

3. Who services the concentrator and the exact workstation at our hospital?

Selected sources: Gradian Health · Gradian Health · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.