MMS / EQUIPMENT BUYING GUIDE

DR retrofit or replacement buying sheet

Keep a useful X-ray room when a compatible detector package resolves the real limitation. Evaluate the whole acquisition and image-transfer workflow.

START HERE

Quote a retrofit after a qualified condition and support review of the generator, tube, table and stands. Include repairs before comparing prices.

PAY MORE WHEN

Replace the room when positioning, mechanical condition or unavailable parts create a material limitation that a new detector leaves unresolved.

WATCH THE BUDGET

Include the detector, workstation, licenses, batteries, networking, commissioning and training, plus future repairs to retained equipment.

CHECK THE SUPPORT

Assign one accountable route for integration faults and document service for both the original room and the new digital components.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact detector, controls and holder combination will you approve?

2. Can you demonstrate a complete examination and successful image transfer?

3. What does detector replacement cost, and who supports the retained generator?

Selected sources: Carestream · Carestream · Fujifilm

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.