MMS / EQUIPMENT BUYING GUIDE

Monitor, ECG and defibrillator buying guide

These are separate purchases: bedside monitoring, a diagnostic 12-lead ECG and resuscitation capability. Start with the job missing at each station.

START HERE

Use EDAN X8 for a basic ward-monitor comparison. Choose Midmark ECG where a supported computer already works, or ELI 230 for standalone acquisition and printing. Compare AED 3 and FRx for a dedicated AED station.

PAY MORE WHEN

Required invasive pressure moves the monitor choice to configured X10/X12 or B105M; X8 cannot supply that expansion. Hospital advanced resuscitation brings R Series Plus and a documented LIFEPAK 15 into consideration.

WATCH THE BUDGET

Price cuffs, probes, leads, pads and batteries before optional displays or networking. Midmark needs usable software rights; ELI needs paper and printer support. Include the FRx child key when required.

CHECK THE SUPPORT

Identify LIFEPAK 15 version, enabled functions and serial-specific inspection/correction records. Confirm recurring electrode and battery supply where the device will live.

THREE QUESTIONS FOR THE SUPPLIER

1. Which patient sizes and parameters does this monitor package cover?

2. Does the ECG price include supported software and retrievable records?

3. Which replacement pads, batteries and child accessories will remain available?

Selected sources: EDAN USA by MDPro · EDAN USA · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.