MMS / EQUIPMENT BUYING GUIDE

Equipment repair or replacement guide

Pay for the next useful step. A defined repair can preserve a complete service; another repair may be poor value when it leaves the main limitation untouched.

START HERE

Obtain a diagnosis, bounded repair scope and agreed acceptance test. Compare that with a complete replacement, including retained accessories, software, training and the cost of keeping work available during either route.

PAY MORE WHEN

Replace or reassign equipment when it cannot supply required functions or its essential parts cannot be supported. Repairing an X8 does not add invasive-pressure capability; a working basic monitor may still have a suitable role.

WATCH THE BUDGET

A repaired ultrasound console can remain dependent on an unreplaceable probe. An LF-GP still needs a complete processing route. Include repair freight and longer-turnaround scenarios without inventing failure probabilities.

CHECK THE SUPPORT

Use fault history and written repair acceptance. King Vision's component phase-out is a reason to verify remaining supplies before further investment; an accessible manual alone does not establish support.

THREE QUESTIONS FOR THE SUPPLIER

1. What specific fault will this repair resolve and how will you prove it?

2. Which essential accessory or part remains unsupported afterward?

3. What finding or additional cost would stop the repair and trigger replacement?

Selected sources: EDAN USA · CHISON · Olympus

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.