MMS / EQUIPMENT BUYING GUIDE

Ventilator buying guide

Separate bedside ventilation from transport, then specify patient groups, oxygen supply and the complete breathing-system package before comparing consoles.

START HERE

Without piped air, compare turbine-based HAMILTON-C1 and Savina 300 Classic. Savina's adult/pediatric scope may fit the ward; C1 deserves comparison when its compactness or purchased neonatal configuration matters.

PAY MORE WHEN

Choose T1 when transfers need its mounting and battery arrangement. Consider pneumatic pNeuton Model A only with a funded compressed-gas and monitoring plan. Avoid building a new fleet around AVEA's ended service program.

WATCH THE BUDGET

Include humidification, flow sensors, valves, circuits, batteries and oxygen. Savina's longer advertised autonomy includes an external battery; published runtime does not establish endurance of a used battery.

CHECK THE SUPPORT

Check Hamilton circuit item/lot numbers and applicable software and control-board corrections separately. New circuits alone do not settle the correction history of a used C1 or T1.

THREE QUESTIONS FOR THE SUPPLIER

1. Which patient groups, modes and monitoring options are enabled?

2. What oxygen input and complete battery arrangement does this package need?

3. Which serial, software and circuit corrective actions are documented as complete?

Selected sources: Hamilton Medical · Dräger · Hamilton Medical

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.