Airway visualization equipment buying guide

Specify video laryngoscopy and flexible-scope work separately. A compact video handle does not supply every function of a flexible intubation scope.

START HERE

Compare next-generation McGRATH MAC 301-000-000 where replacement batteries and blades are dependable, and GlideScope Go 2 where charging and battery service fit better. Include the necessary blade sizes.

PAY MORE WHEN

For intermittent flexible-scope work, compare aScope 4 Broncho/aView 2 Advance with maintaining an intact LF-GP in a functioning reusable-processing service. Buy the complete recurring supply arrangement with either.

WATCH THE BUDGET

Single-use scopes remove reused-scope processing but add freight, expiry and per-procedure costs. Reusable LF-GP economics depend on leak testing, processing, drying, storage and repair access.

CHECK THE SUPPORT

Check McGRATH serials, lawful provenance and applicable actions; next-generation labeling alone is insufficient. King Vision's documented component phase-out makes new fleet standardization dependent on a finite supply.

THREE QUESTIONS FOR THE SUPPLIER

1. Which blade and scope sizes can you replenish, with what expiry?

2. What battery or charging arrangement keeps the complete kit ready?

3. Can you document this handle's origin and completed applicable corrective actions?

Selected sources: Medtronic · Verathon · Ambu

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.