MMS / EQUIPMENT BUYING GUIDE

C-arm detector buying sheet

Compare complete supported C-arm configurations against the same imaging and room requirements. A detector label alone cannot decide the purchase.

START HERE

Request an image-intensifier baseline and a specified flat-panel alternative, identifying detector size, software and required image-export functions.

PAY MORE WHEN

A witnessed comparison demonstrates a useful imaging or positioning improvement, and the hospital can afford the resulting purchase and repair exposure.

WATCH THE BUDGET

Include detector or intensifier replacement, the X-ray tube, software, commissioning, training, freight and engineer travel in the comparison.

CHECK THE SUPPORT

Obtain written component coverage and replacement arrangements for the offered serial and destination. Keep acceptance and corrective-action responsibilities explicit.

THREE QUESTIONS FOR THE SUPPLIER

1. Which detector and licensed options are actually supplied?

2. How will you demonstrate the agreed images, room access and export workflow?

3. What are the complete installed costs of detector and tube replacement?

Selected sources: Siemens Healthineers · Siemens Healthineers · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.