MMS / EQUIPMENT BUYING GUIDE

Endoscopy system buying guide

Buy enough usable scopes to complete the list while one is being processed or repaired. A second tower cannot fix a missing-scope bottleneck.

START HERE

For an established Olympus service, compare CV-190/CLV-190 while retaining verified compatible scopes. For a new service, obtain complete EP-6000 and IMAGINA packages with the required gastroscopes and colonoscopes.

PAY MORE WHEN

Fund an additional accepted scope before a processor upgrade when scope availability limits work. Change platforms when the complete fleet, processing and repair arrangement makes the service more sustainable.

WATCH THE BUDGET

Include scope repairs, leak testing, drying, storage, image capture and processing supplies. PENTAX's i10c compatibility must be explicit; old EPK-1000 scopes cannot be assumed transferable.

CHECK THE SUPPORT

Match every scope to a validated reprocessing route. For SYSTEM 1E, check the exact model's compatibility and handling instructions; the P6700 distinguishes critical-device handling from semicritical-device storage.

THREE QUESTIONS FOR THE SUPPLIER

1. Which scope models and serials are included and accepted for repair?

2. How does our list run with one scope unavailable?

3. What processing equipment, chemistry and connections does each scope require?

Selected sources: Olympus · Olympus · Fujifilm

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.