Laboratory equipment buying guide

Build around the tests you can sustain during quiet weeks. Advertised throughput does not account for controls, repeats, expired reagent or preparation time.

START HERE

For modest chemistry demand, compare HumaLyzer 4000 with trained manual preparation. For three-part CBCs, match XQ-320 or pocH-100i to volume and reagent supply. Preserve funds for microscopy and suitable centrifugation.

CONSIDER ALTERNATIVES WHEN

BK200 automation earns consideration when preparation limits output and its water, temperature and service requirements are funded. Compare human BK-6310 when five-part capability is required; XP-300 is a three-part analyzer.

WATCH THE BUDGET

Price reagent packs, open stability, controls and washing at expected and lower volume. BK200 specifies 15-30°C operation and deionized water. Buy a centrifuge for the required tube, rotor geometry and force.

CHECK THE SUPPORT

Get local fluidics and reagent support. Inspect microscope image quality and focus stability. Replace a microtome only when sectioning limits an otherwise functioning histology pathway.

THREE QUESTIONS FOR THE SUPPLIER

1. What does each accepted result cost at half our expected volume?

2. Which reagent, control and water supplies are required?

3. Which rotor, adapters and relative centrifugal force fit our tube method?

Selected sources: HUMAN · BioBase · Sysmex

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.