MED MISSION SUPPLIES

Equipment Buying Compendium

Equipment buying guides

Choose a topic. Each page gives the essentials and links to its complete online guide.

Ultrasound buying guide 3

C-arm and X-ray buying guide 4

CT and MRI buying guide 5

Surgical table and light buying guide 6

Surgical microscope buying guide 7

Eye-care equipment buying guide 8

Ventilator buying guide 9

Monitor, ECG and defibrillator buying guide 10

Newborn and maternity equipment buying guide 11

Oxygen and respiratory support buying guide 12

Infusion and syringe pump buying guide 13

Suction equipment buying guide 14

Anesthesia machine buying guide 15

Endoscopy system buying guide 16

Electrosurgery buying guide 17

Airway visualization equipment buying guide 18

Laboratory equipment buying guide 19

Point-of-care testing buying guide 20

Sterilizer buying guide 21

Hospital bed and stretcher buying guide 22

Medical refrigerator buying guide 23

Dialysis equipment buying guide 24

Clinic essentials buying guide 25

MED MISSION SUPPLIES

Equipment Buying Compendium

Buying, setup and ownership

Choose a topic. Each page gives the essentials and links to its complete online guide.

Donated, refurbished or new equipment buying guide 26

Equipment quote comparison guide 27

Hospital equipment readiness guide 28

Equipment repair or replacement guide 29

One better machine or two simpler ones: buying guide 30

C-arm detector buying sheet 31

Used CT tube and package buying sheet 32

Ultrasound probe package buying sheet 33

Equipment service coverage buying sheet 34

MRI project readiness buying sheet 35

DR retrofit or replacement buying sheet 36

Image access: follow the whole study 37

Power support: specify the protected outcome 38

Replacement parts: close the whole transaction 39

Shipping: map every handover 40

End of support: make a dated decision 41

Rent or buy: price the full gap 42

Equipment handover buying sheet 43

Maintenance records buying sheet 44

Infusion consumables buying sheet 45

Sterile-processing project buying sheet 46

Facility oxygen buying sheet 47

Equipment software buying sheet 48

MMS / EQUIPMENT BUYING GUIDE

Ultrasound buying guide

Buy the examinations your hospital needs: each requires the right probe, enabled imaging modes and a usable way to save results.

START HERE

Use EBit 30 with a complete everyday probe kit as the first comparison. Keep a working M-Turbo if its useful probes and repair route still support the service.

PAY MORE WHEN

Compare EBit 60 for needed cardiac Doppler capability, SonoBook 6 for frequent room changes, or iQ3/Vscan Air CL when a handheld reaches otherwise inaccessible patients.

WATCH THE BUDGET

A console discount can disappear when probes, batteries, software and a compatible handheld display are added. A handheld subscription may change continuing costs and available functions.

CHECK THE SUPPORT

Get repair acceptance and replacement prices for the exact probes. Manufacturer images do not prove that the quoted software package includes the demonstrated modes.

THREE QUESTIONS FOR THE SUPPLIER

1. Which named probe and enabled mode covers each required examination?

2. What still works without internet or after a subscription expires?

3. Who replaces a failed probe, at what cost and turnaround?

Selected sources: CHISON · CHISON · CHISON

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

C-arm and X-ray buying guide

Choose the imaging job first: outreach radiography, repeated bedside imaging and surgical fluoroscopy need different equipment and support.

START HERE

Start with a complete MinXray Impact Wireless package for outreach. For full-size fluoroscopy, compare OEC One CFD with an inspected, supportable refurbished OEC 9800 or 9900.

PAY MORE WHEN

DRX-Revolution's powered movement can help frequent bedside work. Elite CFD earns consideration when its positioning or cooling addresses a real procedural limit. Demonstrate mini-C-arm anatomy coverage before choosing one.

WATCH THE BUDGET

Include the detector, acquisition computer, software, batteries, positioning equipment and commissioning. OEC One's standby function is for room changes; it does not establish off-grid X-ray operation.

CHECK THE SUPPORT

For a used C-arm, require tube, image-chain and motion-system assessment. A manufacturer parts category does not establish stock or repair coverage for the offered serial.

THREE QUESTIONS FOR THE SUPPLIER

1. Can this generator and detector cover our examination list?

2. Can the C-arm obtain every required view with our table?

3. What happens if commissioning fails or the detector needs replacement?

Selected sources: MinXray · Carestream · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

CT and MRI buying guide

A first scanner is a new service: fund accepted examinations, staffing, installation and continuing repairs together before committing to the machine.

START HERE

For routine CT, compare Revolution Advance and SOMATOM go.Now against the actual examination list. For MRI, evaluate Free.Star where siting and helium logistics constrain the project, alongside the 1.5T MR 5300.

PAY MORE WHEN

Pay for additional protocols, coils or capacity when radiologists accept the results and the complete examination workflow demonstrates a useful gain. A larger slice number alone cannot settle the CT choice.

WATCH THE BUDGET

A donated scanner still needs relocation, room work, installation, software rights and service. CT reconstructed slice counts differ from acquired coverage; low-helium MRI still has substantial operating obligations.

CHECK THE SUPPORT

Name the organization that will repair the exact relocated system. Price major-component exposure and service continuity before treating a low equipment price as a saving.

THREE QUESTIONS FOR THE SUPPLIER

1. Which protocols and coils are included in acceptance?

2. What is the complete installed price and annual fixed cost?

3. Who covers the tube or magnet and recovery after power interruption?

Selected sources: GE HealthCare · Siemens Healthineers · Siemens Healthineers

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Surgical table and light buying guide

Start with the patients, positions and imaging views your theatre uses. A table's largest load number does not describe every operating position.

START HERE

Compare Maquet Lyra as a general-theatre option with the complete supported table already on site. Include compatible armboards, leg supports, clamps and mattress in both packages.

PAY MORE WHEN

Skytron 3603 UltraSlide's longer slide can solve imaging access limits. CMAX X-Ray suits work that uses its floating top. For lights, assess Polaris and HarmonyAIR in the intended room.

WATCH THE BUDGET

Accessory replacement and engineered light installation can change the apparent bargain. Lyra's central and unrestricted load limits differ; CMAX X-Ray specifications do not describe every Cmax-family table.

CHECK THE SUPPORT

Demonstrate approved fallback controls, brakes and positioning. Check suspension drift and replacement light-driver availability. Polaris battery runtime applies to the documented mobile configuration, not every Polaris light.

THREE QUESTIONS FOR THE SUPPLIER

1. What load is allowed in our positions with these accessories?

2. Can our C-arm achieve its views without hitting the base?

3. Which controls and illumination remain available during a power or control failure?

Selected sources: Getinge · Skytron · STERIS

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Surgical microscope buying guide

Choose the working view, focus controls and mounting arrangement the surgeon can use, then price the complete transport or fixed-theatre package.

START HERE

Evaluate Scan Optics SO-111T for table-mounted travel work when manual 5x/10x adjustment is acceptable. Its AC/12 V input supports a portable comparison; the battery and approved mounting still need specification.

PAY MORE WHEN

M220 F12 adds motorized focus and magnification for a fixed ophthalmic workflow. OM-6 is a compact outpatient alternative whose cataract suitability requires review of the actual environment and configuration.

WATCH THE BUDGET

Include objective, stand or clamp, foot control, assistant view, case and power. M220's upgradable XY function is not necessarily installed. SO-111T's differing published weight descriptions need a final packing list.

CHECK THE SUPPORT

Have the surgeon accept the complete optics and controls. Verify service for a donated head/stand combination. Surgilens remains an unresolved production offer, not a confirmed available purchase.

THREE QUESTIONS FOR THE SUPPLIER

1. Can the surgeon demonstrate the intended work with these controls?

2. What is the complete packed weight and specified battery endurance?

3. Which stand, XY, assistant-view and foot-control options are included?

Selected sources: Scan Optics · Takagi Ophthalmic Instruments Europe · Leica Microsystems

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Eye-care equipment buying guide

Separate retinal imaging, slit-lamp examination and tonometry. Buy the missing task before paying for a newer camera or a more portable instrument.

START HERE

For a fixed slit-lamp station, compare SL-2G with an inspected BM 900. For seated/standing tonometry, evaluate IC100. For fixed retinal imaging, compare NW500 and DRSplus using representative, gradable images.

PAY MORE WHEN

IC200 adds supine and reclined positioning. Pictor Prestige can extend retinal capture to patients unable to use a fixed station. PSL Classic or SL-19 addresses bedside slit-lamp access.

WATCH THE BUDGET

Include slit-lamp table/head support, camera software and the actual tonometry supplies. TP01 probes are single-use and must not be cleaned for reuse. Pictor's anterior module does not automatically replace a slit lamp.

CHECK THE SUPPORT

Have intended operators demonstrate capture and export. Check used optics and mechanics, exact Kowa model code, compatible batteries and the repair route for the complete instrument.

THREE QUESTIONS FOR THE SUPPLIER

1. How many representative retinal images are gradable, including repeat attempts?

2. Which patient positions does this tonometer support?

3. Which probes, covers, table accessories and software rights are included?

Selected sources: Topcon Healthcare · Haag-Streit · iCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Ventilator buying guide

Separate bedside ventilation from transport, then specify patient groups, oxygen supply and the complete breathing-system package before comparing consoles.

START HERE

Without piped air, compare turbine-based HAMILTON-C1 and Savina 300 Classic. Savina's adult/pediatric scope may fit the ward; C1 deserves comparison when its compactness or purchased neonatal configuration matters.

PAY MORE WHEN

Choose T1 when transfers need its mounting and battery arrangement. Consider pneumatic pNeuton Model A only with a funded compressed-gas and monitoring plan. Avoid building a new fleet around AVEA's ended service program.

WATCH THE BUDGET

Include humidification, flow sensors, valves, circuits, batteries and oxygen. Savina's longer advertised autonomy includes an external battery; published runtime does not establish endurance of a used battery.

CHECK THE SUPPORT

Check Hamilton circuit item/lot numbers and applicable software and control-board corrections separately. New circuits alone do not settle the correction history of a used C1 or T1.

THREE QUESTIONS FOR THE SUPPLIER

1. Which patient groups, modes and monitoring options are enabled?

2. What oxygen input and complete battery arrangement does this package need?

3. Which serial, software and circuit corrective actions are documented as complete?

Selected sources: Hamilton Medical · Dräger · Hamilton Medical

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Monitor, ECG and defibrillator buying guide

These are separate purchases: bedside monitoring, a diagnostic 12-lead ECG and resuscitation capability. Start with the job missing at each station.

START HERE

Use EDAN X8 for a basic ward-monitor comparison. Choose Midmark ECG where a supported computer already works, or ELI 230 for standalone acquisition and printing. Compare AED 3 and FRx for a dedicated AED station.

PAY MORE WHEN

Required invasive pressure moves the monitor choice to configured X10/X12 or B105M; X8 cannot supply that expansion. Hospital advanced resuscitation brings R Series Plus and a documented LIFEPAK 15 into consideration.

WATCH THE BUDGET

Price cuffs, probes, leads, pads and batteries before optional displays or networking. Midmark needs usable software rights; ELI needs paper and printer support. Include the FRx child key when required.

CHECK THE SUPPORT

Identify LIFEPAK 15 version, enabled functions and serial-specific inspection/correction records. Confirm recurring electrode and battery supply where the device will live.

THREE QUESTIONS FOR THE SUPPLIER

1. Which patient sizes and parameters does this monitor package cover?

2. Does the ECG price include supported software and retrievable records?

3. Which replacement pads, batteries and child accessories will remain available?

Selected sources: EDAN USA by MDPro · EDAN USA · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Newborn and maternity equipment buying guide

Fund the missing care function: enclosed care, open warming, phototherapy or maternity monitoring. Several incomplete machines do not make one usable station.

START HERE

Compare Isolette 8000 plus and Giraffe Incubator Carestation for enclosed care. Wallaby WLB-1000 is a defined open-warmer starting point where 220-240 V fits. Firefly provides a standalone phototherapy arrangement.

CONSIDER ALTERNATIVES WHEN

Panda or Babyroo earns consideration when integrated functions replace equipment the unit needs. F6 Express adds maternal parameters; Avalon FM30 addresses internal-monitoring requirements beyond FM20's external-monitoring role.

WATCH THE BUDGET

Used incubators can need costly doors, seals, probes and humidity parts. Firefly must not be used inside an incubator or radiant warmer; choose the phototherapy layout before comparing prices.

CHECK THE SUPPORT

Verify phototherapy output with the appropriate measurement arrangement. Check supplied Panda blender and Babyroo hose identifiers against their component-specific actions; neither notice applies to every warmer function.

THREE QUESTIONS FOR THE SUPPLIER

1. Which enclosure, warming and resuscitation components complete this station?

2. Is our planned phototherapy arrangement permitted by the instructions?

3. Which fetal and maternal transducers, parameters and replacement parts are included?

Selected sources: Dräger · GE HealthCare · MTTS

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Oxygen and respiratory support buying guide

Size the oxygen source for simultaneous demand before buying downstream equipment. Oxygen generation, high-flow therapy, acute ventilation and sleep therapy are different jobs.

START HERE

Where 1-5 L/min meets the requirement, compare VisionAire 5 with the oxygen-monitor option. For requirements up to 10 L/min, compare the correct DeVilbiss 1025 and NewLife Intensity 10 electrical variants.

CONSIDER ALTERNATIVES WHEN

Price Airvo 2, Airvo 3 and HVT 2.0 as complete high-flow stations. For acute NIV, compare configured C1 or Savina. AirCurve ST-A, Astral and sleep CPAP require their own intended-use comparisons.

WATCH THE BUDGET

High-flow devices do not generate oxygen. Include source capacity, backup, chambers, tubing and interfaces. VisionAire's published 290 W and Intensity 10's 600 W make electricity a meaningful operating-cost input.

CHECK THE SUPPORT

Confirm approved source/device compatibility, concentration at required flow and repair access. Keep Airvo 2 and Airvo 3 correction records distinct; one does not resolve the other.

THREE QUESTIONS FOR THE SUPPLIER

1. What oxygen flow and pressure must our source deliver simultaneously?

2. Which voltage, oxygen monitoring and backup arrangements are included?

3. What is the annual complete consumable basket for this respiratory station?

Selected sources: Drive DeVilbiss · CAIRE · CAIRE

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Infusion and syringe pump buying guide

Choose a sustainable supply of approved tubing and syringes before choosing a pump family. A cheap console can create an expensive recurring dependency.

START HERE

Compare Infusomat/Perfusor compactplus where B. Braun consumables and service are dependable. Give Agilia VP MC WiFi/SP MC equal consideration where Fresenius replenishment and technical support are stronger.

PAY MORE WHEN

Fund drug libraries, profiles and networking when the hospital has people to approve, maintain and update them. Family standardization is useful when it reduces training and supply work; matching fronts alone does not justify replacement.

WATCH THE BUDGET

Compare the same annual approved-set basket, including delivered prices, minimum orders and expiry. Standard Infusomat compactplus does not accept arbitrary tubing; physical syringe fit does not establish validated operation.

CHECK THE SUPPORT

Require calibration/verification and battery support. For donated Spectrum or networked Alaris systems, identify exact serials or software versions and the applicable correction record.

THREE QUESTIONS FOR THE SUPPLIER

1. Which approved set codes and syringe brands/sizes are supplied locally?

2. What is their delivered annual cost at our expected use?

3. Who maintains the pump configuration, drug library and verification schedule?

Selected sources: B. Braun · B. Braun · B. Braun

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Suction equipment buying guide

Choose portable airway, bedside or surgical suction first. The collection system and power arrangement can decide suitability before maximum vacuum enters the comparison.

START HERE

Use battery-equipped Vacu-Aide 7325P-D as a compact portable benchmark. For mains surgical suction, start with Medela Basic when its 30 L/min free-airflow class meets the defined requirement.

PAY MORE WHEN

Compare VX-2 for a suitable transport mount and parts arrangement, or LSU for its battery access and canister/service options. Dominant Flex adds selectable 40/50/60 L/min for a defined higher-flow need.

WATCH THE BUDGET

7325D-D lacks the battery supplied in 7325P-D. VX-2's AC charger cannot run the pump through a failed battery. Medela's movable units remain mains powered. Include jars, lids, filters and tubing.

CHECK THE SUPPORT

Verify flow, vacuum and battery performance with the delivered collection system under the manufacturer's procedure. Choose canister parts the hospital can replenish or reprocess appropriately.

THREE QUESTIONS FOR THE SUPPLIER

1. Does this exact suffix include a battery and our required mount?

2. Can it operate from mains directly, or only charge from mains?

3. Which collection and overflow-protection parts are included and locally replaceable?

Selected sources: Drive DeVilbiss · Drive DeVilbiss · SSCOR

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Anesthesia machine buying guide

Choose around the interruption that stops your operating list: unreliable oxygen, unreliable electricity, missing ventilation capability or an unsustainable repair arrangement.

START HERE

Evaluate Gradian UAM where interrupted utilities shape the service, with its concentrator and ventilation maintenance funded. Compare Carestation 620 where dependable gases, power and GE support already exist.

PAY MORE WHEN

Replace a supportable Aestiva when the new package removes an actual capability or continuity gap. Integra SP VSO2 fits repeated travel only after the complete ventilation, monitoring and gas kit is specified.

WATCH THE BUDGET

UAM's published six-hour backup covers the ventilator and vital-sign monitor. Its concentrator needs mains power; room-air fallback does not preserve oxygen-enriched delivery. Fund backup oxygen separately.

CHECK THE SUPPORT

Require the exact power-loss behavior and applicable corrective-action record. Certain Carestation power boards have a documented battery-transfer correction; the model name alone cannot establish the offered unit's status.

THREE QUESTIONS FOR THE SUPPLIER

1. Which vaporizers, ventilation and monitoring functions are included?

2. What operates after mains loss, and for how long?

3. Who services the concentrator and the exact workstation at our hospital?

Selected sources: Gradian Health · Gradian Health · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Endoscopy system buying guide

Buy enough usable scopes to complete the list while one is being processed or repaired. A second tower cannot fix a missing-scope bottleneck.

START HERE

For an established Olympus service, compare CV-190/CLV-190 while retaining verified compatible scopes. For a new service, obtain complete EP-6000 and IMAGINA packages with the required gastroscopes and colonoscopes.

PAY MORE WHEN

Fund an additional accepted scope before a processor upgrade when scope availability limits work. Change platforms when the complete fleet, processing and repair arrangement makes the service more sustainable.

WATCH THE BUDGET

Include scope repairs, leak testing, drying, storage, image capture and processing supplies. PENTAX's i10c compatibility must be explicit; old EPK-1000 scopes cannot be assumed transferable.

CHECK THE SUPPORT

Match every scope to a validated reprocessing route. For SYSTEM 1E, check the exact model's compatibility and handling instructions; the P6700 distinguishes critical-device handling from semicritical-device storage.

THREE QUESTIONS FOR THE SUPPLIER

1. Which scope models and serials are included and accepted for repair?

2. How does our list run with one scope unavailable?

3. What processing equipment, chemistry and connections does each scope require?

Selected sources: Olympus · Olympus · Fujifilm

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Electrosurgery buying guide

Separate routine operating-room cut/coagulation, office procedures and advanced vessel sealing. Each has a different accessory basket and continuing supply bill.

START HERE

A refurbished Force FX is a credible routine-OR comparison when verified output, approved accessories and practical repairs come with it. Compare IDS-310 when a new platform offers a stronger complete service package.

PAY MORE WHEN

Choose Hyfrecator 2000 for work within its office-procedure scope. FT10 earns additional spending when surgeons need its compatible advanced-energy instruments and the hospital can replenish them consistently.

WATCH THE BUDGET

A generator price can omit pedals, pencils, forceps and return electrodes. Include smoke evacuation and supplies. Confirm voltage; the reviewed Hyfrecator seller package was specifically 115 V.

CHECK THE SUPPORT

Require serial-specific output and electrical-safety results for refurbishment. Match approved accessory part numbers, not just connectors. For FT10, establish software version and update eligibility.

THREE QUESTIONS FOR THE SUPPLIER

1. Which generator suffix, voltage, modes and pedals are supplied?

2. Are the quoted pencils and return electrodes approved for this configuration?

3. What is the recurring instrument cost for our intended advanced-energy cases?

Selected sources: Victori Medical · Aspen Surgical · CONMED

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Airway visualization equipment buying guide

Specify video laryngoscopy and flexible-scope work separately. A compact video handle does not supply every function of a flexible intubation scope.

START HERE

Compare next-generation McGRATH MAC 301-000-000 where replacement batteries and blades are dependable, and GlideScope Go 2 where charging and battery service fit better. Include the necessary blade sizes.

PAY MORE WHEN

For intermittent flexible-scope work, compare aScope 4 Broncho/aView 2 Advance with maintaining an intact LF-GP in a functioning reusable-processing service. Buy the complete recurring supply arrangement with either.

WATCH THE BUDGET

Single-use scopes remove reused-scope processing but add freight, expiry and per-procedure costs. Reusable LF-GP economics depend on leak testing, processing, drying, storage and repair access.

CHECK THE SUPPORT

Check McGRATH serials, lawful provenance and applicable actions; next-generation labeling alone is insufficient. King Vision's documented component phase-out makes new fleet standardization dependent on a finite supply.

THREE QUESTIONS FOR THE SUPPLIER

1. Which blade and scope sizes can you replenish, with what expiry?

2. What battery or charging arrangement keeps the complete kit ready?

3. Can you document this handle's origin and completed applicable corrective actions?

Selected sources: Medtronic · Verathon · Ambu

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Laboratory equipment buying guide

Build around the tests you can sustain during quiet weeks. Advertised throughput does not account for controls, repeats, expired reagent or preparation time.

START HERE

For modest chemistry demand, compare HumaLyzer 4000 with trained manual preparation. For three-part CBCs, match XQ-320 or pocH-100i to volume and reagent supply. Preserve funds for microscopy and suitable centrifugation.

CONSIDER ALTERNATIVES WHEN

BK200 automation earns consideration when preparation limits output and its water, temperature and service requirements are funded. Compare human BK-6310 when five-part capability is required; XP-300 is a three-part analyzer.

WATCH THE BUDGET

Price reagent packs, open stability, controls and washing at expected and lower volume. BK200 specifies 15-30°C operation and deionized water. Buy a centrifuge for the required tube, rotor geometry and force.

CHECK THE SUPPORT

Get local fluidics and reagent support. Inspect microscope image quality and focus stability. Replace a microtome only when sectioning limits an otherwise functioning histology pathway.

THREE QUESTIONS FOR THE SUPPLIER

1. What does each accepted result cost at half our expected volume?

2. Which reagent, control and water supplies are required?

3. Which rotor, adapters and relative centrifugal force fit our tube method?

Selected sources: HUMAN · BioBase · Sysmex

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Point-of-care testing buying guide

Buy a result the clinic needs soon enough to change its work. Test menus, temperature limits and usable pack quantities decide the recurring cost.

START HERE

For a small clinic, first fund sustainable hemoglobin, professional glucose and urine workflows. Hb 301 deserves a warm-site comparison because its stated operating and cuvette-storage range is 10-40°C.

CONSIDER ALTERNATIVES WHEN

For urgent blood gases/basic chemistry, compare epoc with the required i-STAT cartridges. Piccolo fits routine chemistry with refrigerated discs. Add DCA Vantage when sustained demand justifies its assay and control program.

WATCH THE BUDGET

epoc cards require 15-30°C storage; Piccolo discs require 2-8°C. Include delivery conditions, expiry, controls and repeats. A low pack price means little if quiet weeks leave much of it unused.

CHECK THE SUPPORT

Verify exact reader, reagent and software compatibility. The reviewed October 2026 Nova action concerned specified Gen 2.0 models and an unresolved remedy, not every StatStrip device.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact assays, sample types and patients does this package cover?

2. What pack sizes, arrival expiry and control consumption should we budget?

3. Can our delivery and storage conditions meet every consumable's limits?

Selected sources: HemoCue · Siemens Healthineers · Siemens Healthineers

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Sterilizer buying guide

Buy for complete instrument sets ready at the next procedure. Cleaning, drying, cooling and release can constrain output even when chamber capacity looks adequate.

START HERE

Compare Tuttnauer 2540M for a modest workload whose instruments and packaging fit its validated cycles, with staff able to operate and document the process consistently.

PAY MORE WHEN

M11's automatic records and care reminders can justify more spending when they remove staff work. Where instruments require suitable vacuum processing, quote T-Top with Class B actually supplied and enabled.

WATCH THE BUDGET

Add suitable water, packaging, monitoring supplies and enough instrument sets. The reviewed M11-050/-051 offer is medical-only, excluding dental use. The refurbished 2540M listing had conflicting warranty lengths needing resolution.

CHECK THE SUPPORT

Price seals, heaters, valves and control parts through the technician who will maintain the unit. Manual controls alone do not establish easier repairs or longer life.

THREE QUESTIONS FOR THE SUPPLIER

1. Which quoted cycles accept our exact instruments and packaging?

2. How many complete sets can we release on our busiest day?

3. What voltage, refurbishment scope and written warranty apply to this unit?

Selected sources: Tuttnauer · Midmark · Tuttnauer

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Hospital bed and stretcher buying guide

Spend first on safe positioning and transfers: the mattress, rails, brakes, castors and route through the building determine daily usefulness.

START HERE

Compare crank-adjusted Praktika 2 where mechanical positioning suits patients and staffing. Its maximum patient weight is 165 kg; its 200 kg safe working load is a different limit.

PAY MORE WHEN

Eleganza 1's electric positioning can earn its cost with frequent repositioning. A refurbished Advanta 2 fits more naturally into an already supported fleet. For stretchers, compare complete ST1 and P8005 packages.

WATCH THE BUDGET

Include the compatible mattress and freight before judging a frame price. P8005's Steering Plus is optional. An Advanta 2 seller's ICU description does not change the manufacturer's med-surg scope.

CHECK THE SUPPORT

Test loaded maneuverability along actual turns, thresholds and lift entrances. Check rail latches and brakes, then obtain actuator, handset and battery support for powered beds.

THREE QUESTIONS FOR THE SUPPLIER

1. Does this bed, mattress and rail combination meet the intended patient requirements?

2. Which steering and equipment mounts are included on the stretcher?

3. Can our technician obtain replacement brake, castor and positioning parts?

Selected sources: LINET · LINET · Baxter

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Medical refrigerator buying guide

Name the stock, temperature requirements and peak inventory first. Vaccine, blood and pharmacy cabinets solve different storage jobs and need a practical alarm response.

START HERE

For off-grid vaccines fitting 88 L net capacity, compare Dulas VC88SDD with an appropriately qualified Haier HTC-120 package. For blood, compare B51's 45 L with B131's 106 L against actual inventory.

PAY MORE WHEN

Buy extra blood capacity when peak stock needs it. For pharmacy or reagents, compare P55's 45 L with PHCbi MPR-S150H-PE's 165 L and the required validated storage arrangement.

WATCH THE BUDGET

A larger B131 retains B51's +32°C upper ambient limit. Solar pricing must include site design and installation. UNICEF tender benchmarks have specific currencies, delivery terms and exclusions; they are not hospital landed prices.

CHECK THE SUPPORT

Specify monitoring, a named alarm responder and stock-transfer arrangements. B51's published holdover to +10°C is a test endpoint, not permission to store blood at that temperature.

THREE QUESTIONS FOR THE SUPPLIER

1. What usable capacity remains for our products and airflow arrangement?

2. Will this cabinet work within our measured room conditions?

3. Which solar, monitoring, installation and outage provisions are included?

Selected sources: Dulas · UNICEF Supply Division · B Medical Systems

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Dialysis equipment buying guide

Buy a sustainable treatment service before counting stations. Water quality, trained staff, consumables and repair cover determine how much of the installed fleet can work.

START HERE

Use a locally supported Fresenius 4008S classix as the conventional-HD capability benchmark. Compare complete supply and service offers; the focused platform is not automatically the lowest-cost installation.

PAY MORE WHEN

Consider Dialog iQ for needed biofeedback, HDF or single-needle capability. Surdial X is an HD/HDF alternative where its configured package and Nipro support fit better. Fund training and required supplies alongside added functions.

WATCH THE BUDGET

Include water treatment and distribution, chemical and microbiological monitoring, approved bloodlines, concentrates, filters and disinfectants. Price Dialog iQ's dedicated DiaStream iQ lines explicitly at the same treatment volume as competing offers.

CHECK THE SUPPORT

Obtain service response, parts and consumable replenishment commitments at the hospital. A current regional product page does not establish support for an old donated serial.

THREE QUESTIONS FOR THE SUPPLIER

1. What complete per-treatment basket does this modality require?

2. Who specifies and maintains the water system and quality monitoring?

3. How will treatments continue when one station or a supply shipment is unavailable?

Selected sources: Fresenius Medical Care · B. Braun · B. Braun

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Clinic essentials buying guide

Complete the examinations performed every day before adding equipment that creates a new service. Patient sizes and replacement supplies matter more than a large bundled kit.

START HERE

Start with a dependable acoustic stethoscope, purpose-built ear/eye optics, suitable BP cuffs, thermometer and patient-appropriate oximeter. Keep a working Cardiology III; compare HEINE mini 3000 or compatible Welch Allyn additions for optics.

CONSIDER ALTERNATIVES WHEN

Add digital sound, spirometry, dermoscopy or audiometry for a defined task with trained users and follow-up. Spirobank II Basic supports standalone work; Easy on-PC needs a maintained computer.

WATCH THE BUDGET

Budget tips, cuffs, probe covers, breathing interfaces and calibration. HBP-1320 is for age three and above; Nonin 9590 does not cover neonatal or prolonged unattended monitoring. CORE 500's three-lead ECG does not replace diagnostic 12-lead ECG.

CHECK THE SUPPORT

Match reusable instruments to an established processing route and inspect disposable-kit contents separately. Verify software entitlement and actual offline workflow for connected additions.

THREE QUESTIONS FOR THE SUPPLIER

1. Which patient sizes do the included cuffs, probes and tips cover?

2. Which compatible handles and supplies can we keep using?

3. What training, consumables and recordkeeping does the added service require?

Selected sources: Littmann / Solventum · HEINE · Baxter / Hillrom

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Donated, refurbished or new equipment buying guide

Compare three routes to the same missing service: complete a donation, buy a documented refurbishment or fund a new supported installation.

START HERE

Begin with the receiving team's need and inventory. Accept a donation when the exact configuration fills that gap and its completion costs and upkeep have owners and funding.

PAY MORE WHEN

Pay for refurbishment when the defined inspection and repair work produces useful savings. Pay for new equipment when required functions, installation or support resolve gaps left by cheaper offers.

WATCH THE BUDGET

A free endoscopy tower still needs compatible scopes, accessories and processing. Add packing, freight, imports, software rights, missing components and acceptance. Keep unknown completion costs visible instead of treating them as zero.

CHECK THE SUPPORT

Ask where a warranty repair happens, who pays each freight leg and how the service continues. A warranty term or refurbished label does not establish the unit's remaining life.

THREE QUESTIONS FOR THE SUPPLIER

1. What exact work and accessories turn this offer into our required service?

2. What inspection and repair evidence follows the offered serial?

3. Who pays return freight, missing components and continuing replenishment?

Selected sources: WHO · Olympus · STERIS

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Equipment quote comparison guide

Make competing offers cover the same work before comparing totals. Model-family names, upgrade promises and a low console price can conceal different purchases.

START HERE

Write one required configuration and have each supplier price it: patient sizes, functions, accessories, software, delivery, installation, acceptance and initial supplies. Mark every omission as included, priced separately or unresolved.

PAY MORE WHEN

Accept a higher equipment price when it lowers the complete cost or supplies a needed capability. X8 lacks the invasive-pressure option of X10/X12; T-Top Class B must actually be supplied.

WATCH THE BUDGET

Extend the comparison through a common operating period. Add consumable pack sizes, controls, usable expiry and scheduled service. Test quieter activity: unused stock can reverse an apparent per-item saving.

CHECK THE SUPPORT

Attach usable software rights and named repair coverage to each offer. Ultrasound needs named probes and enabled applications; a PC-based ECG needs a supported computer/software arrangement.

THREE QUESTIONS FOR THE SUPPLIER

1. Which required function or accessory is excluded from this total?

2. What changes in the recurring bill at half our expected workload?

3. Which unpriced delivery, installation or support item could reverse the comparison?

Selected sources: EDAN USA · Tuttnauer · MFI Medical

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Hospital equipment readiness guide

Follow the proposed service through an ordinary day and an interruption. Give every essential utility, supply and maintenance dependency a funded owner before shipment.

START HERE

Use the exact installation documents to match equipment with power, oxygen, water, room conditions and storage. Fund the missing enabling infrastructure before adding capacity that cannot operate reliably at the site.

PAY MORE WHEN

Change equipment architecture when it solves a documented constraint: solar-direct-drive vaccine refrigeration can address building-power dependence, while its site, monitoring and installation requirements still need funding.

WATCH THE BUDGET

UAM's ventilator/monitor battery does not power its oxygen concentrator. Airvo is not an oxygen generator. BK200 needs deionized water; dialysis has a separate water-quality system. Include each dependency in the installed cost.

CHECK THE SUPPORT

Carry storage limits through delivery and handover: epoc cards need 15-30°C and Piccolo discs 2-8°C. B51's +32°C upper ambient limit also applies after the installer leaves.

THREE QUESTIONS FOR THE SUPPLIER

1. Which complete utility requirements must our technical team approve?

2. What stops working during our expected power or supply interruption?

3. Who owns replenishment, environmental monitoring and repair after handover?

Selected sources: Gradian Health · Gradian Health · World Health Organization

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Equipment repair or replacement guide

Pay for the next useful step. A defined repair can preserve a complete service; another repair may be poor value when it leaves the main limitation untouched.

START HERE

Obtain a diagnosis, bounded repair scope and agreed acceptance test. Compare that with a complete replacement, including retained accessories, software, training and the cost of keeping work available during either route.

PAY MORE WHEN

Replace or reassign equipment when it cannot supply required functions or its essential parts cannot be supported. Repairing an X8 does not add invasive-pressure capability; a working basic monitor may still have a suitable role.

WATCH THE BUDGET

A repaired ultrasound console can remain dependent on an unreplaceable probe. An LF-GP still needs a complete processing route. Include repair freight and longer-turnaround scenarios without inventing failure probabilities.

CHECK THE SUPPORT

Use fault history and written repair acceptance. King Vision's component phase-out is a reason to verify remaining supplies before further investment; an accessible manual alone does not establish support.

THREE QUESTIONS FOR THE SUPPLIER

1. What specific fault will this repair resolve and how will you prove it?

2. Which essential accessory or part remains unsupported afterward?

3. What finding or additional cost would stop the repair and trigger replacement?

Selected sources: EDAN USA · CHISON · Olympus

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

One better machine or two simpler ones: buying guide

Name the constraint before counting machines: required capability, daily capacity and continuity during repairs are different reasons to spend.

START HERE

Fund the smallest complete change that removes the real bottleneck. That may be another compatible endoscope, an ultrasound probe, more instrument sets or a repair-loaner arrangement before another console.

PAY MORE WHEN

Buy higher capability when the work needs it. A second X8 cannot back up required invasive pressure; another incompatible sterilizer cycle cannot replace the required T-Top load capability.

WATCH THE BUDGET

Two units each need accessories, supplies, maintenance and trained users. Check cleaning, drying, cooling and release before buying a second sterilizer. M11 automation may help staff workload where manual 2540M operation consumes attention.

CHECK THE SUPPORT

Test the backup against shared interruptions. Two oxygen-dependent machines can both stop when their source fails. A handheld ultrasound is useful backup only for examinations and export workflows it can actually cover.

THREE QUESTIONS FOR THE SUPPLIER

1. Which additional work becomes possible with this second complete package?

2. What shared power, oxygen, software or supply failure stops both units?

3. Would a scope, probe, processing improvement or loaner remove more downtime?

Selected sources: Tuttnauer · EDAN USA · Tuttnauer

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

C-arm detector buying sheet

Compare complete supported C-arm configurations against the same imaging and room requirements. A detector label alone cannot decide the purchase.

START HERE

Request an image-intensifier baseline and a specified flat-panel alternative, identifying detector size, software and required image-export functions.

PAY MORE WHEN

A witnessed comparison demonstrates a useful imaging or positioning improvement, and the hospital can afford the resulting purchase and repair exposure.

WATCH THE BUDGET

Include detector or intensifier replacement, the X-ray tube, software, commissioning, training, freight and engineer travel in the comparison.

CHECK THE SUPPORT

Obtain written component coverage and replacement arrangements for the offered serial and destination. Keep acceptance and corrective-action responsibilities explicit.

THREE QUESTIONS FOR THE SUPPLIER

1. Which detector and licensed options are actually supplied?

2. How will you demonstrate the agreed images, room access and export workflow?

3. What are the complete installed costs of detector and tube replacement?

Selected sources: Siemens Healthineers · Siemens Healthineers · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Used CT tube and package buying sheet

A scanner's age and its installed tube's history are different facts. Compare documented condition and a usable replacement remedy.

START HERE

Shortlist accepted systems with traceable tube identities, dated service records and a priced route for replacement at the destination.

PAY MORE WHEN

An identified replacement or broader refurbishment removes an exposure the hospital cannot otherwise fund. Specify the supplied component's condition and warranty.

WATCH THE BUDGET

Include the tube, installation, calibration, freight, site work, utilities, service travel and temporary service arrangements before comparing scanner prices.

CHECK THE SUPPORT

Match serial numbers, condition categories and purchased terms. Establish who pays when commissioning or an early failure requires corrective work.

THREE QUESTIONS FOR THE SUPPLIER

1. Which tube is installed, and what records substantiate its usage history?

2. What is the complete installed replacement price and applicable warranty?

3. Who corrects failed acceptance and covers diagnosis, travel and early faults?

Selected sources: GE HealthCare · GE HealthCare · Siemens Healthineers

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Ultrasound probe package buying sheet

Price the probes that make the required examinations possible, together with their care and a workable repair route.

START HERE

Start with the smallest complete compatible probe bundle. Compare restoring a supported installed console before buying another platform.

PAY MORE WHEN

Required examinations, software support or complete package economics justify changing systems. Demonstrate the exact probes and enabled modes in the offer.

WATCH THE BUDGET

Allow for probe condition, processing supplies, storage, diagnosis, exchange charges, freight, loaners and a compatible backup where justified.

CHECK THE SUPPORT

Name each covered probe and verify the proposed loaner's compatibility, availability, delivery and charges. Obtain test evidence after repair or replacement.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact probes and modes are included and demonstrated?

2. What is the complete repair, exchange or replacement cost?

3. Which processing supplies and backup arrangements must the hospital fund?

Selected sources: GE HealthCare · Philips · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Equipment service coverage buying sheet

Compare who performs and pays for each maintenance and repair obligation. Plan names and annual fees do not define coverage.

START HERE

Request equal-scope full-service, maintenance-only and separately funded repair alternatives. Start with the failures the hospital cannot afford to carry.

PAY MORE WHEN

Broader coverage resolves a material expertise, access or financial gap and supplies a credible response at this destination.

WATCH THE BUDGET

Price major components, overtime, travel, freight, deductibles, limits and exclusions, along with internal staff time, training and diagnostic tools.

CHECK THE SUPPORT

Define acknowledgement, diagnosis, arrival and restoration separately. Confirm parts access, interim arrangements and ownership of the maintenance history.

THREE QUESTIONS FOR THE SUPPLIER

1. Which major components and causes of failure remain excluded?

2. What starts and stops each response clock?

3. Who pays for diagnosis, travel, installation and post-repair verification?

Selected sources: WHO · WHO · Philips

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

MRI project readiness buying sheet

Make the equipment commitment after exact-system feasibility and complete project funding are established. A scanner quote is only one part of the service.

START HERE

Choose a clinically suitable configuration with a specialist-reviewed site, priced installation responsibilities and sustainable operating support.

PAY MORE WHEN

A different design resolves a documented building or operating constraint at an acceptable complete cost. Verify the remaining dependencies separately.

WATCH THE BUDGET

Include design, shielding, utilities, rigging, delivery, coils, software, supporting equipment, commissioning, training and ongoing specialist service.

CHECK THE SUPPORT

Confirm the exact scanner, coils, software and cooling equipment in the service scope, including major exclusions and interruption arrangements.

THREE QUESTIONS FOR THE SUPPLIER

1. Which planning revisions apply to this exact scanner and magnet?

2. Which site findings must be resolved before delivery?

3. Who pays for and verifies every construction, utility and installation interface?

Selected sources: GE HealthCare · GE HealthCare · Philips

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

DR retrofit or replacement buying sheet

Keep a useful X-ray room when a compatible detector package resolves the real limitation. Evaluate the whole acquisition and image-transfer workflow.

START HERE

Quote a retrofit after a qualified condition and support review of the generator, tube, table and stands. Include repairs before comparing prices.

PAY MORE WHEN

Replace the room when positioning, mechanical condition or unavailable parts create a material limitation that a new detector leaves unresolved.

WATCH THE BUDGET

Include the detector, workstation, licenses, batteries, networking, commissioning and training, plus future repairs to retained equipment.

CHECK THE SUPPORT

Assign one accountable route for integration faults and document service for both the original room and the new digital components.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact detector, controls and holder combination will you approve?

2. Can you demonstrate a complete examination and successful image transfer?

3. What does detector replacement cost, and who supports the retained generator?

Selected sources: Carestream · Carestream · Fujifilm

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Image access: follow the whole study

A successful send does not establish a complete archive service. Compare the journey from registration to reading, retention, recovery and eventual export.

START HERE

Test the existing supported workflow first. Map each step to its responsible system, person and acceptance evidence before buying additional software or storage.

PAY MORE WHEN

A demonstrated access, capacity, integration or recovery gap requires a different local, hosted or hybrid arrangement that the team can maintain.

WATCH THE BUDGET

Include interfaces, viewers, storage growth, connectivity, support, backup verification, infrastructure replacement and a future bulk export. Compare identical workload assumptions.

CHECK THE SUPPORT

Match conformance statements to actual versions. Demonstrate current and prior study access during the agreed outage, then verify restoration and reconciliation.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact versions and licensed services will you test together?

2. What remains accessible during an internet or local-network interruption?

3. How do we receive a complete usable export when the agreement ends?

Selected sources: DICOM Standards Committee · DICOM Standards Committee · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Power support: specify the protected outcome

Buy a documented response to the site's actual power events. A UPS label alone does not identify what stays operational or for how long.

START HERE

Commission a qualified assessment, obtain exact equipment requirements and list every load needed for the proposed bridge, continued operation or controlled shutdown.

PAY MORE WHEN

The accepted design identifies an interruption or power-quality problem that the installed arrangement cannot adequately address for the required equipment.

WATCH THE BUDGET

Include assessment, electrical design, site work, commissioning, batteries, monitoring, routine service and the operating needs of any backup supply.

CHECK THE SUPPORT

Assign responsibility for equipment compatibility, generator handover, recovery, alarms and maintenance. Witness agreed tests under professionally controlled conditions before acceptance.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact loads and conditions support your runtime claim?

2. Who verifies transfer to backup supply and return to normal?

3. What maintenance and replacement funding preserves the promised capability?

Selected sources: GE HealthCare · GE HealthCare · WHO, World Bank, IRENA and SEforALL

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Replacement parts: close the whole transaction

A component price is only one part of a repair. Verify the diagnosis, exact compatibility, condition and all return obligations before ordering.

START HERE

Assemble the model, serial, software, options, fault report and full part label. Obtain written compatibility confirmation and identify who will install and verify it.

PAY MORE WHEN

Stronger testing evidence, a usable warranty or dependable delivery resolves a material repair risk that the cheaper offer leaves with the hospital.

WATCH THE BUDGET

Compare diagnosis, freight, handling, installation, calibration, temporary cover and core liability. Show the maximum additional charge if an exchange condition is unmet.

CHECK THE SUPPORT

Record the accepted repair result against the asset. Assign the core return, track its delivery and obtain supplier acceptance before closing the purchase.

THREE QUESTIONS FOR THE SUPPLIER

1. Is this exact part or approved replacement compatible with our configuration?

2. What condition, tests and return remedy are included?

3. What deadline, eligibility rule and charge apply to the old core?

Selected sources: GE HealthCare · US Food and Drug Administration · WHO

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Shipping: map every handover

A freight quote can stop well before equipment reaches an accepted installation. Plan the whole route and assign every responsibility before dispatch.

START HERE

Build an origin-to-room worksheet naming the equipment engineer, packer, carrier, broker, receiving team and installer, with evidence and handover points for each.

PAY MORE WHEN

Specialist packing, handling or coordination resolves a material gap in the route, equipment requirements or the hospital's ability to manage contractors.

WATCH THE BUDGET

Include de-installation, packing, clearance, unloading, rigging, site work, storage, waiting, re-delivery, installation and any uncovered loss exposure.

CHECK THE SUPPORT

Confirm destination-specific requirements with qualified advisers. Inspect and document receipt, preserve claim evidence, and keep delivery separate from installation and technical acceptance.

THREE QUESTIONS FOR THE SUPPLIER

1. Where does your scope end, and who takes responsibility next?

2. Who confirms the destination and documents are ready before dispatch?

3. Who carries transport risk and manages any damage claim?

Selected sources: GE HealthCare · FedEx · FedEx

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

End of support: make a dated decision

A support notice starts a structured review. Determine what changes for the exact installed configuration before deciding to continue, upgrade or replace.

START HERE

Build a support map for parts, field service, software, security and essential accessories. Obtain written dates and limitations from the responsible suppliers.

PAY MORE WHEN

A supported upgrade or replacement resolves a material gap that the hospital cannot acceptably manage through a documented interim plan.

WATCH THE BUDGET

Compare support, controls, spares and interruption provision with complete upgrade or replacement costs, including site work, training, migration and retirement.

CHECK THE SUPPORT

Involve clinical, biomedical and security teams. A continuation decision needs an owner, review date, stop conditions and a funded transition plan.

THREE QUESTIONS FOR THE SUPPLIER

1. Which support activities end for our serial and software configuration?

2. What evidence supports the proposed parts, service and security arrangement?

3. What remains supported after an upgrade, and for what period?

Selected sources: International Medical Device Regulators Forum · Siemens Healthineers · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Rent or buy: price the full gap

Temporary equipment should fill a defined capability gap for a defined period. Compare usable packages and complete costs rather than daily rates alone.

START HERE

State the required configuration, usable start date and event ending the need. Compare rental, confirmed loaner, approved internal capacity and purchase.

PAY MORE WHEN

A better-supported or more complete package can meet the actual deadline and clinical requirement that a cheaper offer cannot demonstrate.

WATCH THE BUDGET

Price the planned interval and plausible extensions. Include transport, setup, accessories, training, minimum commitments, return charges and the exact billing stop event.

CHECK THE SUPPORT

Confirm availability, maintenance, replacement after failure and return obligations. Assign an owner to approve extensions and obtain evidence that billing has ended.

THREE QUESTIONS FOR THE SUPPLIER

1. Which exact package is available for our dates and destination?

2. What starts and stops billing, and which extra charges apply?

3. What would the complete cost be if the need lasts longer?

Selected sources: Agiliti · GE HealthCare · WHO

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Equipment handover buying sheet

Agree the journey from receipt to authorized use before equipment ships. One record should connect the delivered serial, promised configuration and the people responsible for acceptance.

START HERE

Attach a staged handover schedule to the order: receipt, installation, technical acceptance, user preparation and authorization for use. Give every stage an owner and required evidence.

PAY MORE WHEN

Pay for specialist installation, acceptance support or additional training when the receiving hospital lacks the necessary qualified capacity.

WATCH THE BUDGET

Include site work, travel, repeat visits, missing accessories, storage and interim service while any delivery issues are corrected.

CHECK THE SUPPORT

Record warranty start conditions, claim contacts, manuals, training attendance and the first maintenance action. Keep unresolved items open until their evidence is reviewed.

THREE QUESTIONS FOR THE SUPPLIER

1. Who will deliver each installation and acceptance report?

2. Which event starts the warranty?

3. Who pays for corrective work and repeat visits?

Selected sources: WHO · UNICEF Supply Division · UNICEF Supply Division

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Maintenance records buying sheet

Use the equipment register to decide what needs repair, support, accessories or replacement. Begin with verified assets and records that a named person keeps current.

START HERE

Link each local asset identifier to its model, serial, location, service reports, approved maintenance basis, next task and current status.

PAY MORE WHEN

Buy a maintenance management system when several users, scheduled work or a larger fleet exceed the current record's practical limits.

WATCH THE BUDGET

Allow for physical inventory checks, data entry, migration, staff training and ongoing administration. Keep unquoted repair costs visibly unknown.

CHECK THE SUPPORT

Require external engineers to return device-level reports. Test backup, retrieval and export so the hospital retains its history after staff or supplier changes.

THREE QUESTIONS FOR THE SUPPLIER

1. Which tasks and records are included in your service?

2. Can we export the complete history for each serial?

3. Who keeps the register current and resolves overdue actions?

Selected sources: WHO · WHO · WHO Laboratory Quality Stepwise Implementation tool

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Infusion consumables buying sheet

Compare pumps with a full year of the approved sets and syringes the hospital needs. A low equipment quote can leave a difficult recurring supply commitment.

START HERE

Build one basket listing exact pump configuration, approved product references, annual quantities, pack sizes and delivered prices. Have the hospital's responsible team approve it.

PAY MORE WHEN

A change resolves a demonstrated capability, supply or support problem after training, accessories and transition stock are included.

WATCH THE BUDGET

Check minimum orders, expiry on delivery, freight, price validity and unused old stock. Separate opening inventory from annual consumption.

CHECK THE SUPPORT

Require current compatibility evidence and a written replenishment and substitution process. A backup pump also needs its compatible supplies and prepared users.

THREE QUESTIONS FOR THE SUPPLIER

1. Which approved references fit this exact pump and software?

2. What does the complete annual basket cost delivered?

3. What is the supported response to a set shortage?

Selected sources: B. Braun · BD · US FDA

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Sterile-processing project buying sheet

Buy capacity across the full route from used instruments to ready sets. First identify where the hospital's actual workload waits.

START HERE

Map set families, arrival times, return deadlines and capacity at collection, cleaning, preparation, sterilization and storage. Have the responsible teams review the complete pathway.

PAY MORE WHEN

A verified bottleneck requires added capacity and the supporting room, utilities, staff and quality-assurance resources are funded.

WATCH THE BUDGET

Include water treatment, steam generation, drainage, loading accessories, instruments, approved consumables, training and commissioning. Chamber volume alone cannot define useful output.

CHECK THE SUPPORT

Obtain exact-model technical data, compatible-load evidence, maintenance provision and an accountable process-verification plan before routine use.

THREE QUESTIONS FOR THE SUPPLIER

1. Which demonstrated bottleneck will your proposal resolve?

2. What utilities and site work does this configuration need?

3. Who provides the commissioning and verification reports?

Selected sources: WHO · US CDC · STERIS

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

MMS / EQUIPMENT BUYING GUIDE

Facility oxygen buying sheet

Choose an oxygen service from an assessed requirement, then compare the complete source-to-use arrangement. Concentrators, cylinders and a PSA plant each need accountable support and continuity.

START HERE

Obtain a qualified demand assessment, connected-device inlet requirements and site review. Use those same inputs for every supplier comparison.

PAY MORE WHEN

Additional generation or distribution resolves a verified need and the power, personnel, quality assurance, maintenance and backup are funded.

WATCH THE BUDGET

Include delivery logistics, cylinder terms, electricity, site work, distribution, training, spares and ongoing service. Compare a qualified service contract where available.

CHECK THE SUPPORT

Name the operating and technical owners, required commissioning evidence and approved response to power, source or delivery interruption.

THREE QUESTIONS FOR THE SUPPLIER

1. Which assessed demand and device requirements does this package meet?

2. What supply, quality and maintenance responsibilities do you retain?

3. How does service continue when the primary source is unavailable?

Selected sources: WHO · UNICEF · UNICEF Office of Innovation

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.

Equipment software buying sheet

Include software rights, support and recovery in the equipment purchase. Record what the receiving hospital may use and how that service can be restored.

START HERE

Create a schedule of required features, versions, licence terms, destination rights, recurring charges and named support providers. Verify it against the delivered configuration.

PAY MORE WHEN

A documented supported upgrade preserves useful equipment and meets the hospital's capability, security and continuity requirements at an acceptable complete cost.

WATCH THE BUDGET

Include interface work, renewals, activation, authorized recovery, training, migration and any required computer replacement.

CHECK THE SUPPORT

Obtain configuration-specific security information and support dates. Assign account ownership, update review, backups and recovery duties; establish which functions remain during an outage.

THREE QUESTIONS FOR THE SUPPLIER

1. Which feature rights transfer to this hospital?

2. What updates and support dates apply to this configuration?

3. Who restores licensed operation after the computer fails?

Selected sources: US FDA · Philips · GE HealthCare

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-17 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.