MMS / EQUIPMENT BUYING GUIDE

One better machine or two simpler ones: buying guide

Name the constraint before counting machines: required capability, daily capacity and continuity during repairs are different reasons to spend.

START HERE

Fund the smallest complete change that removes the real bottleneck. That may be another compatible endoscope, an ultrasound probe, more instrument sets or a repair-loaner arrangement before another console.

PAY MORE WHEN

Buy higher capability when the work needs it. A second X8 cannot back up required invasive pressure; another incompatible sterilizer cycle cannot replace the required T-Top load capability.

WATCH THE BUDGET

Two units each need accessories, supplies, maintenance and trained users. Check cleaning, drying, cooling and release before buying a second sterilizer. M11 automation may help staff workload where manual 2540M operation consumes attention.

CHECK THE SUPPORT

Test the backup against shared interruptions. Two oxygen-dependent machines can both stop when their source fails. A handheld ultrasound is useful backup only for examinations and export workflows it can actually cover.

THREE QUESTIONS FOR THE SUPPLIER

1. Which additional work becomes possible with this second complete package?

2. What shared power, oxygen, software or supply failure stops both units?

3. Would a scope, probe, processing improvement or loaner remove more downtime?

Selected sources: Tuttnauer · EDAN USA · Tuttnauer

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.