MMS / EQUIPMENT BUYING GUIDE

Oxygen and respiratory support buying guide

Size the oxygen source for simultaneous demand before buying downstream equipment. Oxygen generation, high-flow therapy, acute ventilation and sleep therapy are different jobs.

START HERE

Where 1-5 L/min meets the requirement, compare VisionAire 5 with the oxygen-monitor option. For requirements up to 10 L/min, compare the correct DeVilbiss 1025 and NewLife Intensity 10 electrical variants.

CONSIDER ALTERNATIVES WHEN

Price Airvo 2, Airvo 3 and HVT 2.0 as complete high-flow stations. For acute NIV, compare configured C1 or Savina. AirCurve ST-A, Astral and sleep CPAP require their own intended-use comparisons.

WATCH THE BUDGET

High-flow devices do not generate oxygen. Include source capacity, backup, chambers, tubing and interfaces. VisionAire's published 290 W and Intensity 10's 600 W make electricity a meaningful operating-cost input.

CHECK THE SUPPORT

Confirm approved source/device compatibility, concentration at required flow and repair access. Keep Airvo 2 and Airvo 3 correction records distinct; one does not resolve the other.

THREE QUESTIONS FOR THE SUPPLIER

1. What oxygen flow and pressure must our source deliver simultaneously?

2. Which voltage, oxygen monitoring and backup arrangements are included?

3. What is the annual complete consumable basket for this respiratory station?

Selected sources: Drive DeVilbiss · CAIRE · CAIRE

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.