MMS / EQUIPMENT BUYING GUIDE

Equipment quote comparison guide

Make competing offers cover the same work before comparing totals. Model-family names, upgrade promises and a low console price can conceal different purchases.

START HERE

Write one required configuration and have each supplier price it: patient sizes, functions, accessories, software, delivery, installation, acceptance and initial supplies. Mark every omission as included, priced separately or unresolved.

PAY MORE WHEN

Accept a higher equipment price when it lowers the complete cost or supplies a needed capability. X8 lacks the invasive-pressure option of X10/X12; T-Top Class B must actually be supplied.

WATCH THE BUDGET

Extend the comparison through a common operating period. Add consumable pack sizes, controls, usable expiry and scheduled service. Test quieter activity: unused stock can reverse an apparent per-item saving.

CHECK THE SUPPORT

Attach usable software rights and named repair coverage to each offer. Ultrasound needs named probes and enabled applications; a PC-based ECG needs a supported computer/software arrangement.

THREE QUESTIONS FOR THE SUPPLIER

1. Which required function or accessory is excluded from this total?

2. What changes in the recurring bill at half our expected workload?

3. Which unpriced delivery, installation or support item could reverse the comparison?

Selected sources: EDAN USA · Tuttnauer · MFI Medical

Read the full guide, model comparisons and source notes online →

Reviewed 2026-10-16 · Purchasing guidance; confirm the exact unit, local requirements and support before ordering.