



SISMOGRAMA

Start	February 10, 2026 11:58:09	Serial Number	UM16385
Finish	February 10, 2026 12:00:52	Model Number	Micromate ISEE 11.0AK
Number of Intervals/Interval	16.25/10 sec	Battery Level	3.8 volts
Sample Rate	1024 sps	Unit Calibration	April 22, 2025 by VMA LTDA
Setup File Name	SLM.mmb	Event File Name	BARRAGEM DE
Operator	THIAGO FERREIRA DA S	USB Sensor Support	Disables-UM16385_20260210115809.IDFH
		Scaled Distance	13528.3 (4278.0 m, 0.1 kg)
Notes			
Location:	BARRAGEM DE REJEITOS		
Client:	MVV		
User Name:	MINERACAO VALE VERDE		
General:			

Post Event Notes No text to be displayed.

Geophone	Tran	Vert	Long
Peak Particle Velocity	<0.127 mm/s	<0.127 mm/s	<0.127 mm/s
Zero Crossing Frequency	26.9 Hz	51.2 Hz	32.0 Hz
Date	Feb 10, 2026	Feb 10, 2026	Feb 10, 2026
Time	12:00:09	12:00:09	12:00:09
Sensor Check	✓ Passed	✓ Passed	✓ Passed
Frequency	7.3 Hz	7.3 Hz	7.1 Hz
Overswing Ratio	3.9	3.8	4.1
Peak Vector Sum	0.143 mm/s at February 10, 2026 12:00:09		

ISEE Linear Microphone	
Peak Sound Pressure Level	8.77 pa
Date	Feb 10, 2026
Time	12:00:52
Zero Crossing Frequency	N/A
Sensor Check	✗ Check
Frequency	0.0 Hz
Test Amplitude	0 mv

