



SISMOGRAMA

Start	February 10, 2026 11:57:24	Serial Number	UM16668	
Finish	February 10, 2026 12:02:15	Model Number	Micromate ISEE 11.0AK	
Number of Intervals/Interval	29.08/10 sec	Battery Level	3.5 volts	
Sample Rate	1024 sps	Unit Calibration	May 28, 2025 by VMA LTDA	
Setup File Name	MVV.MMB	Event File Name	COMUNIDADE DE	
Operator	THIAGO FERREIRA DA S	USB Sensor Support	TORROES-UM16668_20260210115724	
Notes		GPS Location	(1) IDFH	
Location:	COMUNIDADE DE TORROES	Source Location	Latitude	Longitude
Client:	MVV	Sensor Location	009 40.050 S	036 43.941 W
User Name:	MINERACAO VALE VERDE	Distance	009 39.898 S	036 43.091 W
General:		Scaled Distance	1577.4 m	
			4988.1 (1577.4 m, 0.1 kg)	

Post Event Notes No text to be displayed.

Geophone	Tran	Vert	Long
Peak Particle Velocity	0.457 mm/s	0.331 mm/s	0.497 mm/s
Zero Crossing Frequency	26.9 Hz	32.0 Hz	34.1 Hz
Date	Feb 10, 2026	Feb 10, 2026	Feb 10, 2026
Time	12:00:14	12:00:14	12:00:14
Sensor Check	Passed	Passed	Passed
Frequency	7.3 Hz	7.5 Hz	7.3 Hz
Overswing Ratio	4.4	4.3	4.4

Peak Vector Sum 0.598 mm/s at February 10, 2026 12:00:14

ISEE Linear Microphone	
Peak Sound Pressure Level	2.72 pa
Date	Feb 10, 2026
Time	11:59:04
Zero Crossing Frequency	5.7 Hz
Sensor Check	Passed
Frequency	20.5 Hz
Test Amplitude	1110 mv

