Bureau
International des
Poids et
Mesures

Le système international d'unités
The International System of Units

9 édition 2003

Annexe 2
Annex 2

Valeurs recommandées des fréquences étalons destinées à la mise en pratique de la définition du mètre et aux représentations secondaires de la seconde

Recommended values of standard frequencies for applications including the practical realization of the metre and secondary representations of the second

Partie 2
Part 2

Iodine ()

Iodine ()



CCL-CCTF Frequency Standards Working Group

Bureau International
des Poids et Mesures

Recommended values of standard frequencies for applications including the practical realization of the metre and secondary representations of the second
Part 2: Iodine ()

CCL-CCTF Frequency Standards Working Group

9th edition 2003

Appendix 2

1. CIPM recommended values

The values $f = 468\ 218\ 332.4$ MHz
$\lambda = 640\ 283\ 468.7$ fm

with a relative standard uncertainty of 4.5×10^{-10} apply to the radiation of a He-Ne laser stabilized with an internal iodine cell having a cold-finger temperature of (16 ± 1) °C[(1)] and a frequency modulation width, peak-to-peak, of (6 ± 1) MHz.

[(1)] For the specification of operating conditions, such as temperature, modulation width and laser power, the symbols ± refer to a tolerance, not an uncertainty.

2. Source data

Adopted value

$f = 468\ 218\ 332.4\ (2)$ MHz
$u_c/y = 4.5 \times 10^{-10}$

for which:
$\lambda = 640\ 283\ 468.7\ (3)$ fm
$u_c/y = 4.5 \times 10^{-10}$

calculated from

f/ kHz	u_c/y	source data
468 218 332 419	1.0×10^{-10}	Section 2.1
468 218 332 310	1.2×10^{-10}	Section 2.2
468 218 332 069	4.6×10^{-10}	Section 2.3
Weighted mean	$f = 468\ 218\ 332\ 366$ kHz	

Given the small number of determinations, the CCL considered it prudent to assume a relative standard uncertainty of 4.5×10^{-10}.

Source data

2.1.

Bennet and Mills-Baker [1] give $\lambda_{a9} = 640.283\ 468\ 6$ nm.

From this paper the ratio f_{a9}/f_i is calculated as

$f_{a9}/f_i = 0.988\ 611\ 184\ 191$
$u_c/y = 1 \times 10^{-10}$.

Using the recommended value of the absorbing molecule $^{127}I_2$, a_{16} or f component, R(127) 11-5 transition (see iodine at $\lambda \approx 633$ nm and frequency differences listed in corresponding Table 1, see p. 8) one obtains

$f_i = 473\ 612\ 214\ 712$ kHz
$u_c/y = 2.2 \times 10^{-11}$,

one calculates

$f_{a9} = 468\ 218\ 332\ 434$ kHz
$u_c/y = 1.0 \times 10^{-10}$

at a cold-finger temperature of 14.3 °C (iodine pressure = 16 Pa) and a modulation width of 7 MHz. For a reference temperature of 16 °C (iodine pressure = 18.9 Pa) and a modulation width of 6 MHz, corrections of −23 kHz and + 8 kHz has to be applied to this value assuming a pressure dependence of −7.8 kHz Pa^{-1} and a modulation dependence of −7.6 kHz MHz^{-1}, similar to that reported in [2], giving

$f_{a9} = 468\ 218\ 332\ 419$ kHz

$u_c/y = 1.0 \times 10^{-10}$.

2.2.

Zhao et al. [2], [3] give $\lambda_{a9} = 640.283\,468\,8$ nm
$3 \times (u_c/y) = 1.1 \times 10^{-9}$.

Bönsch [4] gives $\lambda_i/\lambda_{a9} = 0.988\,611\,183\,8$
$u_c/y = 12 \times 10^{-11}$.

Using the recommended value of the absorbing molecule $^{127}I_2$, a_{16} or f component, R(127) 11-5 transition (see iodine at $\lambda \approx 633$ nm and frequency differences listed in corresponding Table 1, see p. 8) one obtains

$f_i = 473\,612\,214\,712$ kHz
$u_c/y = 2.2 \times 10^{-11}$,

one calculates

$f_{a9} = 468\,218\,332\,277$ kHz
$u_c/y = 1.2 \times 10^{-10}$

at a cold-finger temperature of 18 °C (iodine pressure = 22.6 Pa) and a modulation width of 6.5 MHz. For a reference temperature of 16 °C (iodine pressure = 18.9 Pa) and a modulation width of 6 MHz, corrections of $+29$ kHz and $+4$ kHz have to be applied to this value knowing a pressure dependence of -7.8 kHz Pa^{-1} and a modulation dependence of -7.6 kHz MHz^{-1}, giving

$f_{a9} = 468\,218\,332\,310$ kHz
$u_c/y = 1.2 \times 10^{-10}$.

2.3.

Reference [5], [6] give λ_{a9} (17 °C)/λ_e (20 °C) $= 1.011\,520\,341\,04$
$u_c/y = 4.6 \times 10^{-10}$.

Using the recommended value of the absorbing molecule $^{127}I_2$, a_{16} or f component, R(127) 11-5 transition (see iodine at $\lambda \approx 633$ nm and frequency differences listed in corresponding Table 1, see p. 8) one obtains

$f_e = 473\,612\,366\,967$ kHz
$u_c/y = 2.2 \times 10^{-11}$,

one calculates

$f_{a9} = 468\,218\,332\,055$ kHz
$u_c/y = 4.6 \times 10^{-10}$

at a cold-finger temperature of 17 °C (iodine pressure = 20.7 Pa). For a reference temperature of 16 °C (iodine pressure = 18.9 Pa), a correction of $+14$ kHz has to be applied to this value, assuming a pressure dependence of -7.8 kHz Pa^{-1} similar to that reported in [2], giving

$f_{a9} = 468\,218\,332\,069$ kHz

$u_c/y = 4.6 \times 10^{-10}$.

3. Absolute frequency of the other transitions related to those adopted as recommended and frequency intervals between transitions and hyperfine components

These tables replace those published in BIPM Com. Cons. Long., 2001, **10**, 188 and *Metrologia*, 2003, **40**, 128.

The notation for the transitions and the components is that used in the source references. The values adopted for the frequency intervals are the weighted means of the values given in the references.

For the uncertainties, account has been taken of:

- the uncertainties given by the authors;
- the spread in the different determinations of a single component;
- the effect of any perturbing components;
- the difference between the calculated and the measured values.

In the tables, uc represents the estimated combined standard uncertainty (1σ).

All transitions in molecular iodine refer to the B-X system.

Table 1

$\lambda \approx 640$ nm $^{127}I_2$ P(10) 8-5					
a_n	$[f(a_n) - f(a_9)]$/MHz	u_c/ MHz	a_n	$[f(a_n) - f(a_9)]$/MHz	u_c/ MHz
a_1	-495.4	0.4	a_9	0	—
a_2	-241.5	0.7	a_{10}	77.84	0.03
a_3	-233.0	0.4	a_{11}	186.22	0.07
a_4	-177.8	1.3	a_{12}	199.51	0.07
a_5	-175.2	0.6	a_{13}	256.6	0.2
a_6	-130.8	0.1	a_{14}	272.75	0.07
a_7	-82.45	0.03	a_{15}	374.0	0.2
a_8	-61.85	0.14			
Frequency referenced to	a_9, P(10) 8-5, $^{127}I_2$: $f = 468\,218\,332.4$ MHz [7]				
Ref. [8], [9], [10], [11], [12], [13], [14], [15]					

Table 2

$\lambda \approx 640$ nm $^{127}I_2$ R(16) 8-5		
b_n	$[f(b_n) - f(a_9)]$/MHz	u_c/ MHz
b_1	62.834	0.01

b_2	329.8	0.2
b_3	335.99	0.02

Frequency referenced to a_9, P(10) 8-5, $^{127}I_2$: $f = 468\ 218\ 332.4$ MHz [7]

Ref. [8], [9], [10], [11], [12], [13], [14], [15]

Absorbing molecule $^{127}I_2$, a_9 component, P(10) 8-5 transition[1]

[1] All transitions in I_2 refer to the $B^3\Pi\, 0_u^+ - X^1\, \Sigma_g^+$ system.

References

[1] Bennett S. J., Mills-Baker P., Iodine Stabilized 640 nm Helium-Neon laser, *Opt. Commun.*,1984, **51**, 322-324.

[2] Zhao K. G., Blabla J., Helmcke J., $^{127}I_2$-Stabilized ^{3}He-^{22}Ne Laser at 640 nm Wavelength, *IEEE Trans. Instrum. Meas.*, 1985, **IM-34**, 252-256.

[3] CCDM/92-10a, NIM, Research findings in realizing the definition of the metre measurement/intercomparison of frequency (wavelength) and geometrical standard of length.

[4] Bönsch G., Simultaneous Wavelength Comparison of Iodine-Stabilized Lasers at 515 nm, 633 nm, and 640 nm, *IEEE Trans. Instrum. Meas.*, 1985, **IM-34**, 248-251.

[5] CCDM/92-20a, BIPM, Reply to the Questionnaire for the CCDM.

[6] CCDM/92-6a, IMGC, Reply to questionnaire CCDM/92-1, 5 June 1992.

[7] Recommendation CCL3 (*BIPM Com. Cons. Long.*, 10th Meeting, 2001) adopted by the Comité International des Poids et Mesures at its 91th Meeting as Recommendation 1 (CI-2002).

[8] Gläser M., Hyperfine Components of Iodine for Optical Frequency Standards *PTB-Bericht*, 1987, **PTB-Opt-25**.

[9] Bertinetto F., Cordiale P., Fontana S., Picotto G. B., Recent Progresses in He-Ne Lasers Stabilized to $^{127}I_2$, *IEEE Trans. Instrum. Meas.*, 1985, **IM-34**, 256-261.

[10] Bennett S. J., Cérez P., Hyperfine Structure in Iodine at the 612-nm and 640-nm Helium-Neon Laser Wavelengths, *Opt. Commun.*, 1978, **25**, 343-347.

[11] Kegung D., Xu J., Li C.-Y., Liu H.-T., Hyperfine Structure in Iodine Observed at the 612 nm and 640 nm ^{3}He-^{22}Ne Laser Wavelengths, *Acta Metrologica Sinica*, 1982, **3**, 322-323.

[12] Zhao K., Li H., Hyperfine structure of iodine at 640 nm ^{3}He-^{22}Ne laser wavelength and identification, *Acta Metrologica Sinica*, 1983, **3**, 673-677.

[13] Zhao K.-G., Li H., Analysis and Calculation of Hyperfine Lines of Iodine Molecule, *Acta Metrologica Sinica*, 1985, **6**, 83-88.0-2c.

[14] Gläser M., Identification of Hyperfine Structure Components of the Iodine Molecule at 640 nm Wavelength, *Opt. Commun.*, 1985, **54**, 335-342.

[15] Zhao K.-G., Li C.-Y., Li H., Xu J., Way H., Investigations of $^{127}I_2$-Stabilized He−Ne Laser at 640 nm, *Acta Metrologica Sinica*, 1987, **8**, 88-95.