

Le système international d'unités The International System of Units

9 édition 2005

Annexe 2

Annex 2

Valeurs recommandées des fréquences étalons
destinées à la mise en pratique de la définition du
mètre et aux représentations secondaires de la
seconde


**Recommended values of standard frequencies
for applications including the practical
realization of the metre and secondary
representations of the second**

Partie 2

Part 2

Calcium () (2005)

Calcium () (2005)



**Recommended values of standard
frequencies for applications including
the practical realization of the metre and
secondary representations of the second
Part 2: Calcium () (2005)**

CCL-CCTF Frequency Standards Working Group

9th edition 2005

1. CIPM recommended values

The values $f = 455\,986\,240\,494\,140\text{ Hz}$
 $\lambda = 657\,459\,439.291\,683\text{ fm}$

with a relative standard uncertainty of 1.8×10^{-14} apply to the radiation of a laser stabilized to Ca atoms. The values correspond to the mean frequency of the two recoil-split components for atoms which are effectively stationary, i.e. the values are corrected for the second-order Doppler shift.

2. Source data

Adopted value $f = 455\,986\,240\,494\,140\,(8)\,\text{Hz}$

$$u_c/y = 1.8 \times 10^{-14}$$

for which:

$$\lambda = 657\,459\,439.291\,683\,(12)\,\text{fm}$$

$$u_c/y = 1.8 \times 10^{-14}$$

calculated from

f/Hz	u_c/y	source data
455 986 240 494 144	1.2×10^{-14}	[1]
455 986 240 494 135.8	7.5×10^{-15}	[2], [3]
Unweighted mean:	$f = 455\,986\,240\,494\,140\,\text{Hz}$	

The CCL decided to adopt the unweighted mean of the two values, with an uncertainty of 8 Hz, equal to the difference between the values.

Absorbing atom ^{40}Ca , $^1\text{S}_0 - ^3\text{P}_1$; $\Delta m_J = 0$ transition

References

- [1] Degenhardt C., Stoeck H., Lisdat Ch., Wilpers G., Schnatz H., Lipphardt B., Nazarova T., Pottie P.-E., Sterr U., Helmcke J., Riehle F., Calcium Optical Frequency Standard with Ultracold Atoms: Approaching 10^{-15} Relative Uncertainty, *Phys. Rev. A*, **72**, 062111/1-17, 2005.
- [2] Wilpers G., Oates C. W., Hollberg L., Improved Uncertainty Budget for Optical Frequency Measurements with Microkelvin Neutral Atoms: Results for a High-Stability ^{40}Ca Optical Frequency Standard, *Appl. Phys. B* **85**, pp. 31-44, 2006.
- [3] Wilpers G., Oates C. W., Diddams S. A., Bartels A., Fortier T. M., Oskay W. H., Bergquist J. C., Jefferts S. R., Heavner T. P., Parker T. E., Hollberg L., Absolute Frequency Measurement of the Neutral ^{40}Ca Optical Frequency Standard at 657 nm based on Microkelvin Atoms, *Metrologia*, **44**, pp. 146-151, 2007.