Bureau International des Poids et Mesures

Comparison of ozone reference standards of the CHMI the BIPM

3rd edition 2018

Bureau International
des Poids et Mesures

3rd edition 2018

Copyright statement

This document is distributed under the terms of the Creative Commons Attribution 4.0 International License (http://creativecommons.org/licenses/by/4.0/:), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made.

Édité par le BIPM
Pavillon de Breteuil
F-92312 Sèvres Cedex
FRANCE

Foreword

A comparison of the ozone reference standards of the Czech Hydrometeorological Institute (CHMI) and of the Bureau International des Poids et Mesures (BIPM) has been performed. Both institutes maintain Standard Reference Photometers (SRPs), developed by the National Institute of Standards and Technology (NIST), as their reference standards. The instruments were compared over a nominal ozone mole fraction range of 0 $\mathrm{nmol mol^{-1}}$ to 860 $\mathrm{nmol mol^{-1}}$ and the results showed good agreement.

Comparison of ozone reference standards of the CHMI and the BIPM

Contents

1. Introduction 7

2. Terms and definitions 8

3. Measurement schedule 9

4. Measurement protocol 10

4.1. Ozone generation 10
4.2. Comparison procedure 10
4.3. Comparison repeatability 11
4.4. SRP27 stability check 11

5. Measurement standards 12

5.1. Measurement equation of a NIST SRP 12
5.2. Absorption cross–section for ozone 13
5.3. Condition of the BIPM SRPs 13
5.4. Uncertainty budget of the common reference BIPM-SRP27 13
5.5. Covariance terms for the common reference BIPM-SRP27 14
5.6. Condition of the CHMI SRP17 15
5.7. Uncertainty budget of the CHMI SRP17 15

6. Analysis of the measurement results by generalized least-squares regression 16

7. Results of the comparison 17

7.1. Measurement results 17
7.2. Repeatability 18
7.3. History of comparisons between BIPM SRP27, SRP28 and CHMI SRP17 19

8. Conclusion 21

References 22

1. Introduction

A comparison of the ozone reference standards of the Czech Hydrometeorological Institute (CHMI) and of the Bureau International des Poids et Mesures (BIPM) has been performed. Both institutes maintain Standard Reference Photometers (SRPs) developed by the National Institute of Standards and Technology (NIST) as their reference standards. This comparison was performed following the protocol established for the key comparison BIPM.QM-K1, adapted to the measurement range normally covered by the CHMI as part of their calibration service, which is 10 $\mathrm{nmolmol^{-1}}$ to 870 $\mathrm{nmolmol^{-1}}$. The measurement protocol is described briefly in Chapter 4. A description of the standards is given in Chapter 5 of this report, together with their uncertainty budgets. The data treatment is explained in Chapter 6, and the results of the comparison are given in Chapter 7.

2. Terms and definitions

No terms and definitions are listed in this document.

x_{nom} nominal ozone mole fraction in dry air supplied by the ozone generator.

$x_{\mathrm{A},i}$ ith measurement of the nominal value x_{nom} by the photometer A.

$\bar{x}_{\mathrm{A}}$ the mean of N measurements of the nominal value x_{nom} measured by the photometer A : $\bar{x}_{\mathrm{A}} = \frac{1}{N}\sum_{i=1}^{N} x_{\mathrm{A},i}$

s_{A} standard deviation of N measurements of the nominal value x_{nom} measured by the photometer A : $s_{\mathrm{A}}^2 = \frac{1}{N-1}\sum_{i=1}^{N}\left(x_{\mathrm{A},i} - \bar{x}_{\mathrm{A}}\right)^2$

- The result of the linear regression fit performed between two sets of data measured by photometers A and B during a comparison is written as: $x_{\mathrm{A}} = a_{\mathrm{A,B}} x_{\mathrm{B}} + b_{\mathrm{A,B}}$. With this notation, photometer A is compared to photometer B. $a_{\mathrm{A,B}}$ is dimensionless and $b_{\mathrm{A,B}}$ is expressed in units of $\mathrm{nmolmol}^{-1}$.

3. Measurement schedule

The measurements were performed on 20 September 2017 at the BIPM.

4. Measurement protocol

This comparison was performed following the protocol established for the key comparison BIPM.QM-K1, with a modified measurement range: the instruments were compared over a nominal ozone mole fraction range of 0 $\mathrm{nmolmol^{-1}}$ to 860 $\mathrm{nmolmol^{-1}}$ (instead of a nominal range of 0 $\mathrm{nmolmol^{-1}}$ to 500 $\mathrm{nmolmol^{-1}}$).

The comparison protocol is summarized in this section. The complete version can be downloaded from the BIPM website (http://www.bipm.org/utils/en/pdf/BIPM.QM-K1_protocol.pdf).

This comparison was performed following protocol A, corresponding to a direct comparison between the CHMI national standard SRP17 and the common reference standard BIPM-SRP27 maintained at the BIPM. A comparison between two (or more) ozone photometers consists of producing ozone-air mixtures at different amount-of-substance fractions over the required range, and measuring these with the photometers.

4.1. Ozone generation

The same source of purified air is used for all the ozone photometers being compared. This air is used to provide reference air as well as the ozone–air mixture to each ozone photometer. Ambient air is used as the source for reference air. The air is compressed with an oil-free compressor. It is dried and scrubbed using a commercial purification system so that the mole fraction of ozone and nitrogen oxides remaining in the air is below detectable limits. The relative humidity of the reference air is monitored and the mole fraction of water in air is typically less than 3 $\mathrm{\mu molmol^{-1}}$. The mole fraction of volatile organic hydrocarbons in the reference air was measured (November 2002), with no mole fraction of any detected component exceeding 1 $\mathrm{nmolmol^{-1}}$.

A common dual external manifold in Pyrex is used to supply the necessary flows of reference air and ozone–air mixtures to the ozone photometers. The two columns of this manifold are vented to atmospheric pressure.

4.2. Comparison procedure

All the instruments were switched on and allowed to stabilize for at least 8 hours before the comparison commenced. The pressure and temperature measurement systems of the instruments were checked at this time. If any adjustments were required, these were noted. For this comparison, no adjustments were necessary.

One comparison run normally includes ten different amount-of-substance fractions distributed to cover the range, together with the measurement of reference air at the beginning and end of each run. The nominal amount-of-substance fractions were measured in a sequence imposed by the BIPM.QM-K1 protocol, plus three additional values to extend the range (0, 220, 80, 420, 120, 320, 30, 370, 170, 500, 270, 0, 570, 840, 700 and 0) $\mathrm{nmolmol^{-1}}$. Each of these points is an average of ten single measurements.

For each nominal value of the ozone mole fraction x_{nom} supplied by the ozone generator, the standard deviation s_{SRP27} on the set of ten consecutive measurements $x_{\mathrm{SRP27},i}$ recorded by

BIPM-SRP27 was calculated. The measurement results were considered as valid if s_{SRP27} was less than 1 nmolmol^{-1}, which ensures that the photometers were measuring a stable ozone concentration. If not, another series of ten consecutive measurements was performed.

4.3. Comparison repeatability

The comparison procedure was repeated continuously for BIPM.QM-K1 to evaluate its repeatability. The participant and the BIPM jointly decided when both instruments were stable enough to start recording a set of measurement results to be considered as the official comparison results. The measurements reported here were recorded in an additional run that was performed immediately afterwards.

4.4. SRP27 stability check

A second ozone reference standard, BIPM-SRP28, was included in the comparison to verify its agreement with BIPM-SRP27 and thus follow its stability over the period of the on-going key comparison.

5. Measurement standards

All instruments included in this comparison were Standard Reference Photometers (SRP) built by the NIST. More details on the instrument's operating principle and capabilities can be found in [1]. The following section describes the SRP measurement principle and uncertainty budget.

5.1. Measurement equation of a NIST SRP

The measurement of ozone mole fraction by an SRP is based on the absorption of radiation at 253.7 nm by ozonized air in the gas cells of the instrument. One unique characteristic of the instrument design is the use of two gas cells to overcome the instability of the light source. The measurement equation is derived from the Beer–Lambert and ideal gas laws. The number concentration (C) of ozone is calculated from:

$$C = \frac{-1}{2\sigma L_{\text{opt}}} \frac{T}{T_{\text{std}}} \frac{P_{\text{std}}}{P} \ln(D) \tag{1}$$

where

σ is the absorption cross-section of ozone at 253.7 nm under standard conditions of temperature and pressure, 1.1476×10^{-17} cm^2/molecule ISO 13964:1996;

L_{opt} is the optical path length of one of the cells;

T is the measured temperature of the cells;

T_{std} is the standard temperature (273.15 K);

P is the measured pressure of the cells;

P_{std} is the standard pressure (101.325 kPa);

D is the product of transmittances of two cells, with the transmittance (T_r) of one cell defined as

$$T_r = \frac{I_{\text{ozone}}}{I_{\text{air}}} \tag{2}$$

where

I_{ozone} is the UV radiation intensity measured from the cell when containing ozonized air; and

I_{air} is the UV radiation intensity measured from the cell when containing pure air (also called reference or zero air).

Using the ideal gas law, Equation (1) can be reformulated in order to express the measurement results as a mole fraction (x) of ozone in air:

$$x = \frac{-1}{2\sigma L_{\text{opt}}} \frac{T}{P} \frac{R}{N_{\text{A}}} \ln(D) \tag{3}$$

where

N_{A} is the Avogadro constant, 6.022142×10^{23} mol^{-1}, and

R is the gas constant, 8.314472 $\mathrm{Jmol^{-1}K^{-1}}$.

The formulation implemented in the SRP software is:

$$x = \frac{-1}{2\alpha_x L_{opt}} \frac{T}{T_{std}} \frac{P_{std}}{P} \ln(D) \quad (4)$$

where α_x is the linear absorption coefficient under standard conditions, expressed in $\mathrm{cm^{-1}}$, linked to the absorption cross–section with the relation:

$$\alpha_x = \sigma \frac{N_A}{R} \frac{P_{std}}{T_{std}} \quad (5)$$

5.2. Absorption cross–section for ozone

The linear absorption coefficient at standard conditions α_x used within the SRP software algorithm is 308.32 $\mathrm{cm^{-1}}$. This corresponds to a value for the absorption cross–section σ of 1.1476×10^{-17} $\mathrm{cm^2/molecule}$, rather than the more often quoted 1.147×10^{-17} $\mathrm{cm^2/molecule}$. In the comparison of two SRP instruments, the absorption cross section can be considered to have a conventional value and its uncertainty can be set to zero. However, in the comparison of different methods or when considering the complete uncertainty budget of the method the uncertainty of the absorption cross–section should be taken into account. A consensus value of 2.12 % at a 95 % level of confidence for the uncertainty of the absorption cross–section has been proposed by the BIPM and the NIST in a recent publication [3].

5.3. Condition of the BIPM SRPs

Compared to the original design described in [1], SRP27 and SRP28 have been modified to deal with two biases revealed by a study conducted by the BIPM and the NIST [3]. In 2009, an “SRP upgrade kit” was installed in the instruments, as described in [4].

5.4. Uncertainty budget of the common reference BIPM-SRP27

The uncertainty budget for the ozone mole fraction in dry air (x) measured by the instruments BIPM-SRP27 and BIPM-SRP28 in the nominal range 0 $\mathrm{nmolmol^{-1}}$ to 900 $\mathrm{nmolmol^{-1}}$ is given in Table 1.

Table 1. Uncertainty budget for the SRPs maintained by the BIPM

Component (y)	Uncertainty $u(y)$	Sensitivity coefficient $\lvert c_i \rvert = u(x)/u(y)$	Contribution to $u(x)$ nmolmol^{-1}

	Source	Distribution	Standard Uncertainty	Combined standard uncertainty $u(y)$	
Optical Path L_{opt}	Measurement scale	Rectangular	0.0006 cm	0.52 cm	$-\frac{x}{L_{opt}}2.89 \times 10^{-3}x$
	Repeatability	Normal	0.01 cm		
	Correction factor	Rectangular	0.52 cm		
Pressure P	Pressure gauge	Rectangular	0.029 kPa	0.034 kPa	$-\frac{x}{P}3.37 \times 10^{-4}x$
	Difference between cells	Rectangular	0.017 kPa		
Temperature T	Temperature probe	Rectangular	0.03 K	0.07 K	$\frac{x}{T}2.29 \times 10^{-4}x$
	Temperature gradient	Rectangular	0.058 K		
Ratio of intensities D	Scaler resolution	Rectangular	8×10^{-6}	1.4×10^{-5}	$\frac{x}{D\ln(D)}$ 0.28
	Repeatability	Triangular	1.1×10^{-5}		
Absorption Cross section σ	Hearn value		1.22×10^{-19} cm²/molecule	1.22×10^{-19} cm²/molecule	$-\frac{x}{\sigma}8.06 \times 10^{-2}x$

As explained in the comparison protocol, following this budget the standard uncertainty associated with the ozone mole fraction measurement with the BIPM SRPs can be expressed as a numerical equation (numerical values expressed as $\mathrm{nmol\,mol^{-1}}$):

$$u(x) = \sqrt{(0.28)^2 + (2.92 \cdot 10^{-3}x)^2} \tag{6}$$

5.5. Covariance terms for the common reference BIPM-SRP27

Correlations in-between the results of two measurements, performed at two different ozone amount-of-substance fractions with the BIPM–SRP27, were taken into account in the OzonE software. More details on the covariance expression can be found in the protocol. The following expression was applied:

$$u(x_i, x_j) = x_i \cdot x_j \cdot u_b^2 \tag{7}$$

where:

$$u_b^2 = \frac{u^2(T)}{T^2} + \frac{u^2(P)}{P^2} + \frac{u^2(L_{opt})}{L_{opt}^2} \tag{8}$$

The value of u_b is given by the expression of the measurement uncertainty: $u_b = 2.92 \times 10^{-3}$.

5.6. Condition of the CHMI SRP17

Compared to the original design, the CHMI SRP17 has been modified to deal with the two biases revealed in [3]. In August 2007, an "SRP upgrade kit" was installed by NIST, as already described in the previous comparison report [5].

5.7. Uncertainty budget of the CHMI SRP17

The uncertainty budget for the ozone mole fraction in dry air x measured by the CHMI standard SRP17 in the nominal range 0 nmolmol^{-1} to 900 nmolmol^{-1} is given in Table 2.

Following this budget, as explained in the protocol of the comparison, the standard uncertainty associated with the ozone mole fraction measurement with the CHMI SRP17 can be expressed as a numerical equation (numerical values expressed as nmolmol^{-1}):

$$u(x) = \sqrt{(0.28)^2 + (2.92 \cdot 10^{-3} x)^2} \tag{9}$$

No covariance term for the CHMI standard SRP17 was included in the calculations.

Table 2. SRP17 uncertainty budget

Component (y)	Uncertainty $u(y)$				Sensitivity coefficient $c_i = \frac{\partial x}{\partial y}$	Contribution to $u(x)$ nmolmol^{-1}
	Source	Distribution	Standard Uncertainty	Combined standard uncertainty $u(y)$		
Optical Path L_{opt}	Measurement scale	Rectangular	0.005 cm	0.52 cm	$-\frac{x}{L_{opt}}$	$-2.89 \times 10^{-3} x$
	Variability	Rectangular	0.004 cm			
	Divergence	Rectangular	0.52 cm			
Pressure P	Pressure gauge	Rectangular	0.029 kPa	0.034 kPa	$-\frac{x}{P}$	$-3.37 \times 10^{-4} x$
	Difference between cells	Rectangular	0.017 kPa			
Temperature T	Temperature probe	Rectangular	0.03 K	0.07 K	$\frac{x}{T}$	$2.29 \times 10^{-4} x$
	Temperature gradient	Rectangular	0.058 K			
Ratio of intensities D	Scaler resolution	Rectangular	8×10^{-6}	1.4×10^{-5}	$\frac{x}{D\ln(D)}$	0.28
	Repeatability	Triangular	1.1×10^{-5}			
Absorption Cross section α	Hearn value		1.22×10^{-19} cm^2/molecule	1.22×10^{-19} cm^2/molecule	$-\frac{x}{\alpha}$	$-1.06 \times 10^{-2} x$

6. Analysis of the measurement results by generalized least-squares regression

The relationship between the national and reference standards was first evaluated with a generalized least-squares regression fit, using the OzonE software. This software, which is documented in a publication [6], is an extension of the previously used software B_Least recommended by ISO standard 6143:2001 ISO 6143.2:2001. It includes the possibility to take into account correlations between measurements performed with the same instrument at different ozone mole fractions. It also facilitates the use of a transfer standard, by the handling of unavoidable correlations, which arise since this instrument needs to be calibrated by the reference standard.

In a direct comparison, a linear relationship between the ozone amount-of-substance fractions measured by SRP*n* and SRP27 is obtained:

$$x_{\mathrm{SRP}n} = a_0 + a_1 x_{\mathrm{SRP27}} \tag{10}$$

The associated uncertainties on the slope $u(a_1)$ and the intercept $u(a_0)$ are given by OzonE, as well as the covariance between them and the usual statistical parameters to validate the fitting function.

7. Results of the comparison

SRP17, maintained by the CHMI, was compared with the SRPs maintained by the BIPM following the general procedure outlined above. A cycle of twelve comparison runs between SRP27, SRP28 and SRP17 were performed. Ozone was generated using the Environics 6100 generator with airflow of 10 L min^{-1}. The nominal ozone mole fraction range covered during this exercise was 0 nmolmol^{-1} to 500 nmolmol^{-1} for the eleven first runs, then 0 nmolmol^{-1} to 840 nmolmol^{-1} for the last run. The result of the last comparison run is presented in detail in the Section 7.1. The repeatability of the results over the twelve runs is shown in Section 7.2, and the stability of the BIPM standards in Section 7.3.

7.1. Measurement results

The measurement results of the last of the twelve recorded runs are shown in Table 3. For each nominal ozone mole fraction, the standard deviation is reported on the ten successive measurements that were recorded. The values reported here show that both instruments were in a stable regime.

An interesting way to look at these results is to display the difference between the ozone amount-of-substance fractions measured by SRP17 and SRP27 (x_{17}–x_{27}) versus the ozone mole fraction measured by SRP27. Figure 1, shows that measurement results from both instruments differ little from each other over the entire concentration range of the comparison. Considering a coverage factor of $k = 2$ (95 % confidence interval), the difference between both standards is much lower than their combined measurement uncertainties.

Table 3. measurement results of the comparison between the CHMI standard SRP17 and the BIPM reference standard SRP27

Reference standard BIPM-SRP27 (RS)			National standard CHMI SRP17 (NS)		
x_{RS} nmolmol^{-1}	s_{RS} nmolmol^{-1}	$u(x_{RS})$ nmolmol^{-1}	x_{NS} nmolmol^{-1}	s_{NS} nmolmol^{-1}	$u(x_{NS})$ nmolmol^{-1}
0.18	0.35	0.28	0.04	0.32	0.28
201.04	0.25	0.65	201.71	0.14	0.65
72.00	0.23	0.35	72.26	0.22	0.35
386.31	0.24	1.16	387.34	0.21	1.17
108.73	0.17	0.42	109.17	0.15	0.42
292.78	0.20	0.90	293.60	0.27	0.90
25.94	0.26	0.29	26.04	0.22	0.29
339.56	0.47	1.03	340.71	0.49	1.03
154.23	0.25	0.53	154.80	0.27	0.53
465.61	0.22	1.39	466.93	0.38	1.39
247.91	0.32	0.78	248.57	0.17	0.78
0.12	0.17	0.28	-0.05	0.27	0.28
576.5	0.5	1.71	578.12	0.38	1.71
842.1	0.4	2.47	844.58	0.33	2.48
697.9	0.2	2.06	700.08	0.22	2.06

−0.10	0.10	0.28	−0.04	0.28	0.28




Figure 1 — difference between the ozone amount-of-substance fractions measured by SRP17 and SRP27 versus the ozone mole fraction measured by SRP27.

The relationship between SRP17 and SRP27 is given by the result of the generalized least-square regression performed following the method described in Chapter 6:

$$x_{\mathrm{SRP17}} = 1.0034 x_{\mathrm{SRP27}} - 0.06 \tag{11}$$

The standard uncertainties on the parameters of the regression are $u(a_1) = 0.0032$ for the slope and $u(a_0) = 0.21$ nmolmol^{-1} for the intercept. The covariance between the two parameters is cov $(a_0, a_1) = -1.66 \times 10^{-4}$.

The least-squares regression results confirm that a linear fit is appropriate, with a sum of the the squared deviations (SSD) of 0.30 and a goodness of fit (GoF) equal to 0.17.

To assess the agreement of the standards using Equation (10), the difference between the calculated slope value and unity, and the intercept value and zero, together with their measurement uncertainties, need to be considered. In this comparison, the value of the intercept is consistent with an intercept of zero, considering the uncertainty in the value of this parameter; i.e $|a_0| < 2u(a_0)$, and the value of the slope is consistent with a slope of 1; i.e. $|1 - a_1| < 2u(a_1)$.

7.2. Repeatability

The results of the twelve comparison runs performed successively between SRP17 and SRP27 are reported in Table 4. With a standard deviation of 0.04 nmolmol^{-1} on the intercept and a relative standard deviation of 0.03 % on the slope, these data show a good repeatability.

Table 4. results of the ten comparison runs repeated successively

Run	Slope a_1	$u(a_1)$	Intercept a_0 / (nmolmol^{-1})	$u(a_0)$ / (nmolmol^{-1})	cov (a_0, a_1)	GoF
1	1.0027	0.0033	0.09	0.22	$-2.05\ 10^{-4}$	0.395
2	1.0030	0.0033	−0.04	0.22	$-2.05\ 10^{-4}$	0.273
3	1.0030	0.0033	0.04	0.22	$-2.04\ 10^{-4}$	0.391
4	1.0025	0.0033	0.05	0.22	$-2.04\ 10^{-4}$	0.363
5	1.0025	0.0033	0.05	0.22	$-2.04\ 10^{-4}$	0.400
6	1.0022	0.0033	0.09	0.22	$-2.04\ 10^{-4}$	0.756
7	1.0032	0.0033	−0.01	0.22	$-2.04\ 10^{-4}$	0.277
8	1.0030	0.0033	0.09	0.22	$-2.04\ 10^{-4}$	0.333
9	1.0022	0.0033	0.08	0.22	$-2.04\ 10^{-4}$	0.381
10	1.0032	0.0033	0.00	0.22	$-2.05\ 10^{-4}$	0.473
11	1.0026	0.0033	0.07	0.22	$-2.04\ 10^{-4}$	0.629
12	1.0032	0.0032	−0.02	0.18	$-1.19\ 10^{-4}$	0.260

7.3. History of comparisons between BIPM SRP27, SRP28 and CHMI SRP17

Results of the previous comparison performed with CHMI, on the same measurement range and since the first one in December 2002, are shown in Figure 2 together with the results of this comparison. The slopes a_1 of the linear relation $x_{\mathrm{SRP}n} = a_0 + a_1 x_{\mathrm{SRP27}}$ are represented together with their associated uncertainties calculated at the time of each comparison. Results of previous comparisons have been corrected to take into account the changes in the reference BIPM-SRP27 described in [4], which explains the larger uncertainties associated with the corresponding slopes.

Figure 2 shows that all standards included in these comparisons remained in close agreement.




Figure 2 — Results of previous comparisons between SRP27, SRP28 and CHMI-SRP17 realised at the BIPM. Uncertainties are calculated at $k = 2$, with the uncertainty budget in use at the time of each comparison.

8. Conclusion

For the sixth time in fifteen years, a direct comparison between the CHMI national standard SRP17 and the BIPM reference standard SRP27 has been conducted at the BIPM. The measurement range of the SRP17 at the CHMI is 10 $\mathrm{nmolmol^{-1}}$ to 870 $\mathrm{nmolmol^{-1}}$. This entire range was examined during the comparison, and the relative difference between the CHMI SRP17 and BIPM SRP27 measurement results was found to be much smaller than their combined measurement uncertainties.

References

[1] Paur R.J., Bass A.M., Norris J.E. and Buckley T.J., 2003, Standard reference photometer for the assay of ozone in calibration atmospheres *Env. Sci. Technol.* **NISTIR 6369**, 25 pp

[2] ISO 13964:1996 *ISO 13964 : 1996 Ambient air — Determination of ozone — Ultraviolet photometric method (International Organization for Standardization)*

[3] Viallon J., Moussay P., Norris J.E., Guenther F.R. and Wielgosz R.I., 2006, A study of systematic biases and measurement uncertainties in ozone mole fraction measurements with the NIST Standard Reference Photometer *Metrologia* **43** 441-450

[4] Viallon J., Moussay P., Idrees F. and Wielgosz R.I., 2010, Upgrade of the BIPM Standard Reference Photometers for Ozone and the effect on the on-going key comparison BIPM.QM-K1, **Rapport BIPM-2010/07**, 16pp

[5] Viallon J., Moussay P., Wielgosz R.I., Novák J. and Vokoun M., 2008, Final report of the on-going key comparison BIPM.QM-K1: Ozone at ambient level, comparison with CHMI, 2007 *Metrologia* **45**, ***Tech. Supl.*** 08005

[6] Bremser W., Viallon J. and Wielgosz R.I., 2007, Influence of correlation on the assessment of measurement result compatibility over a dynamic range *Metrologia* **44** 495-504

[7] ISO 6143.2:2001 ISO 6143.2 : 2001 Gas analysis — Determination of the composition of calibration gas mixtures — Comparison methods (International Organization for Standardization)

Document Control

Authors: Philippe Moussay (BIPM), Faraz Idrees (BIPM), Robert Wielgosz (BIPM), Jan Šilhavý (CHMI), and Miroslav Vokoun (CHMI)


BIPM
ΜΕΤΡΩ ΧΡΩ