Recording 1216: what synchronising the clocks does (dotted lines: Primus airway-pressure pauses)
A. Each clock against the synchronised timeline, as recorded
clock time minus synchronised time (s)
SNUADC record stamps
BIS EEG stamps
Primus airway-pressure stamps
B. After synchronisation
typical error left: SNUADC about 3 ms, BIS 42 ms, Primus 43 ms (bands)
C. BIS against SNUADC, from shared bursts, as recorded
EEG minus ECG, same burst (s)
D. The same bursts, after synchronisation
spread after: 26 ms (median |deviation|)
E. Primus against SNUADC, from the breath cycle, as recorded
airway-to-CVP lag, minus its median (s)
hours into the recording
F. After synchronisation
spread after: 43 ms (median |leave-one-out error|)
hours into the recording