Timing in VitalDB: a user guide

How to judge whether two signals are aligned well enough for your question

Draft, October 2026

Abstract

Every time in a clinical database carries two kinds of uncertainty: how far it may be from the true time (*accuracy*) and how wide an interval it stands for (*resolution*). In VitalDB these range from about a millisecond for waveform samples to five minutes for anaesthesia records and a whole day for admission dates. This guide summarises what we measured for each kind of time, gives a rule of thumb for deciding whether two signals can be compared at the timescale of a research question, and works through examples. It follows the framework of Goodwin et al. [1] and applies it to VitalDB [2].

Note (October 2026). This guide uses the *accuracy* and *resolution* of Goodwin et al. [1]. The metrology atlas (vitaldb.metrology-atlas.org) now states the same quantities in the vocabulary of the VIM and GUM: a systematic error (bias), corrected where known, and a standard uncertainty with named components, the interval a value describes being part of what the value is. Its numbers supersede the ones here; in particular the raw BIS stamps come about 0.79 s *early*, not late.

1 Why timing matters

Most analyses of VitalDB compare things recorded by different devices or people: a drug infusion and a blood pressure response, a laboratory result and the waveforms around it, an anaesthesia event and a monitor trend. Each source has its own clock, its own delay and its own rounding. Goodwin et al. [1] describe how such timing errors arise in clinical databases and why they are easy to overlook: the data look precise, because every value carries a timestamp to the millisecond, even when the event it describes is known only to the nearest few minutes. This guide makes those uncertainties explicit for VitalDB, so you can check before you analyse.

2 Two numbers for every time

Accuracy The offset between a recorded time and the true time on a common reference. Here the reference is the recorder clock, as reconstructed from the SNUADC sample count (the *stamp model*). Accuracy includes clock offsets, transmission and processing delays, and documentation lag.

Resolution The width of the interval a time stands for. A 500 Hz sample stands for 2 ms. A heart rate averaged over 6 s stands for 6 s, even though its stamp is to the millisecond. An event entered on a five-minute grid stands for five minutes.

Two further points matter in practice:

- **The stamp is not the event.** A numeric value is stamped when the monitor reports it, which is after its averaging window closes. Treat the stamp as the time the value became *available*, not the time the physiology happened. This is also the rule to follow to avoid leaking future information into a prediction model.

- **Rounding direction matters.** If a time is rounded up, the true time lies before it; if rounded to nearest, it lies on either side. Section 6 shows that VitalDB's five-minute anaesthesia entries are rounded up.

3 The accuracy and resolution map

Figure 1 places every kind of time in VitalDB on one log–log map. Each ellipse spans the measured (solid) or estimated (dashed) range on each axis. Arrows run from raw device stamps to their position after synchronisation. The key gives the evidence for each ellipse.




Figure 1: Accuracy (offset from the recorder clock, x) and resolution (width of the interval a time stands for, y) for each kind of time in VitalDB. Blue: waveform samples. Orange: numeric values. Green: records, anaesthesia (EMR) and laboratory times.

Table 1 lists typical values. Numbers are medians or ranges over the cases examined; your case may differ.

4 Rule of thumb

To decide whether two signals A and B can be compared for a given question:

1. **Find both on the map**, at their synchronised position if you apply the synchronisation.

Source	Accuracy (offset)	Resolution
SNUADC waveforms (ECG, ART, PLETH, CVP)	reference; raw stamps sawtooth up to ∼480 ms	2–10 ms (100–500 Hz)
BIS EEG	−1.05 to −0.53 s raw; 16–33 ms after sync	7.8 ms (128 Hz)
Primus airway pressure and CO_2	0.4–4 s raw; ∼50 ms after breath matching	16 ms (62.5 Hz)
CardioQ ABP and flow	0.4–4.4 s raw; ∼0.24 s after beat matching	5.6 ms (180 Hz)
Solar8000 HR, ART, CVP	1.9–5 s after the averaging window closes	3–6 s windows plus 2 s polling
SpO_2 and PLETH_HR	onset ∼4 s	10–20 s smoothing (∼12 beats for PLETH_HR)
Primus $ETCO_2$, PIP, RR	1.8–8.5 s	last breath to 30 s; ∼7 s records
EV1000 ART_MBP	1.6 s median, p95 3.3 s	20 s
Vigilance HR_AVG	window centre 6–21 min behind	running median over ∼20–30 min
NIBP	result 20–40 s after the cuff measurement	one reading
Infusion pumps (Orchestra)	∼1 s; propofol → BIS response 132 s median	1 s records
Anaesthesia/operation events, 5-min entries	rounded up: true time 0–300 s earlier	300 s
Anaesthesia events, minute entries	documentation lag (minutes)	60 s
Laboratory results	result 3–12 min after the sample	draw time not recorded
Admission, discharge	—	1 day

Table 1: Typical accuracy and resolution by source.

2. **Take the largest edge.** For each ellipse, read its right edge (worst accuracy) and its top edge (widest resolution). The relative alignment uncertainty U is the largest of these four numbers:

$$U \approx \max\left(a_A,\ a_B,\ r_A,\ r_B\right).$$

   If A and B come from the same device clock (for example ECG and ART, both from SNUADC), their offsets cancel: drop a_A and a_B and use the resolutions alone. Prefer the right edge to the centre, so the estimate is conservative.

3. **Compare with the timescale T of your question**, the shortest time difference your conclusion depends on:

   - $U < T/10$: **safe**. Timing error will not change your answer.
   - $T/10 \le U \le 10\,T$: **marginal**. Model the uncertainty explicitly (for example, treat times as uniform intervals, shift by the known lag, or run a sensitivity analysis over the plausible offsets).
   - $U > 10\,T$: **unsafe**. The data cannot answer the question at this timescale. Ask a coarser question or use a better-timed source.

The rule uses the maximum rather than a sum of squares because the largest term usually dominates by an order of magnitude, and because offsets are often systematic rather than random.

5 Worked examples

1. Pulse arrival time from ECG and arterial waveform. Both from SNUADC, same clock: drop accuracy. Resolution 2 ms each, so $U \approx 2$ ms (plus up to 0.8 ms skew from sequential channel reads). Pulse arrival time is ∼200–300 ms and changes of interest are ∼10 ms: $U/T \approx 0.2$. *Marginal*: average over beats or interpolate fiducial points.

2. Arterial waveform against the BIS EEG. After synchronisation BIS sits at ∼33 ms accuracy; resolution 8 ms. $U \approx 33$ ms. For a question about EEG changes over seconds ($T = 1$ s), $U/T = 0.03$: *safe*. Without synchronisation the raw offset is up to 1 s: *marginal*.

3. Heart rate numeric against arterial pressure numeric. Both Solar8000 numerics: same clock, but each value is an average over a window and reported after a delay. The widest resolution is ∼6 s; delays differ between tracks by ∼2 s. $U \approx 6$ s. For trends over minutes, *safe*; for beat-to-beat coupling, use the waveforms instead.

4. Propofol infusion start against BIS. Pump records are ∼1 s; BIS is a 1 s numeric with vendor smoothing (estimated tens of seconds). U is tens of seconds. The median response is 132 s, so for onset latency $U/T \approx 0.2$: *marginal*, and the measured latencies (none negative across 497 cases) are plausible.

5. SpO_2 against a drug recorded in the anaesthesia record. The anaesthesia entry has 300 s resolution. For a 30 s desaturation response, $U/T = 10$: *unsafe*. For whether saturation fell in the 30 minutes after induction, $U/T = 0.17$: *marginal but usable*.

6. A laboratory value against the waveforms at the time of sampling. The result time is 3–12 min after the draw, and the draw time is not recorded. Using the result time as the sample time gives $U \approx 12$ min. Questions on a scale of an hour are marginal; minute-scale questions are unsafe unless the draw is located from the arterial line access artefact, which brings U down to the access duration (∼0.5–2 min).

6 Anaesthesia record times

Anaesthesia start and end, and operation start and end, are mostly entered on absolute five-minute clock marks (92–98% of cases; 52% for anaesthesia start, which mixes minute and five-minute entries). We compared these times with device markers on the recorder clock: infusion start and stop, loss and return of BIS, start and end of ventilation and of end-tidal CO_2, and the end of the recording. Across all 6,388 cases, five-minute entries sit 90–290 s later relative to these markers than minute entries do. For all eleven markers a rounding-up model fits better than rounding to nearest or down.

Practical reading: for a five-minute entry at time t, the true time lies in $(t - 300\,\text{s}, t]$; the best estimate is $t - 150\,\text{s}$ with a standard deviation of $300/\sqrt{12} \approx 87\,\text{s}$, plus any documentation lag. Minute entries carry 60 s resolution and their own lag. Do not treat the two kinds as equally precise.

7 Numerics: lag without editing timestamps

Monitor numerics are computed from the waveforms over a window and reported after a delay. We fitted the window and delay for several tracks (for example, Solar8000 HR is close to a median over ∼6 s, reported ∼3.6 s after the window closes). Three conventions help:

- Keep the recorded stamps. They are the availability time, which is correct for real-time and prediction work.
- When you need the physiological time, subtract the delay *and* half the window to reach the window centre.
- Treat the window as the resolution of that value when applying the rule of thumb.

The connectors for TimeNet and AtriumDB attach these properties as labels that span each recording or signal run, rather than altering the data.

8 Value resolution, briefly

Timing and value resolution interact when a threshold defines an event. SNUADC waveforms are 10-bit: about 1 mmHg per count for arterial pressure and 10 µV for ECG, with the arterial signal saturating near 400 mmHg and ECG often clipping. Laboratory values in the released CSV drop trailing zeros, and haematocrit is recorded to an integer when drawn with a blood gas and to 0.1 otherwise. Check value labels before using thresholds near these limits.

9 Checklist before you analyse

1. List every source your analysis compares and find each on Figure 1.
2. Decide whether you need availability time (prediction) or physiological time (mechanism).
3. Apply the rule of thumb at the shortest timescale your conclusion depends on.
4. If marginal, model the uncertainty; if unsafe, change the question or the source.
5. Report the timing assumptions with your results, following the recommendations in [1].

References

[1] Andrew J Goodwin, Danny Eytan, William Dixon, Sebastian D Goodfellow, Zakary Doherty, Robert W Greer, Alistair McEwan, Mark Tracy, Peter C Laussen, Azadeh Assadi, et al. Timing errors and temporal uncertainty in clinical databases—a narrative review. *Frontiers in Digital Health*, 4:932599, 2022.

[2] Hyung-Chul Lee, Yoonsang Park, Soo Bin Yoon, Seong Mi Yang, Dongnyeok Park, and Chul-Woo Jung. VitalDB, a high-fidelity multi-parameter vital signs database in surgical patients. *Scientific Data*, 9(1):279, 2022.