




Paper



Code



## Motivation

- Ultrasound Resolution is limited by e.g. diffraction and finite apertures that **constrain diagnostic capabilities**
- Retrieving the **exact location** of reflectors and scatterers based on a processed B-mode image is **difficult** due to unknown acquisition parameters and processing step



## Image Formation Process

### Signal Processing Pipeline




### Convolutional Model

The interaction between acoustic wave and tissue can be modeled by a convolution between a filter and the echogenicity map.

The filter is called the **Point Spread Function (PSF)** of the imaging system and incorporates acquisition parameters such as aperture size, acquisition frequency, etc.

$$y(z') = h(z') \otimes \gamma(z') + n(z')$$

$y$  : radio-frequency data,  $h$  : PSF,  $\gamma$  : echogenicity map,  $n$  : gaussian noise,  $\otimes$  : convolution



## Deconvolution

Deconvolution model **estimate the PSF** of the imaging system to **convolve** the acquired image with **its inverse**.

### Blind vs. Non-Blind Deconvolution

**Blind Deconvolution:** Estimate the PSF **based on the acquired data** [2]

**Non-Blind Deconvolution:** Estimate an **analytical** form of the PSF **based on equations** [3]

The derived PSF is used as the filter in deconvolution algorithms, such as the Wiener Filter or the Richardson-Lucy Algorithm.

→ But these methods **lack flexibility**, necessitate **careful tuning of regularizers**, and require **high computational resources** when applied to large datasets

## Point-Spread-Function

Adapting the model from [4] we define the PSF to:

$$h(x', z') = h_{lat}(x') \otimes h_{ax}(z')$$

The **lateral pressure** distribution at the focal point can be written as a **convolution** of the Fourier transformation of the **transmit**  $A_T$  and **receive** function  $A_T$  :


$$h_{lat}(x' | \lambda, D, r) = \mathcal{F}(A_T) \otimes \mathcal{F}(A_T) = D^2 \text{sinc}^2\left(\frac{Dx'}{r\lambda}\right)$$

We remove the envelope detection step, in accordance with [3], and model the **axial signal** to:


$$h_{ax}(z' | \sigma_z) = \exp\left(-\frac{z'^2}{2 * \sigma_z^2}\right)$$

Where  $\sigma_z$  = scaling factor based on the number of cycles for the axial pulse.

Where  $D$  = Aperture and  $\lambda$  = wavelength.



## Proposed Approach



### (1) Implicit Neural Representation (INR)

- For a given **input coordinate x**, the surrounding grid vertices lead to an entry in the **learnable** look-up tables. The look-up tables contains one vector per vertex, which is used as the input to a linear interpolation.
- An L-dimensional **multi-resolution grid** provides L entries which are concatenated to form the input to the MLP. The MLP outputs a single scalar **echogenicity value**.

### (2) Differentiable Image processing pipeline

- The output of the MLP is **convolved** with an isotropic PSF
- A **differentiable** signal **processing pipeline** transforms the convolved data via log-compression to the imaging domain


### (3) Loss Computation

- Loss function= Combination of a total variation loss (TV), structural similarity index (SSIM) loss, and a mean-squared error (MSE) loss

## Results

### Synthetic Results (CIRS Phantom)


- Comparison with the Richardson-Lucy Algorithm
- Synthetic images are created with a known PSF (f = 10 MHz, pulses= 5, f-number = 2.0)



	Wire Targets		Cylindrical Inclusions	
	PSNR	SSIM	PSNR	SSIM
Richardson-Lucy	17.35	0.06	16.89	0.21
Proposed Method	<b>17.85</b>	<b>0.07</b>	<b>17.85</b>	<b>0.29</b>


### In-Silico Results (CIRS Phantom)

**10 of 12 wires** are correctly identified with a mean radius of 0.053 mm (+- 0.01), expected value is 0.040 mm, based on processed data (thresholding, filtering and clustering)



### In-Vivo Results (Carotid Data)

Using the predicted **echogenicity map** we can change the “acquisition parameters” to generate new B-Mode Images



## Summary

We effectively leverage an **INR** for learning a continuous representation of the **echogenicity map** based only on the B-mode image and a deconvolutional model including a **fully differentiable** rendering pipeline which integrates a PSF that is **based on equations** derived from the wave equations.

## References

- [1] Hyun, Dongwoon. “Advanced Ultrasound Imaging”, Summer semester 2023, Stanford, University. Lecture.
- [2] Li, X., Zhang, X., Fan, C., Chen, Y., Zheng, J., Gao, J., & Shen, Y. (2024). Deconvolution based on sparsity and continuity improves the quality of ultrasound image. Computers in Biology and Medicine, 169, 107860.
- [3] Dalitz, C., Pohle-Frohlich, R., & Michalk, T. (2015). Point spread functions and deconvolution of ultrasonic images. IEEE transactions on ultrasonics, ferroelectrics, and frequency control, 62(3), 531-544
- [4] Walker, W. F., & Trahey, G. E. (1998). The application of k-space in pulse echo ultrasound. IEEE transactions on ultrasonics, ferroelectrics, and frequency control, 45(3), 541-558.
- [5] Müller, T., Evans, A., Schied, C., & Keller, A. (2022). Instant neural graphics primitives with a multiresolution hash encoding. ACM transactions on graphics (TOG), 41(4), 1-15.