

UltraScatter: Ray-Based Simulation of Ultrasound Scattering


Felix Duelmer^{1,2,3,4}, Mohammad Farid Azampour^{1,2}, and Nassir Navab^{1,2}

¹Chair of Computer Aided Medical Procedures (CAMP), School of Computation, Information and Technology, Technical University of Munich, Munich, Germany

²Munich Centre for Machine Learning (MCML), Munich, Germany

³Institute of Biological and Medical Imaging, Bioengineering Center, Helmholtz Zentrum München, Neuherberg, Germany

⁴Chair of Biological Imaging, Central Institute for Translational Cancer Research (TranslaTUM), School of Medicine and Health & School of Computation, Information and Technology, Technical University of Munich, Munich, Germany



Poster No. A1P-19

Introduction & Motivation

- **Ultrasound simulation** is critical for designing/improving reconstruction algorithms, transducers, and training ML models.
- However, **wave-based solvers** such as k-space methods [1] are computationally intensive, and their high accuracy is not always required.
- Faster convolution-based tools such as Field II [2] or delay-based simulators like SIMUS [3] cut runtime yet still take *minutes* per B-mode frame and assume weak, single scattering.

→ **UltraScatter** replaces waves with ray fronts and leverages highly parallelized **Monte-Carlo ray tracing** to model multiple scattering, attenuation and absorption *in seconds*.

Methodology

We are interested in the received pressure P at element e over time t :

$$P(e, t) = \int_{\Omega} \int_A P_i(x, t, \omega_i) f_a(\omega_i) d\omega da,$$

Scattering in the heterogeneous media is defined by the **scattering probability** (extinction coefficient) σ_t and the **scattering amplitude** (albedo) a .

$$P_{scattered}(x, t, \omega_o) = \int_{\Omega} a(x) p(\omega_i, \omega_o) P_i(x, t, \omega_i) d\omega_i$$

Following the methodology of **free flight delta tracking** with null interactions (Woodcock tracking [5]):

- The distance s a ray is travelling until an interaction is triggered is sampled based on the **extinction majorant** μ (largest defined extinction value) and a randomly sampled variable $\xi \sim \mathcal{U}(0,1)$:

$$s = s_{min} - \frac{1}{\mu} \ln(1 - \xi)$$

- This interaction can lead to a real or null interaction event based on the **extinction value** at the sampled position x :

$$\xi' < \frac{\sigma_t(x)}{\mu} \Rightarrow \text{real interaction, else null interaction}$$

- In case of a **real** interaction event the scattering amplitude defines whether the ray is absorbed or scatters

If the ray is scattered, secondary rays are cast to **all individual elements**

The phase function p defines the subsequent direction ω_o of the primary ray


CONCLUSION

- UltraScatter approximates wave propagation using parallel ray tracing, enabling **clinical-like B-mode imaging** at a fraction of the cost of frequency-domain solvers.
- UltraScatter is **modular** and can be easily extended for elevational focusing, heterogeneous speed of sound, etc.
- **Future work:** Speckle coherence during probe translation, full 3-D transmit/receive, refraction & harmonic generation.


Idea & Implementation

- Physics-Based Rendering (PBR) has achieved remarkable image fidelity and advanced simulators (such as Mitsuba 3 [4]) facilitate synthetic image generation
- By using a **Monte-Carlo** ray tracing approach multiple nested integrals can be solved computationally efficiently
- We are building on top of **Mitsuba 3 [4]** – a modular framework that allows us to add modules (e.g. transducer) but also customize and adapt existing software (e.g. integrator)

Pipeline



Results



REFERENCES

- [1] B. E. Treeby and B. T. Cox, "k-wave: Matlab toolbox for the simulation and reconstruction of photoacoustic wave fields," *Journal of biomedical optics*, vol. 15, no. 2, pp. 021 314–021 314, 2010.
- [2] J. A. Jensen, "A model for the propagation and scattering of ultrasound in tissue," *The Journal of the Acoustical Society of America*, vol. 89, no. 1, pp. 182–190, 1991.
- [3] D. Garcia, "Simus: An open-source simulator for medical ultrasound imaging. part i: Theory & examples," *Computer Methods and Programs in Biomedicine*, vol. 218, p. 106726, 2022.
- [4] W. Jakob, S. Speierer, N. Roussel, M. Nimier-David, D. Vicini, T. Zeltner, B. Nicolet, M. Crespo, V. Leroy, and Z. Zhang, "Mitsuba 3 renderer," 2022, <https://mitsuba-renderer.org>.
- [5] E. Woodcock, T. Murphy, P. Hemmings, and S. Longworth, "Techniques used in the gem code for monte carlo neutronics calculations in reactors and other systems of complex geometry," in *Proc. Conf. Applications of Computing Methods to Reactor Problems*, vol. 557, no. 2. Argonne National Laboratory, 1965.