

FIELD INSTRUMENT CALIBRATION RECORDS - CALIBRATION LOG - PRP

Project Site/FacID: _____

Calibrated by (Print)/Affiliation: Kayla Wingard / CBA

Boldly "X" this box if there is qualified data on this page.

Temperature (Quarterly)

Date of Last Temp Verification: _____

See log book: CBA Quarterly Temp Ver. Log

DISSOLVED OXYGEN (DO) (REFERENCE: DEP SOP FT 1500)

Acceptance Criteria +/-0.3 mg DO/L

Meter/Instrument Name and Unique ID: CBA Blue 451

CAL	ICV	CCV	Initials	Date	Time	Standard (DO %)	Temp °C	DO Saturation mg/L (100%)**	Response DO (%)	Response mg DO/L	Deviation mg DO/L	Pass or Fail
CAL	ICV	CCV	KW	9-8-25	0908	100%	23.0	8.578	100.7	—	—	P F
CAL	ICV	CCV	I	1	0909	100%	1	1	100.2	8.59	0.012	P F
CAL	ICV	CCV	KW	9-12-25	1146	100%	22.5	8.660	100.0	8.65	0.01	P F
CAL	ICV	CCV				100%						P F
CAL	ICV	CCV				100%						P F
CAL	ICV	CCV				100%						P F

** See Table FS 2200-2 and/or Table FT 1500-1 for Dissolved Oxygen 100% Saturation (mg/L) corresponding to Temperature.

SPECIFIC CONDUCTANCE (REFERENCE: DEP SOP FT 1200)

Acceptance Criteria +/-5% the standard

Meter/Instrument Name and Unique ID: CBA Blue 451

CAL	ICV	CCV	Initials	Date	Time	Standard (µmho/cm)	Exp. Date	Lot #	Response (µmho/cm)	Deviation (%)	Pass or Fail
CAL	ICV	CCV	KW	9-8-25	0852	1413	10-8-25	256100424	1360	3.75	P F
CAL	ICV	CCV	I	1	1	1	1	1	1414	0.07	P F
CAL	ICV	CCV	I	1	0855	12900	10-2-25	2509041	12959	0.45	P F
CAL	ICV	CCV	KW	9-12-25	1147	1413	10-8-25	256100424	1413	0.0	P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

OXIDATION-REDUCTION POTENTIAL (ORP)

Acceptance Criteria +/-10 mV

REFERENCE: EPA Region 4, Operating Procedure, Field Measurement of Oxidation-Reduction Potential (ORP)

Meter/Instrument Name and Unique ID: _____

CAL	ICV	CCV	Initials	Date	Time	Standard (mV)	Exp. Date	Lot #	Response (mV)	Deviation (mV)	Pass or Fail
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F
CAL	ICV	CCV									P F

Perform ICVs and CCVs only in "READ/RUN" mode.

CAL - Calibration; ICV - Initial Calibration Verification; and, CCV - Continuing Calibration Verification.

Deviation (%) = 100-((Response/Standard)*100)

FIELD INSTRUMENT CALIBRATION RECORDS - CALIBRATION LOG - PRP

Project Site/FacID: _____

Calibrated by (Print)/Affiliation: Kayla Wingard / CBA

Boldly "X" this box if there is qualified data on this page.

TURBIDITY (REFERENCE: DEP SOP FT 1600)

Meter/Instrument Name and Unique ID: CBA Blue YSI

Std=0.1-10 NTU +/-10%			Std=11-40 NTU +/-8%			Std=41-100 NTU +/-6.5%			Std>100 NTU +/-5%		
CAL	ICV	CCV	Initials	Date	Time	Standard (NTU)	Exp. Date	Lot #	Response (NTU)	Deviation (%)	Pass or Fail
☒	ICV	CCV	KW	9-8-25	0915	Filtered DI water (0.8 FNU)			0.67		P F
☒	ICV	CCV				124 FNU	2-1-26	25G24023167	125.1	0.89	☒ P F
☒	ICV	CCV			0926	12.4 FNU	11-1-25	25C24019290	12.95	4.44	☒ P F
☒	ICV	CCV	KW	9-12-25	1131	12.4 FNU	11-1-25	25C4019290	12.89	3.95	☒ P F
☒	ICV	CCV			1137	124 FNU	2-1-26	25G24023167	124.06	0.05	☒ P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F

pH (REFERENCE: DEP SOP FT 1100)

Acceptance Criteria +/-0.2 SU

Meter/Instrument Name and Unique ID: CBA Blue YSI

CAL	ICV	CCV	Initials	Date	Time	Standard (SU)	Exp. Date	Lot #	Response (SU)	Deviation (SU)	Pass or Fail
☒	ICV	CCV	KW	9-8-25	0931	7.00	3-8-26	25F28	6.90	0.1	P F
☒	ICV	CCV				4.00	2-1-26	25F1R	3.96	0.04	P F
☒	ICV	CCV				10.00		25F3T	9.93	0.07	P F
☒	ICV	CCV			0936				10.01	0.01	☒ P F
☒	ICV	CCV	KW	9-12-25	1153	4.00	3-12-26	25F1R	4.02	0.02	☒ P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F
☐	ICV	CCV									P F

Perform ICVs and CCVs only in "READ/RUN" mode.

CAL - Calibration; ICV - Initial Calibration Verification; and, CCV - Continuing Calibration Verification.

Deviation (%) = 100-((Response/Standard)*100)