DATA LOG

File Created At: 2026/07/27 01:15:37

Device Information

Device Code : RC-4HPro
Serial Number : EFX267100110
Mode Code : TLCFFA089
Probe Type : Temperature & Humidity(internal)
Firmware Version : V1.4

Trip Information

Trip Id: 0000001
Description : Temperature recording.

Mark Event

Configuration Information

Start Mode : Manual
Start Delay : 0s
Time Base : UTC +00:00
Log Interval : 5m
Stop Mode : Manual + Software

Alarm Zone	Allow Time	Alarm Type	Total Time	Violations	Status
TH: **over** 60.0 °C	0s	**Sin**	0s	0	OK
Ideal Region	**unlimited**		25m		
TL: **below** -30.0 °C	0s	**Sin**	0s	0	OK
HH: **over** 90.0 %	0s	**Sin**	20m	1	Alarm
Ideal Region	**unlimited**		5m		
HL: **below** 10.0 %	0s	**Sin**	0s	0	OK

Logging Summary

Highest : 12.1 °C / 90.8 %
Lowest : 11.9 °C / 83.8 %
Average : 12.0 °C / 89.3 %
Alarm At(Te): N/A
Alarm At(Hu): 2026/07/27 00:54:12
Start Time : 2026/07/27 00:49:12
Stop Time : 2026/07/27 01:14:12(temporary)
Elapsed Time : 25m
Data Points : 6



Date	Time	°C	%RH	Date	Time	°C	%RH	Date	Time	°C	%RH	Date	Time	°C	%RH	Date	Time	°C	%RH	Date	Time	°C	%RH
26/07/27	00:49:12	11.9	83.8																				
26/07/27	00:54:12	12.0	90.8																				
26/07/27	00:59:12	12.1	90.5																				
26/07/27	01:04:12	12.0	90.6																				
26/07/27	01:09:12	12.0	90.6																				
26/07/27	01:14:12	12.0	89.5																				