

Quality Report

Pika L 100121-278 Sensor SN: 23020627

Status: PASS

Average FWHM across all bands: 2.3 pixels

FWHM: PASS

Maximum Smile: 0.3 pixels.

Smile: PASS

Tilt < 1 pixel: PASS

Maximum wavelength error: 0.9 nm Wavelength Calibration: **PASS**

Slit transmission: PASS

Spectral Line Location: PASS

Vibration Testing Performed: July 2020 (.99 g-rms, 20-1k Hz, 11 Hrs)

Technician:JG

Signature:

Date: 2020-07-27




Calibration Report

Pika L 100121-278 Sensor SN: 23020627

Wavelength Calibration Status: PASS

Calibration Results:

A	В	С	
0.0001115	0.933445	88.273	



Technician: JG

Signature:

Date: 2020-07-27