

Endovascular Today - US Device Guide - Cerebral Stents

Company Name	Product Name	Materials Used	Stent Diameters (mm)	Stent Lengths (mm)	Recommended Guidewire Size (inch)	Recommended Guide Catheter Size (inch)
Cerenovus	Codman Enterprise 2 Vascular Reconstruction Device	Nitinol, laser-cut, closed-cell design with platinum markers	5 (for vessels 2.5–4)	16, 23, 30, 39	No guidewire	5- or 6-F guiding catheter for a 0.021 delivery microcatheter
Cerenovus	Codman Enterprise Vascular Reconstruction Device	Nitinol	4.5 (for vessels 2.5–4)	14, 22, 28, 37	No guidewire	5- or 6-F guiding catheter for a 0.021 delivery microcatheter
Stryker	Neuroform Atlas Stent System	Nitinol	3, 4, 4.5	15, 21, 24, 30	0.014	0.0165
Stryker	Neuroform EZ Stent System	Nitinol	2.5, 3, 3.5, 4, 4.5	10, 15, 20, 30	0.014	ID: 0.027
Stryker	Wingspan Stent System	Nitinol	2.5, 3, 3.5, 4, 4.5	9, 15, 20	0.014	Minimum ID: 0.064
Terumo Neuro	LVIS Intraluminal Support Device	Nitinol with tantalum markers	3.5, 4, 4.5, 5.5	12–33	0.014	ID: 0.021
Terumo Neuro	LVIS Jr Intraluminal Support Device	Nitinol with tantalum markers	2.5, 3.5	13–33	0.014	ID: 0.0165, 0.017