COBRE Scan Information

Below is more information on the directory structure for the COBRE imaging data. Also below are the imaging parameters for each series.

Directory structure:
var/www/html/dropbox/1139_anonymized/human: internal directory structure
/dicom OR /behavioral: this directory is the data type. **/dicom** contains DICOM files and **/behavioral** contains timing and log files for the GATE functional task.
/triotim: scanner name
/vcalhoun: PI
/cobre_07325: Study Name and Study Number
/A000XXXXX: Participant ID
/Session_1: scan session number. Due to the fact that COBRE data was collected across several different sub studies, some participants have more than one scan session. The scan session numbers are chronological.

Series Label Description
dti_35dir
Diffusion tensor imaging (DTI) is a magnetic resonance imaging technique that enables the measurement of the restricted diffusion of water in tissue in order to produce neural tract images instead of using this data solely for the purpose of assigning contrast or colors to pixels in a cross sectional image.
GATE_V01_R01, GATE_V01_R02, GATE_V01_R03, GATE_V01_R04
During the gating task, participants passively listened to single tones (2000 or 3000 Hz) or pairs of identical (two 2000 Hz tones) or pairs of non-identical (2000 and 3000 Hz) tones. The gating task therefore exhibits the brain's ability to inhibit repetitive versus novel stimuli.
mprage_5e/mprage_5e_RMS
T1-weighted scans refer to a set of standard scans that depict differences in the spin-lattice (or T1) relaxation time of various tissues within the body. T1 weighted images can be acquired using either spin echo or gradient-echo sequences.
RSTpre_V01_R01
Subjects rest passively with their eyes open or closed. Often used as a baseline for comparison for other tasks.
t1_midline_Sag
T1-weighted scans refer to a set of standard scans that depict differences in the spin-lattice (or T1) relaxation time of various tissues within the body. T1 weighted images can be acquired using either spin echo or gradient-echo sequences.
t2_tse_tra_192
Image made with a sequence with long TR and TE to show contrast in tissues with varying T2 relaxation times; water gives a strong signal.

mprage_5e

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 6:03    PAT: 2    Voxel size: 1.0×1.0×1.0 mm    Rel. SNR: 1.00    USER: tfl_mgh_multiecho

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slab group 1
    Slabs                         1
    Dist. factor                  50 %
    Position                      R2.7 A35.1 H10.8
    Orientation                   S > T-2.8 > C2.6
    Phase enc. dir.               A >> P
    Rotation                      0.00 deg
  Phase oversampling              0 %
  Slice oversampling              0.0 %
  Slices per slab                 192
  FoV read                        256 mm
  FoV phase                       100.0 %
  Slice thickness                 1.00 mm
  TR                              2530 ms
  TE 1                            1.64 ms
  TE 2                            3.5 ms
  TE 3                            5.36 ms
  TE 4                            7.22 ms
  TE 5                            9.08 ms
  Averages                        1
  Concatenations                  1
  Filter                          Prescan Normalize
  Coil elements                   HEA;HEP
Contrast
  Magn. preparation               Non-sel. IR
  TI                              1200 ms
  Flip angle                      7.0 deg
  Fat suppr.                      None
  Water suppr.                    None
  ...............................................
  Averaging mode                  Long term
  Reconstruction                  Magnitude
  Measurements                    1
  Multiple series                 Each measurement
Resolution
  Base resolution                 256
  Phase resolution                100 %
  Slice resolution                100 %
  Phase partial Fourier           Off
  Slice partial Fourier           Off
  Interpolation                   Off
  ...............................................
  PAT mode                        GRAPPA
  Accel. factor PE                2
  Ref. lines PE                   32
  Accel. factor 3D                1
  Matrix Coil Mode                Auto (Triple)

  Reference scan mode             Integrated
  ...............................................
  Image Filter                    Off
  Distortion Corr.                Off
  Unfiltered images               Off
  Prescan Normalize               On
  Normalize                       Off
  B1 filter                       Off
  Raw filter                      Off
  Elliptical filter               Off
Geometry
  Multi-slice mode                Single shot
  Series                          Interleaved
System
  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                REF
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Save uncombined                 Off
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Tune up
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      Isocenter
    Orientation                   Transversal
    Rotation                      0.00 deg
    R >> L                        350 mm
    A >> P                        263 mm
    F >> H                        350 mm
Physio
  1st Signal/Mode                 None
  ...............................................
  Dark blood                      Off
Inline
  Subtract                        Off
  Std-Dev-Sag                     Off
  Std-Dev-Cor                     Off
  Std-Dev-Tra                     Off
  Std-Dev-Time                    Off
  MIP-Sag                         Off
  MIP-Cor                         Off
  MIP-Tra                         Off
  MIP-Time                        Off
  Save original images            On
Sequence
  Introduction                    On
  Dimension                       3D
  Elliptical scanning             Off
  Asymmetric echo                 Off
  Contrasts                       5
                              5/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Bandwidth 1                     651 Hz/Px
  Bandwidth 2                     651 Hz/Px
  Bandwidth 3                     651 Hz/Px
  Bandwidth 4                     651 Hz/Px
  Bandwidth 5                     651 Hz/Px
  Flow comp. 1                    No
  Flow comp. 2                    No
  Flow comp. 3                    No
  Flow comp. 4                    No
  Flow comp. 5                    No
  Echo spacing                    12.2 ms
  ...............................................
  RF pulse type                   Fast
  Gradient mode                   Fast
  Excitation                      Non-sel.
  RF spoiling                     On
  ...............................................
  Readout polarity                Positive
  Readout trajectory              Bipolar
  Add. scale factor               4.0
  Gradient spoiling               Integral
  Gradient moment factor          3.0
  Siemens reconstruction          On
  Save raw k-space data           Off
  Averaging                       RMS
```

GATE_V01_R01

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 3:48    PAT: Off    Voxel size: 3.8 ×3.8×3.5 mm    Rel. SNR: 1.00    USER: ep2d_complex

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slice group 1
    Slices                        33
    Dist. factor                  30 %
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Phase enc. dir.               A >> P
    Rotation                      0.00 deg
  Phase oversampling              0 %
  FoV read                        240 mm
  FoV phase                       100.0 %
  Slice thickness                 3.5 mm
  TR                              2000 ms
  TE                              29 ms
  Averages                        1
  Concatenations                  1
  Filter                          None
  Coil elements                   HEA;HEP
Contrast
  MTC                             Off
  Flip angle                      75 deg
  Fat suppr.                      Fat sat.
  ...............................................
  Averaging mode                  Long term
  Reconstruction                  Magnitude
  Measurements                    112
  Delay in TR                     0 ms
  Multiple series                 Off
Resolution
  Base resolution                 64
  Phase resolution                100 %
  Phase partial Fourier           Off
  Interpolation                   Off
  ...............................................
  PAT mode                        None
  Matrix Coil Mode                Triple
  ...............................................
  Distortion Corr.                Off
  Prescan Normalize               Off
  Raw filter                      On
  Elliptical filter               Off
  Hamming                         Off
Geometry
  Multi-slice mode                Interleaved
  Series                          Ascending
  ...............................................
  Special sat.                    None
System

  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                FIX
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Standard
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Rotation                      0.00 deg
    R >> L                        240 mm
    A >> P                        240 mm
    F >> H                        150 mm
Physio
  1st Signal/Mode                 None
BOLD
  GLM Statistics                  Off
  Dynamic t-maps                  Off
  Starting ignore meas            0
  Ignore after transition         0
  Model transition states         Off
  Temp. highpass filter           Off
  Threshold                       4.00
  Paradigm size                   20
  Meas[1]                         Baseline
  Meas[2]                         Baseline
  Meas[3]                         Baseline
  Meas[4]                         Baseline
  Meas[5]                         Baseline
  Meas[6]                         Baseline
  Meas[7]                         Baseline
  Meas[8]                         Baseline
  Meas[9]                         Baseline
  Meas[10]                        Baseline
  Meas[11]                        Active
  Meas[12]                        Active
  Meas[13]                        Active
  Meas[14]                        Active
  Meas[15]                        Active
  Meas[16]                        Active
  Meas[17]                        Active
  Meas[18]                        Active
  Meas[19]                        Active
  Meas[20]                        Active
  Motion correction               Off
  Spatial filter                  Off
Sequence
  Introduction                    Off
  Bandwidth                       2170 Hz/Px
                              17/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Free echo spacing               Off
  Echo spacing                    0.53 ms
  ...............................................
  EPI factor                      64
  RF pulse type                   Normal
  Gradient mode                   Fast
```

GATE_V01_R02

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 3:48    PAT: Off    Voxel size: 3.8 ×3.8×3.5 mm    Rel. SNR: 1.00    USER: ep2d_complex

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slice group 1
    Slices                        33
    Dist. factor                  30 %
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Phase enc. dir.               A >> P
    Rotation                      0.00 deg
  Phase oversampling              0 %
  FoV read                        240 mm
  FoV phase                       100.0 %
  Slice thickness                 3.5 mm
  TR                              2000 ms
  TE                              29 ms
  Averages                        1
  Concatenations                  1
  Filter                          None
  Coil elements                   HEA;HEP
Contrast
  MTC                             Off
  Flip angle                      75 deg
  Fat suppr.                      Fat sat.
  ...............................................
  Averaging mode                  Long term
  Reconstruction                  Magnitude
  Measurements                    112
  Delay in TR                     0 ms
  Multiple series                 Off
Resolution
  Base resolution                 64
  Phase resolution                100 %
  Phase partial Fourier           Off
  Interpolation                   Off
  ...............................................
  PAT mode                        None
  Matrix Coil Mode                Triple
  ...............................................
  Distortion Corr.                Off
  Prescan Normalize               Off
  Raw filter                      On
  Elliptical filter               Off
  Hamming                         Off
Geometry
  Multi-slice mode                Interleaved
  Series                          Ascending
  ...............................................
  Special sat.                    None
System

  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                FIX
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Standard
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Rotation                      0.00 deg
    R >> L                        240 mm
    A >> P                        240 mm
    F >> H                        150 mm
Physio
  1st Signal/Mode                 None
BOLD
  GLM Statistics                  Off
  Dynamic t-maps                  Off
  Starting ignore meas            0
  Ignore after transition         0
  Model transition states         Off
  Temp. highpass filter           Off
  Threshold                       4.00
  Paradigm size                   20
  Meas[1]                         Baseline
  Meas[2]                         Baseline
  Meas[3]                         Baseline
  Meas[4]                         Baseline
  Meas[5]                         Baseline
  Meas[6]                         Baseline
  Meas[7]                         Baseline
  Meas[8]                         Baseline
  Meas[9]                         Baseline
  Meas[10]                        Baseline
  Meas[11]                        Active
  Meas[12]                        Active
  Meas[13]                        Active
  Meas[14]                        Active
  Meas[15]                        Active
  Meas[16]                        Active
  Meas[17]                        Active
  Meas[18]                        Active
  Meas[19]                        Active
  Meas[20]                        Active
  Motion correction               Off
  Spatial filter                  Off
Sequence
  Introduction                    Off
  Bandwidth                       2170 Hz/Px
                              19/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Free echo spacing               Off
  Echo spacing                    0.53 ms
  ...............................................
  EPI factor                      64
  RF pulse type                   Normal
  Gradient mode                   Fast
```

GATE_V01_R03

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 3:48    PAT: Off    Voxel size: 3.8 ×3.8×3.5 mm    Rel. SNR: 1.00    USER: ep2d_complex

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slice group 1
    Slices                        33
    Dist. factor                  30 %
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Phase enc. dir.               A >> P
    Rotation                      0.00 deg
  Phase oversampling              0 %
  FoV read                        240 mm
  FoV phase                       100.0 %
  Slice thickness                 3.5 mm
  TR                              2000 ms
  TE                              29 ms
  Averages                        1
  Concatenations                  1
  Filter                          None
  Coil elements                   HEA;HEP
Contrast
  MTC                             Off
  Flip angle                      75 deg
  Fat suppr.                      Fat sat.
  ...............................................
  Averaging mode                  Long term
  Reconstruction                  Magnitude
  Measurements                    112
  Delay in TR                     0 ms
  Multiple series                 Off
Resolution
  Base resolution                 64
  Phase resolution                100 %
  Phase partial Fourier           Off
  Interpolation                   Off
  ...............................................
  PAT mode                        None
  Matrix Coil Mode                Triple
  ...............................................
  Distortion Corr.                Off
  Prescan Normalize               Off
  Raw filter                      On
  Elliptical filter               Off
  Hamming                         Off
Geometry
  Multi-slice mode                Interleaved
  Series                          Ascending
  ...............................................
  Special sat.                    None
System

  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                FIX
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Standard
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Rotation                      0.00 deg
    R >> L                        240 mm
    A >> P                        240 mm
    F >> H                        150 mm
Physio
  1st Signal/Mode                 None
BOLD
  GLM Statistics                  Off
  Dynamic t-maps                  Off
  Starting ignore meas            0
  Ignore after transition         0
  Model transition states         Off
  Temp. highpass filter           Off
  Threshold                       4.00
  Paradigm size                   20
  Meas[1]                         Baseline
  Meas[2]                         Baseline
  Meas[3]                         Baseline
  Meas[4]                         Baseline
  Meas[5]                         Baseline
  Meas[6]                         Baseline
  Meas[7]                         Baseline
  Meas[8]                         Baseline
  Meas[9]                         Baseline
  Meas[10]                        Baseline
  Meas[11]                        Active
  Meas[12]                        Active
  Meas[13]                        Active
  Meas[14]                        Active
  Meas[15]                        Active
  Meas[16]                        Active
  Meas[17]                        Active
  Meas[18]                        Active
  Meas[19]                        Active
  Meas[20]                        Active
  Motion correction               Off
  Spatial filter                  Off
Sequence
  Introduction                    Off
  Bandwidth                       2170 Hz/Px
                              21/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Free echo spacing               Off
  Echo spacing                    0.53 ms
  ...............................................
  EPI factor                      64
  RF pulse type                   Normal
  Gradient mode                   Fast
```

RSTpre_V01_R01

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 5:04    PAT: Off    Voxel size: 3.8 ×3.8×3.5 mm    Rel. SNR: 1.00    USER: ep2d_complex

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slice group 1
    Slices                        33
    Dist. factor                  30 %
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Phase enc. dir.               A >> P
    Rotation                      0.00 deg
  Phase oversampling              0 %
  FoV read                        240 mm
  FoV phase                       100.0 %
  Slice thickness                 3.5 mm
  TR                              2000 ms
  TE                              29 ms
  Averages                        1
  Concatenations                  1
  Filter                          None
  Coil elements                   HEA;HEP
Contrast
  MTC                             Off
  Flip angle                      75 deg
  Fat suppr.                      Fat sat.
  ...............................................
  Averaging mode                  Long term
  Reconstruction                  Magnitude
  Measurements                    150
  Delay in TR                     0 ms
  Multiple series                 Off
Resolution
  Base resolution                 64
  Phase resolution                100 %
  Phase partial Fourier           Off
  Interpolation                   Off
  ...............................................
  PAT mode                        None
  Matrix Coil Mode                Triple
  ...............................................
  Distortion Corr.                Off
  Prescan Normalize               Off
  Raw filter                      On
  Elliptical filter               Off
  Hamming                         Off
Geometry
  Multi-slice mode                Interleaved
  Series                          Ascending
  ...............................................
  Special sat.                    None
System

  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                FIX
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Standard
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      L1.9 A8.8 H38.5
    Orientation                   T > C-22.5 > S3.4
    Rotation                      0.00 deg
    R >> L                        240 mm
    A >> P                        240 mm
    F >> H                        150 mm
Physio
  1st Signal/Mode                 None
BOLD
  GLM Statistics                  Off
  Dynamic t-maps                  Off
  Starting ignore meas            0
  Ignore after transition         0
  Model transition states         Off
  Temp. highpass filter           Off
  Threshold                       4.00
  Paradigm size                   20
  Meas[1]                         Baseline
  Meas[2]                         Baseline
  Meas[3]                         Baseline
  Meas[4]                         Baseline
  Meas[5]                         Baseline
  Meas[6]                         Baseline
  Meas[7]                         Baseline
  Meas[8]                         Baseline
  Meas[9]                         Baseline
  Meas[10]                        Baseline
  Meas[11]                        Active
  Meas[12]                        Active
  Meas[13]                        Active
  Meas[14]                        Active
  Meas[15]                        Active
  Meas[16]                        Active
  Meas[17]                        Active
  Meas[18]                        Active
  Meas[19]                        Active
  Meas[20]                        Active
  Motion correction               Off
  Spatial filter                  Off
Sequence
  Introduction                    Off
  Bandwidth                       2170 Hz/Px
                              23/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Free echo spacing               Off
  Echo spacing                    0.53 ms
  ...............................................
  EPI factor                      64
  RF pulse type                   Normal
  Gradient mode                   Fast
```

t2_tse_tra_192

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 3:08 PAT: 2 Voxel size: 1.1×1.1×1.5 mm Rel. SNR: 1.00 SIEMENS: tse

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments
  Auto open inline display        Off
  Start measurement without       On
  further preparation
  Wait for user to start          Off
  Start measurements              single
Routine
  Slice group 1
    Slices                        120
    Dist. factor                  0 %
    Position                      L0.9 A32.1 H4.9
    Orientation                   T > C-4.3 > S1.7
    Phase enc. dir.               R >> L
    Rotation                      90.00 deg
  Phase oversampling              12 %
  FoV read                        220 mm
  FoV phase                       100.0 %
  Slice thickness                 1.5 mm
  TR                              15500 ms
  TE                              77 ms
  Averages                        1
  Concatenations                  1
  Filter                          Prescan Normalize, Elliptical
                                  filter
  Coil elements                   HEA;HEP
Contrast
  MTC                             Off
  Magn. preparation               None
  Flip angle                      155 deg
  Fat suppr.                      None
  Water suppr.                    None
  Restore magn.                   Off
  ...............................................
  Averaging mode                  Short term
  Reconstruction                  Magnitude
  Measurements                    1
  Multiple series                 Each measurement
Resolution
  Base resolution                 192
  Phase resolution                100 %
  Phase partial Fourier           Off
  Trajectory                      Cartesian
  Interpolation                   Off
  ...............................................
  PAT mode                        GRAPPA
  Accel. factor PE                2
  Ref. lines PE                   26
  Matrix Coil Mode                Auto (Triple)
  Reference scan mode             Integrated
  ...............................................
  Image Filter                    Off
  Distortion Corr.                Off
  Unfiltered images               Off
  Prescan Normalize               On

  Normalize                       Off
  B1 filter                       Off
  Raw filter                      Off
  Elliptical filter               On
  Mode                            Inplane
Geometry
  Multi-slice mode                Interleaved
  Series                          Interleaved
  ...............................................
  Special sat.                    None
  ...............................................
  ...............................................
  Tim CT mode                     Off
System
  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                REF
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
  Sagittal                        R >> L
  Coronal                         A >> P
  Transversal                     F >> H
  Save uncombined                 Off
  Coil Combine Mode               Adaptive Combine
  Auto Coil Select                Default
  ...............................................
  Shim mode                       Standard
  Adjust with body coil           Off
  Confirm freq. adjustment        Off
  Assume Silicone                 Off
  ? Ref. amplitude 1H             0.000 V
  Adjustment Tolerance            Auto
  Adjust volume
    Position                      L0.9 A32.1 H4.9
    Orientation                   T > C-4.3 > S1.7
    Rotation                      90.00 deg
    A >> P                        220 mm
    R >> L                        220 mm
    F >> H                        180 mm
Physio
  1st Signal/Mode                 None
  ...............................................
  Dark blood                      Off
  ...............................................
  Resp. control                   Off
Inline
  Subtract                        Off
  Std-Dev-Sag                     Off
  Std-Dev-Cor                     Off
  Std-Dev-Tra                     Off
  Std-Dev-Time                    Off
  MIP-Sag                         Off
  MIP-Cor                         Off
  MIP-Tra                         Off
  MIP-Time                        Off
  Save original images            On
Sequence
  Introduction                    On
  Dimension                       2D
  Compensate T2 decay             Off
  Reduce Motion Sens.             Off
                              7/+
SIEMENS MAGNETOM TrioTim syngo MR B17

  Contrasts                       1
  Bandwidth                       199 Hz/Px
  Flow comp.                      No
  Allowed delay                   60 s
  Echo spacing                    11 ms
  ...............................................
  Define                          Turbo factor
  Turbo factor                    11
  Echo trains per slice           11
  RF pulse type                   Fast
  Gradient mode                   Normal
```

dti_35dir

```
SIEMENS MAGNETOM TrioTim syngo MR B17

TA: 5:42 PAT: 2 Voxel size: 2.0×2.0×2.0 mm Rel. SNR: 1.00 SIEMENS: ep2d_diff

Properties
  Prio Recon                      Off
  Before measurement
  After measurement
  Load to viewer                  On
  Inline movie                    Off
  Auto store images               On
  Load to stamp segments          Off
  Load images to graphic          Off
  segments

  Special sat.                    None
  ...............................................
System
  Body                            Off
  HEP                             On
  HEA                             On
  ...............................................
  Positioning mode                FIX
  Table position                  H
  Table position                  0 mm
  MSMA                            S - C - T
```