SYNTHETIC DEMO DOCUMENT – FICTIONAL DATA – NO REAL PATIENTS, PEOPLE, INSTITUTIONS OR PRODUCTS

Demo Orthopaedic Centre Westshire (fictional), Department of Trauma and Orthopaedics
Westshire, England – 10 November 2026
Our ref: DOCW/TO/2026/118

To: Fictivia Medical Ltd. (fictional), Vigilance Department

Re: Femoral stem fracture followed by death – patient P-GB-2208

Dear Vigilance Team,

I am writing about our patient (pseudonym P-GB-2208, 67 years), who underwent a primary total hip replacement in England on 19 August 2025 with a Fictivia OA-S5 femoral stem (UDI-DI (01)00499002100052, LOT OA-S5-2506-C) and a ceramic head.

On 30 October 2026 the patient presented with sudden thigh pain and inability to bear weight after getting up from a chair; there was no fall. Radiographs showed a fracture of the stem at the neck–body junction. Revision surgery was performed on 2 November 2026. On 6 November 2026 the patient collapsed on the ward and died. A post-mortem examination confirmed a massive pulmonary embolism.

The explanted stem has been retained in quarantine and can be made available for analysis. The patient's weight was 104 kg (BMI 34). No previous problems with the implant had been reported.

Yours sincerely,
Consultant Orthopaedic Surgeon (fictional role, no name used in the demo)