SYNTHETIC DEMO DOCUMENT – FICTIONAL DATA – NO REAL PATIENTS, PEOPLE, INSTITUTIONS OR PRODUCTS

Demo Regional Orthopaedic Service (fictional), Northern Ireland
11 November 2026

To: Fictivia Medical Ltd. (fictional), Vigilance Department

Re: Liner dissociation – patient P-NI-3390

Dear Colleagues,

Our patient (pseudonym P-NI-3390, 61 years) underwent a primary total hip replacement in our unit in Northern Ireland on 2 April 2026. A Fictivia OA-L36 polyethylene liner (LOT OA-L36-2603-B) was used together with an OA-C shell.

On 4 November 2026 the patient attended the emergency department with a sudden painful clunk in the hip and was unable to walk. Radiographs showed an eccentric position of the femoral head. The patient was admitted and underwent revision surgery on 6 November 2026. Intraoperatively, the polyethylene liner had dissociated from the acetabular shell; the locking ring was deformed. The liner was exchanged and the patient is recovering well.

There was no trauma and the components had been correctly seated according to the operation note. We have not seen this before with this system. The explanted liner is available for return.

Kind regards,
Consultant Orthopaedic Surgeon (fictional role, no name used in the demo)