SYNTHETIC DEMO DOCUMENT – FICTIONAL DATA – NO REAL PATIENTS, PEOPLE, INSTITUTIONS OR PRODUCTS

Demo Atlantic University Hospital (fictional), Department of Cardiothoracic Surgery
Ireland – 13 November 2026

To: Fictivia Medical Ltd. (fictional), Vigilance Department

Re: Early leaflet tears in three patients – same lot of Fictivia CB-21

Dear Vigilance Team,

We wish to inform you of three patients who received a Fictivia CB-21 aortic bioprosthesis from the same lot (LOT CB21-2607-A) at our hospital between 18 August and 22 September 2026:

- P-IE-1102, implanted 18 August 2026
- P-IE-1107, implanted 1 September 2026
- P-IE-1115, implanted 22 September 2026

Between 3 and 11 November 2026, all three patients developed acute severe aortic regurgitation caused by a torn leaflet, only six to ten weeks after implantation. All three required urgent redo surgery. P-IE-1107 needed intensive care treatment for five days after the redo operation; all three patients are currently stable.

We have stopped using the remaining valves from this lot and placed them in quarantine. Two further patients with valves from the same lot have been recalled for echocardiography. In our view this cluster may represent a risk to other patients.

Yours sincerely,
Consultant Cardiothoracic Surgeon (fictional role, no name used in the demo)