GOC – Without IRDC

Equipment Room

Main Disconnect Panel — CB3, TB1-15-16, CB2

Remote MRU

E-Off Switch

O_2 Monitor Control Unit — Meter, Ctrl Board, Alarm, Light

Customer-Supplied Outlet

Site E-Net

Facility Power Filters

Facility Utilities — Facility Coolant Supply, Facility Coolant Return

Mag Mon — J5, J10, J8, J9, J7

HEC

- **HEC Power Box** — Cryo, Power Supply, GC, PE, Blower
- **Signal Box** — J24, J11, J12
- Grad HtXchr
- Grad Res
- Grad Pump
- Flow Meter, Filter
- Blower HtXchgr
- **PGR Pump**, **PGR Res**, PGR HtXchr
- **Grad/RF Coil Blower**

JR Power **Cryo-Cooler Compressor** JC — Water Return, Water Supply, Gas Return, Gas Supply

Operator Workspace

Temp/Hum Sensor — FJ7, E3018, E3019, E3011

GOC

- **Host PC** — Ethernet B Lwr Graphic Card Port 1; RS232; USB; USB; USB; USB; Upr Graphic Card Port 1; Ethernet A; Line In; Line Out
- **USB-Serial Converter**
- Server (Option) — Site Provided
- **Ethernet Switch** 1, 2
- Service Laptop — Site Provided
- **GOCAA** — J12, J14, J17; J18 Pin3, Pin8; J15, J7; Pin1, Pin2, J6
- Optional Speaker — Site Provided
- J10, J4, J6, J9, J2, GOC, J5, J1, PDU, J3

FJ6, E3016; FJ5, E3016; E3017; E3031; E3047

SCIM Spkr, Mic

Global Modem (Option) — E3045

Key-board — E3048

Mouse

Primary Display — E3046

Pneumatic Patient Alert System

Scan Room

PGR

- **PDM** — L1/L2/L3; J29, J25; J27 J28 J24
- **VRE 2** (32-ch) — J5, Net 3, Net 2, Net Mgmt, J1
- **VRE 1** — J1, Net Mgmt, Net 2, Net 3, J5
- **Term Server** — ETH 1, 2, 3, 4
- **Aux Board** — J4; J6; J1, J2/J3, J9, J11, J10
- XRFD; LS1-A, LS1-B, LS2-A, LS2-B, LS3-A, LS3-B, LS4-A, Term; PDU Top 1, PDU Top 1, PDU Front
- J4 XGD Fan x3
- CB1, J1, J4, J3; CB3; RECON; RF/MNS, J5, J2, PEN; **PGR PDU**; Host; J6/J7/J8, J11/J12/J13
- J3, J6, J2, J102, J19, J18, J17, E3015
- J14, J15, J13, J12, J4
- CB2 J1 J2; **NB Amp**; J4, J3; J8, J7, J5, J6 J19 J18; J14, J13, J10, J9; J3 J2; Service Panel; J21, J22, J11, J1 HSS Test
- **16 Port E-net Sw** — 3, 2, 1, 4, 5, 6, 15, 12, 11, 7; 14, 13
- **Cronus Chassis** (Cronus Rack Control Node, Workstation only)
  - XPS CB — J13/J16, J12/J15, J11/J14; P1
  - XGA CB — J13/J16, J12/J15, J11/J14
  - J10/J11, J6, J9, **GMC**, J8, J7; J1, J2
  - J3, J4, J5
- Pwr **XPS x3** — J2, J10, J1, J11, J8, J9, J3, J4/J5, J4/J5
- **XGA x3** — J8, J7, J16, J13, J15, J14, J5, J12, J6, J17, J18, J1/J2, J1/J2
- Manifold Supply — SMA Coax 9-pin; LEM Icoil x3 — Grad Pulse In, Grad Pulse Out
- Term Block x3 (+ / −)
- Manifold Return
- J2 **CAM Chassis** — UPM1 Det J1, J2, Proc J3; UPM2 Det J1, J2, Proc J3; ETH SCP3 PMC; 10/100 Base AGP2 J3; J3 NB AIF J8; STIF J4 J3, J18/J19; J23 I/O Panel J10, IRF3 J8, J14 J13 J11
- J8, J20, J22 J21 J11 J1, J5 J31 J16, J103

SPW

- J115, J116, J117, J106, J16
- J118, J17
- J100 Wvgde, J101 Wvgde, J111 Wvgde, J110 Wvgde, Patient Alert Wvgde
- Term Conn — F1, F2, F3, F4
- X Y Z **Gradient Filter** X Y Z
- J128
- J121 J120 **Xmit Switch** J123, J122
- J104 Wvgde
- J105 Wvgde

Primary MRU

E-Off Switch

O_2 Monitor — Sensor, Alarm, Light

Mag Enclosure Chimney

Gradient Bus Bar — X, Y, Z

Mag I/O — M3031; J1, J3, J6, J4, J2, J5, J12, J15, J14, J11, J13, J8, J7, J18, J19, J9, J10, J17, J16

FO I/F — J28; J20, J21

Mag Bore

- J8, J6, J9, J7
- **Body Coil** J2, J3, J9
- Bore-PAL L, R
- Left Arc Cover — Ambient Temp Snsr
- Bore Temp Sensors
- Front Mic — M3345
- Airflow Sensors
- X Y Z **XRMB Gradient** Temp Sensor
- J101 MAG-Grad Coil, J100 MAG-Grad Coil
- J104 Grad Coil Air
- Patient Air

Magnet

- J3 P403 Terminal Box
- Pressure Transducer
- Reed Switch
- J2 RUO PreAmp J1
- Coldhead Pwr Sleeve; Coldhead Supply Return
- Airflow Sensor Module
- **SRI4** — J23, J10, J34, J3, J20, J32, J24, J1, J4, J5, J33, J11/J12, J13/J14, J21, J31, J2, J22, J17; M3338, M3339, M3346
- **Pre-Amp** — Output Input; J1, J4, J2, J3, J6; **BHYB**; M3348, M3347
- J3, J5, J2, J4, J1 **7SD**; M3349
- J9/J10 J2 **PAC**
- J11, J2, J12 **TDM** J1, J9, J10
- J9 **RRx1** J7, J1, J4/J5 J8 J2
- J7 **RRx2** (32-ch) J1, J4/J5 J2

MAG-PAC J4

PAC I/F

Squeeze Bulb

Speaker

Left E-Stop Button

Rt E-Stop Button

Patient Table

- Coil Port P3
- Coil Port P4
- Touch N Go_R
- Touch N Go_L
- Table Hydraulics

Dock

- J2 Mag
- J1 Dock
- J3 Left
- ODU

Rear Pedestal

- PED Antenna
- Lights
- RF Hub — J102 J101, J104, J103; Hub Backplane (bias, power, status/control)
- J1, J3, J2, J4 PDB, J5, J6
- MCDB
- J9, J13, J12 Debug Port; J16, J17, J10 RFCB J8; J14/J15; J1 J3, J7 J6 J2 J4
- J1–J6 (ch 17-32); RFSB (ch 1-16) J1–J6
- J4 J6, J3 TIF, J1 J2 J5
- POS Enc
- **Long Drive** J1 **ASM** (MTR) J2
- **Load A**
- J3 EOT Sensor, HOME Sensor, EGRESS Sensor **LPCA**
- J4 J2, J5 Port A J1; Port P2 J4; Port P1 J5; Rear Mic
- J2 PED I/O J1
- **Reroute Box** — J10 J7; P2A P3A, P2B P3B, P1A P4A, P1B P4B; J8 J9
- **Load B**
- Rear Pedestal Air

Door Switch

Coolant Removal Kit

PEN

- J14, J1, J13, J11, J12; Pin2, J10, Pin8; F/O Wvgde
- J19, J15, J18, J17, J3, J9, J5, J6
- J18 J7, J23, J12, J22 **Exciter**, J19, J20, J21, J29, J32, J30; J15 J17 J1
- J1 **Power Distribution** J4 J2 J3
- J18, J19, J7/J8, J1/J2, J5/J6, J3/J4 **CFB** J17
- J14 J19 J17, J11, J12 **Driver Module** J2, J7, J6, J1, J25, J24, J23 J21 J22
- J94 J96 J92 J95 J93, J76 J75, **DD Filter**, J74 J72 J73
- J12, Pin2, J9, Pin8, J37/J38, J2, J43, J11, J39, J54, J41, J1, J5, J40, J10, **PHPS** J42, J44, J35, J36, J8
- J6, J5, J1, **SRPS** J51, J52, J46, J48, J45, J50, J47, J3
- J39, J54, J41, J40, J42, J35, J36, J51, J52, J46, J48, J45, J50, J47, J29, J32, J30, J34, F11, F10, F9, F8, F7, F6, F5

Temp/Hum Sensor

Scan Room Blwr Box — Patient Comfort, Hub

Key — TBD = To Be Determined (Future Enhancement)

Cabinet; Chassis/ Sub-system; Part; Data Cables, Coolant, Power, Pulse, Data Paths

E-Stop/E-Off Circuitry

E-Stop Circuit; E-Off Circuit; PDU Power Off

© 2009, GE Healthcare