GOC – Main

Equipment Room

Main Disconnect Panel — CB3, TB1-15-16, CB2

Remote MRU

E-Off Switch

O_2 Monitor Control Unit — Meter, Ctrl Board, Alarm, Light

Customer-Supplied Outlet

Site E-Net

Facility Power Filters

Facility Utilities — Facility Coolant Supply, Facility Coolant Return

Operator Workspace

Temp/Hum Sensor

Server (Option) — Site Provided

Service Laptop — Site Provided

Optional Speaker — Site Provided

GOC

IRDC (Option) — J5, J4, J3, J2, J1

USB-Serial Converter

Ethernet Switch — 1, 2

GOCAA — J12, J14, J17, J18 (Pin3, Pin8), J15, J7, J6 (Pin1, Pin2)

Host PC — Ethernet B, Lwr Graphic Card Port 1, RS232, USB, USB, USB, USB, Upr Graphic Card Port 1, Ethernet A, Line In, Line Out

J10, J4, J6, J9, J2, GOC, J5, J1, PDU, J3

Global Modem (Option)

Key-board

Mouse

Primary Display

SCIM Spkr, Mic

Pneumatic Patient Alert System

PGR

PDM — L1/L2/L3, J29, J25, J27 J28 J24

VRE 2 (32-ch) — J5, Net 3, Net 2, Net Mgmt, J1

VRE 1 — J1, Net Mgmt, Net 2, Net 3, J5

Term Server — ETH 1, 2, 3, 4

Aux Board — J4, J6, J1, J2/J3, J9, J11, J10

XFRD, LS1-A, LS1-B, LS2-A, LS2-B, LS3-A, LS3-B, LS4-A, Term; PDU Top 1, PDU Top 1, PDU Front

J4 XGD Fan x3

CB1, CB3, RECON, J1, J4, J3, **PGR PDU**, Host, RF/MNS, J5, J2, PEN, J6/J7/J8, J11/J12/J13

J3, J6, J2, J102, J19, J18, J17, E3015

J14, J15

CB2 J1 J2, Service Panel, **NB Amp**, J4, J3, J14, J13, J10, J9, J8, J7, J5, J6 J19 J18, J21, J22, J11, J1 HSS Test, J3 J2

J13, J12

16 Port E-net Sw — 3, 2, 1, 4, 5, 6, 15, 12, 11, 7, 13, 14

J4

J2

CAM Chassis

UPM1 — Det J1 J2, Proc J3

UPM2 — Det J1 J2, Proc J3

ETH, SCP3, PMC

10/100 Base, AGP2, J3

J3 NB AIF J8

STIF — J4 J3, J18/J19

J23 I/O Panel, J10, IRF3, J8, J14 J13 J11

J8

J20, J22 J21 J11 J1, J5 J31 J16, J103

Cronus Chassis

P1

XPS CB — J13/J16, J12/J15, J11/J14

XGA CB — J13/J16, J12/J15, J11/J14

Cronus Backplane (Multi-N, Multi-Protocol only)

J3, J4, J5

J10/J11, J6, J9 **GMC**, J8, J7, J1, J2

Pwr **XPS x3** — J2, J10, J1, J11, J8, J9, J3, J4/J5, J4/J5

XGA x3 — J8, J7, J16, J13, J15, J14, J5, J12, J6, J17, J18, J1/J2, J1/J2

Manifold Supply

SMA Coax 9-pin — LEM Icoil x3 — Grad Pulse In, Grad Pulse Out

Term Block x3

Manifold Return

SPW

J115, J116, J117, J106, J16

Term Conn — F1, F2, F3, F4

X Y Z **Gradient Filter** X Y Z

J128

J121 J120 **Xmit Switch** J123 J122

J104 Wvgde

J105 Wvgde

P2026

J118 J17

J100 Wvgde, J101 Wvgde, J111 Wvgde, J110 Wvgde, Patient Alert Wvgde

Primary MRU

E-Off Switch

O_2 Monitor — Sensor, Alarm, Light

Mag Enclosure Chimney

Gradient Bus Bar — X, Y, Z

Mag I/O — M3031 — J1, J3, J6, J4, J2, J5, J12, J15, J14, J11, J13, J8, J7, J18, J19, J9, J10, J17, J16

FO I/F — P2026 — J28, J20, J21

Mag Bore

J8, J6, J9, J7

Body Coil — J2, J3, J9

Bore-PAL L R

Left Arc Cover — Ambient Temp Snsr

Bore Temp Sensors

Front Mic — M3345

Airflow Sensors

X Y Z **XRMB Gradient** Temp Sensor

J101 MAG-Grad Coil, J100 MAG-Grad Coil

J104 Grad Coil Air

Patient Air

Scan Room

Magnet

J3 P403 Terminal Box

Pressure Transducer

Reed Switch

J2 RUO PreAmp J1

Coldhead Pwr Sleeve, Coldhead Supply, Coldhead Return

Airflow Sensor Module — M3339

SRI4 — J23, J10, J34, J3, J20, J32, J24, J1, J4, J5, J33, J11/J12, J13/J14, J21, J31, J2, J22, J17

Pre-Amp — Output Input

J1, J4, J2, J3, J6 **BHYB**

M3348, M3347

LCP J11 J3

RCP J11 J3

J9/J10 J2 **PAC**

M3346

IRD J1, P2026

J11 J2, J12 **TDM** J1, J9, J10

J9 **RRx1** J7, J1, J4/J5 J8 J2

J7 **RRx2** (32-ch) J1, J4/J5 J2

MAG-PAC J4

PAC I/F

Squeeze Bulb

Speaker

Left E-Stop Button

Rt E-Stop Button

Patient Table

Coil Port P3

Coil Port P4

Touch N Go_R

Touch N Go_L

Table Hydraulics

Dock

J2 Mag, J1 Dock, J3 Left

ODU

Rear Pedestal

PED Antenna

Lights

RF Hub — Hub Backplane (bias, power, status/control) — J102 J101, J104, J103

J1 J3, J2, J4 PDB, J5, J6

MCDB

J9 J13 J12, J16, J17 RFCB, J10, J8, J14/J15, J1 J3, J7 J6 J2 J4

Debug Port

J1–J6 (ch 17-32)

RFSB (ch 1-16) J1–J6

J4 J6, J3 TIF, J1 J2 J5

POS Enc

Long Drive J1 **ASM (MTR)** J2

Load A

J3 EOT Sensor, HOME Sensor, EGRESS Sensor **LPCA**

J4 J2, J5 Port A J1, Port P2 J4, Port P1 J5

Rear Mic

J2 PED I/O J1

Reroute Box — J10 J7, P2A P3A, P2B P3B, P1A P4A, P1B P4B, J8 J9

Load B

Rear Pedestal Air

PEN

J14, J1, J13, J11, J12, J10 (Pin2, Pin8), F/O Wvgde

J19, J15, J18, J17, J3, J9, J5, J6

J18 J7, J23, J12, J22 **Exciter**, J19, J20, J21, J29, J32, J30, J15 J17 J1

J1 **Power Distribution** J4 J2 J3

Temp/Hum Sensor

J18, J19, J7/J8, J1/J2, J5/J6, J3/J4 **CFB** J17

J14 J19 J17, J11, J12 **Driver Module** J2, J7, J6, J1, J25, J24, J23 J21 J22

J94 J96 J92 J95 J93, J76 J75, **DD Filter**, J74 J72 J73

J12, J9 (Pin2, Pin8), J37/J38, J2, J43, J11, J39, J54, J41, J1, J5, J10, J40, **PHPS**, J42, J44, J35, J36, J8

J6, J5, J1, J51, J52, J46, J48, J45, J50, J47, **SRPS**, J3

J39, J54, J41, J40, J42, J35, J36, J51, J52, J46, J48, J45, J50, J47, J29, J32, J30, J34, F11, F10, F9, F8, F7, F6, F5

Scan Room Blwr Box — Patient Comfort — Hub

Door Switch

Coolant Removal Kit

Mag Mon

J5, J10, J8, J9, J7

HEC

HEC Power Box — Cryo, Power Supply, GC, PE, Blower

Signal Box — J24, J11 J12

Grad HtXchr

Grad Res

Grad Pump

Blower HtXchgr

Flow Meter

Filter

PGR Pump

PGR Res

PGR HtXchr

Grad/RF Coil Blower

JR Power **Cryo-Cooler Compressor** JC — Water Return, Water Supply, Gas Return, Gas Supply

E3011, E3045, E3019, E3018, FJ7, P2026, E3048, E3046, FJ6, E3016, FJ5, E3017, E3047, E3031

© 2009, GE Healthcare