Gradient – Pulse

Equipment Room

Main Disconnect Panel — CB3, TB1-15-16, CB2

Remote MRU

O_2 Monitor Control Unit — Meter, Ctrl Board, Alarm, Light

Customer-Supplied Outlet

Site E-Net

Facility Power Filters

Facility Utilities — Facility Coolant Supply, Facility Coolant Return

E-Off Switch

Operator Workspace

Temp/Hum Sensor

Server (Option)

Service Laptop

Optional Speaker

GOC

IRDC (Option) — J5, J4, J3, J2, J1

USB-Serial Converter

Ethernet Switch — 1, 2

GOCAA — J12, J14, J17, J18, J15, J7, J6

Host PC — Ethernet B, Lwr Graphic Card Port 1, RS232, USB, USB, USB, USB, Upr Graphic Card Port 1, Ethernet A, Line In, Line Out

J10, J4, J6, J9, J2, GOC, J5, J1, PDU, J3

Global Modem (Option)

Key-board

Mouse

Primary Display

Pneumatic Patient Alert System

SCIM Spkr, Mic

PGR

PDM — L1/L2/L3, J29, J25, J27 J28 J24

VRE 2 (32-ch) — J5, Net 3, Net 2, Net Mgmt, J1

VRE 1 — J1, Net Mgmt, Net 2, Net 3, J5

Term Server — ETH, 1, 2, 3, 4

Aux Board — J4, J6, J1, J2/J3, J9, J11, J10

XRFD, LS1-A, LS1-B, LS2-A, LS2-B, LS3-A, LS3-B, LS4-A, Term; PDU Top 1, PDU Top 1, PDU Front

J4 XGD Fan x3

CB1, CB3, RECON, J1, J4, J3, **PGR PDU**, Host, J6/J7/J8, J11/J12/J13, RF/MNS, J5, J2, PEN

J3, J6, J2, J102, J19, J18, J17

J14, J15, CB2 J1, J2, J4, J3, **NB Amp**, J8, J7, J5, J6 J19 J18, J14, J13, J10, J9, Service Panel, J21, J22, J11, J1, HSS Test, J3 J2

J13, J12

16 Port E-net Sw — 3, 2, 1, 4, 5, 6, 15, 12, 11, 14, 13 7

J4

J2, UPM1, Det J1 J2, Proc J3, **CAM Chassis**, UPM2, Det J1 J2, Proc J3, ETH, SCP3 PMC, 10/100 Base, AGP2, J3

J8, J3 NB AIF J8, STIF, J4 J3, J18/J19, J23 I/O Panel, J10, IRF3, J8, J14 J13 J11

J20, J22 J21 J11 J1, J5, J31 Unused, J16, J103

Cronus Chassis

P1, XPS CB, J13/J16, J12/J15, J11/J14

Cronus Backplane (data & clocks, fiber/custom cards)

XGA CB, J13/J16, J12/J15, J11/J14

J3, J4, J5

J10/J11, J6, J9 **GMC**, J8, J7, J1, J2

Pwr **XPS x3**, J2, J10, J1, J11, J8, J9, J3, J4/J5, J4/J5

XGA x3, J8, J7, J16, J13, J15, J14, J5, J12, J6, J17, J18, J1/J2, J1/J2

Manifold Supply

SMA Coax, 9-pin, LEM Icoil x3, Grad Pulse In, Grad Pulse Out

Term Block x3

Manifold Return

E3317, E3318, E3319

SPW

J115, J116, J117, J106, J16

Term Conn, F1, F2, F3, F4

X Y Z **Gradient Filter** X Y Z

J128

J121, J120, **Xmit Switch**, J123, J122

J104 Wvgde

J105 Wvgde

J118, J17

J100 Wvgde, J101 Wvgde, J111 Wvgde, J110 Wvgde, Patient Alert Wvgde

Primary MRU

E-Off Switch

O_2 Monitor — Sensor, Alarm, Light

Mag Enclosure Chimney

Gradient Bus Bar — M3317 X, M3318 Y, M3319 Z

Mag I/O — J1, J3, J6, J4, J2, J5, J12, J15, J14, J11, J13, J8, J7, J18, J19, J9, J10, J17, J16

FO I/F — J28, J20, J21

Mag Bore

J8, J9, J6, J7, **Body Coil**, J2, J3, J9

Bore-PAL L R

Left Arc Cover — Ambient Temp Snsr

Bore Temp Sensors

Front Mic

Airflow Sensors

X Y Z **XRMB Gradient** Temp Sensor

J101 MAG-Grad Coil, J100 MAG-Grad Coil

J104 Grad Coil Air

Patient Air

Scan Room

Magnet

J3 Terminal P403 Box

Pressure Transducer

Reed Switch

J2 RUO PreAmp J1

Coldhead Pwr Sleeve, Coldhead Supply Return

Airflow Sensor Module

SRI4 — J23, J10, J34, J3, J20, J32, J24, J1, J4, J5, J33, J11/J12, J13/J14, J21, J31, J2, J22, J17

Pre-Amp — Output Input

J1, J4, J2, J3, J6, **BHYB**

LCP J11 J3

RCP J11 J3

J9/J10 J2 **PAC**

IRD J1

J11 J2, J12 **TDM** J1, J9, J10

J9 **RRx1** J7, J1, J4/J5 J8 J2

J7 **RRx2** (32-ch) J1, J4/J5 J2

MAG-PAC J4

PAC I/F

Squeeze Bulb

Speaker

Left E-Stop Button

Rt E-Stop Button

Patient Table

Coil Port P3

Coil Port P4

Touch N Go_R

Touch N Go_L

Table Hydraulics

Dock

J2 Mag, J1 Dock, J3 Left

ODU

Rear Pedestal

PED Antenna

RF Hub — Hub Backplane (bias, power, status/control), J102J101, J104, J103

J1, J3, J2, J4 PDB, J5, J6

MCDB

J9, J13, J12, Debug Port, J16, J17 RFCB, J10, J8, J14/J15, J1, J3, J7J6, J2, J4

J1–J6 (ch 17-32) RFSB (ch 1-16) J1–J6

Lights

J4, J6, J3 TIF, J1, J2, J5

POS Enc

Long Drive ASM (MTR) J1 J2

Load A

J2 PED I/O J1

J3 EOT Sensor, HOME Sensor, EGRESS Sensor **LPCA**

J4, J2, Port A, J5, J1, Port P2 J4, Port P1 J5

Rear Mic

J10, J7, **Reroute Box**, P2A, P2B, P1A, P1B, P3A, P3B, P4A, P4B, J8, J9

Load B

Rear Pedestal Air

PEN

J14, J1, J13, J11, J12, J10, F/O Wvgde

J19, J15, J18, J17, J23, J12, J22 **Exciter**, J19, J20, J21, J29, J32, J30, J15 J17, J1

Power Distribution J4 J2 J3

J18, J19, J7/J8, J1/J2, J5/J6, J3/J4, **CFB**, J17

J12, J9, J37/J38, J2, J43, J11, J39, J54, J41, J1, J5, J10, J40, **PHPS**, J42, J44, J35, J36, J8

J6, J51, J52, J5, J46, J1, **SRPS**, J48, J45, J50, J3, J47

J14, J19, J17, J11, J20, J12 **Driver Module** J1, J2, J7, J6, J25, J24, J23, J21, J22

J94, J96, J92, J95, J93, **DD Filter**, J76, J75, J74, J72, J73

J3, J9, J5, J6

J39, J54, J41, J40, J42, J35, J36, J51, J52, J46, J48, J45, J50, J47, J29, J32, J30, J34, F11, F10, F9, F8, F7, F6, F5

Temp/Hum Sensor

Scan Room Blwr Box

Patient Comfort

Hub

HEC

Mag Mon — J5, J10, J8, J9, J7

HEC Power Box — Cryo, Power Supply, GC, PE, Blower

Signal Box — J24, J11 J12

Grad HtXchr

Grad Res

Grad Pump

Flow Meter

Filter

Blower HtXchgr

JR Power **Cryo-Cooler Compressor** JC — Water Return, Water Supply, Gas Return, Gas Supply

PGR Pump

PGR Res

PGR HtXchr

Grad/RF Coil Blower

Door Switch

Coolant Removal Kit

Key

TBD = To Be Determined (Future Enhancement)

Cabinet

Chassis/ Sub-system

Part

Data Cables, Coolant, Power, Pulse, Data Paths

© 2009, GE Healthcare