Patient Handling – Control/Display-DVMR

Equipment Room

Main Disconnect Panel — CB3, TB1-15-16, CB2

Remote MRU

O_2 Monitor Control Unit — Meter, Ctrl Board, Alarm, Light

Customer-Supplied Outlet

Site E-Net

Facility Power Filters

Facility Utilities — Facility Coolant Supply, Facility Coolant Return

E-Off Switch

PGR

PDM L1/L2/L3 — J29 J25, J27 J28 J24

VRE 2 (32-ch) — J5, Net 3, Net 2, Net Mgmt, J1

VRE 1 — J1, Net Mgmt, Net 2, Net 3, J5

ETH 1 2 3 4 Term Server

Aux Board — J4, J6, J1 J2/J3 J9 J11 J10

XRFD LS1-A LS1-B LS2-A LS2-B LS3-A LS3-B LS4-A Term — PDU Top 1, PDU Top 1, PDU Front

J4 XGD Fan x3

CB1 J1 J4 J3 — CB3 — RECON — RF/MNS J5 J2 PEN — PGR PDU — Host J6/J7/J8 J11/J12/J13

J14 J15

CB2 J1 J2 — Service Panel — J4 J3 — NB Amp — J14 J13 J10 J9 — J8 J7 J5 — J6 J19 J18 — J21 J22 J11 J1 HSS Test — J3 J2

16 Port E-net Sw — 3 2 1 4 5 15 12 11 — 14 13 7

J13 J12

J4

Cronus Chassis — P1 — XPS CB — J13/J16, J12/J15, J11/J14 — Cronus Backplane (bus, clock, trig, synchronization signals) — XGA CB — J13/J16, J12/J15, J11/J14 — J3 J4 J5 — J10/J11 J6 J9 J8 J7 GMC — J1 J2

Pwr XPS x3 — J2 J10, J1 J11, J8, J9 J3, J4/J5, J4/J5

XGA x3 — J8 J7, J16, J13 J15, J14 J5, J12 J6, J17, J18, J1/J2, J1/J2

Manifold Supply — SMA Coax 9-pin — LEM Icoil x3 — Grad Pulse In, Grad Pulse Out

Term Block x3

Manifold Return

J2 — CAM Chassis — ETH — SCP3 PMC — 10/100 Base — AGP2 J3

UPM1 — Det J1 J2, Proc J3

UPM2 — Det J1 J2, Proc J3

J8 — J3 NB AIF J8

STIF J4 J3 J18/J19

J23 I/O Panel J10 — IRF3 J8 — J14 J13 J11

J20 J22 J21 J11 J1 J5 J31 J16 J103

Door Switch

Coolant Removal Kit

PEN

J14 J1 J13 J11 J12 J10 — F/O Wvgde

J19 J15 J18 J17 — J18 J7, J23, J12, J22 Exciter, J19 J20 J21, J29 J32 J30, J15 J17 J1

J1 Power Distribution J4 J2 J3

J18 J19 J7/J8 J1/J2 J5/J6 J3/J4 CFB J17

J12 J9 J37/J38 J2 J43 J39 J54 J41 — J11 J1 J5 J10 — PHPS — J44 J8 — J40 J42 J35 J36

J6 J5 J1 — SRPS — J3 — J51 J52 J46 J48 J45 J50 J47

J14 J19 J17 J11 J20 J12 Driver Module J2 J7 J1 J6 J25 J24 J23 J21 J22 — J3 J9 J5 J6

J94 J92 J93 J96 J95 DD Filter J76 J75 J74 J72 J73

J39 J54 J41 J40 J42 J35 J36 J51 J52 J46 J48 J45 J50 J47 J29 J32 J30 J34 F11 F10 F9 F8 F7 F6 F5

M3302 M3314

Temp/Hum Sensor

Scan Room Blwr Box — Patient Comfort — Hub

Mag Mon — J5, J10, J8 J9, J7

JR Power **Cryo-Cooler Compressor** JC — Water Return, Water Supply, Gas Return Gas Supply

HEC — HEC Power Box — Cryo, Power Supply, GC, PE Blower

Signal Box — J24, J11 J12

Grad HtXchr

Grad Res

Grad Pump

Flow Meter — Filter

Blower HtXchgr

PGR Pump

PGR Res

PGR HtXchr

Grad/RF Coil Blower

SPW

J115, J116, J117, J106, J16

J118 J17

J100 Wvgde, J101 Wvgde, J111 Wvgde, J110 Wvgde, Patient Alert Wvgde

Primary MRU

E-Off Switch

O_2 Monitor — Sensor, Alarm, Light

Term Conn — F1, F2, F3, F4

X Y Z Gradient Filter X Y Z — J128

J121 J120 Xmit Switch J123 J122

J104 Wvgde

J105 Wvgde

P2026

Mag Enclosure Chimney

Gradient Bus Bar — X, Y, Z

Mag I/O — J1, J3, J6, J4, J2, J5, J12, J15, J14, J11, J13, J8, J7, J18, J19, J9, J10, J17, J16

FO I/F — P2026 — J28, J20, J21

Mag Bore

J8 J9 J6 J7 — Body Coil J2 J3 J9

Bore-PAL L R

Left Arc Cover — Ambient Temp Snsr

Bore Temp Sensors

Front Mic

Airflow Sensors

X Y Z XRMB Gradient Temp Sensor

J101 MAG-Grad Coil, J100 MAG-Grad Coil

J104 Grad Coil Air

Patient Air

Operator Workspace

Temp/Hum Sensor

P2026

Server (Option)

Service Laptop

Optional Speaker

GOC

J5 J4 J3 — IRDC (Option) — J2 J1

USB-Serial Converter

Ethernet 1 2 Switch

GOCAA — J12 J14 J17, J18, J15 J7, J6

Host PC — Ethernet B, Lwr Graphic Card Port 1, USB, RS232, USB, USB, USB, Upr Graphic Card Port 1, Ethernet A, Line In, Line Out

J10 J4 — J6 J2 GOC PDU — J9 J5 J1 J3

SCIM Spkr, Mic

Global Modem (Option)

Key-board

Mouse

Primary Display

Pneumatic Patient Alert System

Scan Room

Magnet

J3 P403 Terminal Box

Pressure Transducer

Reed Switch

J2 RUO PreAmp J1

Coldhead Pwr Sleeve — Coldhead Supply Return

Airflow Sensor Module

J9/J10 J2 PAC — M3336

Pre-Amp — Output Input

J1 J4 J2 J3 J6 BHYB

LCP J11 J3

RCP J11 J3

M3337

SRI4 — J23 J10, J34 J3 J20 J32 J24, J1 J4, J5, J33 J11/J12 J13/J14 J21 J31 J2 J22 J17

IRD J1 — P2026 — M3366

J11 J2 J12 TDM J1 J9 J10

J9 RRx1 J7 J1 J4/J5 J8 J2

J7 RRx2 (32-ch) J1 J4/J5 J2

MAG-PAC J4

PAC I/F

Squeeze Bulb

Speaker

Left E-Stop Button

Rt E-Stop Button

Patient Table

Coil Port P3

Coil Port P4

Touch N Go_R

Touch N Go_L

Table Hydraulics

Rear Pedestal

PED Antenna

RF Hub — Hub Backplane (bus, power, status/control) — J102J101 — J104 J103

J1 J3 J2 J4 PDB J5 J6

MCDB

J9 J13 J12 J16 J17 RFCB J10 J8 J14/J15 J1 J3 J7J6 J2 J4 — Debug Port

J1–J6 (ch 17-32)

RFSB (ch 1-16) J1–J6

Lights

J4 J6 J3 TIF J1 J2 J5

POS Enc

Long Drive J1 ASM (MTR) J2

Load A

J2 PED I/O J1

J3 EOT Sensor, HOME Sensor, EGRESS Sensor LPCA

J4 J2 Port A J5 J1

Port P2 J4

Port P1 J5

Rear Mic

Reroute Box — J10 J7, P2A P3A, P2B P3B, P1A P4A, P1B P4B, J8 J9

Load B

Rear Pedestal Air

Dock

J2 Mag, J1 Dock, J3 Left

ODU

Key

TBD = To Be Determined (Future Enhancement)

Cabinet

Chassis/ Sub-system

Part

Data Cables, Coolant, Power, Pulse, Data Paths

E-Stop/E-Off Circuitry

E-Stop Circuit

E-Off Circuit

PDU Power Off

© 2009, GE Healthcare