

De-Installation for MR Upgrades

Upgrade Documentation (each has De-Install Info and Instructions where applicable) The Latest Update Can Be found On Line Indicated

•Direction 2143793, *Signa Horizon System Upgrades*

This manual upgrades existing pre-LCC magnet systems (Advantage 4.x & 5.x, Horizon 5.x, & LX) to 9.x or 10.x. The magnet is not upgraded. <http://3.87.118.28/optec3/common/8xdocs/install/2143793/upgrade.htm>

•Direction 2223753, *Signa MR/i & CV/i System Upgrades*

This manual is used to upgrade early LCC magnets that were originally installed with the old "box" enclosure to the new WideOpen style enclosure <http://3.87.118.28/optec3/common/8xdocs/install/2223753/signa.htm>

•Direction 2248332, *Magnet & System Deinstallation for Upgrade to Signa WideOpen*

This manual is being obsoleted. The deinstallation information is being incorporated into Direction 2248333 <http://3.87.118.28/optec3/common/8xdocs/install/2248332.pdf>

•Direction 2248333, *Magnet & System Upgrade to Signa WideOpen*

This manual contains information for deinstalling a magnet and system that is being upgraded to an LCC magnet with a WideOpen enclosure. It also contains instructions for upgrading to a 9.x or 10.x (MGD) system. It must be on site before deinstallation of a magnet and system begins. <http://3.87.118.28/optec3/common/8xdocs/install/2248333/magnet.htm>

•Direction 2304881, *Signa 1.5T TwinSpeed System Upgrade Installation*

This manual contains instructions for upgrading an existing LCC magnet and system to TwinSpeed. It also contains instructions for a site that had a pre-LCC magnet removed (see Direction 2316677) and is now installing a TwinSpeed system. The manual also has information for upgrading to EXCITE Technology (MGD).

http://3.87.118.28/optec3/common/1_mgd/install/2304881/upgrade.htm

•Direction 2316677, *Magnet & System Deinstallation for Upgrade to Signa 1.5T TwinSpeed*

This manual removes the magnet and obsolete system components for upgrade to TwinSpeed. This manual is not part of any catalog structure. It must be on site before deinstallation begins.

http://3.87.118.28/optec3/common/1_mgd/install/2316677/2316677.pdf

•Direction 2334999, *Signa System Upgrade for MGD*

This manual is used to upgrade any existing LX system to one with EXCITE Technology. It contains information of removing cables, cabinets, and modules. If a site is having a magnet removed and the system is being upgraded to one with MGD, it would be good to have this manual on site prior to deinstallation so the applicable equipment can be removed at that time.

http://3.87.118.28/optec3/common/1_mgd/install/2333499/2333499.pdf