

FCT SERVICE NOTE

SN07010810a

Class M information is Property of GE - For GE Service Personnel Only. Disclosure to Third Parties Prohibited.

Date: January 8, 2007

Subject: UPDATE: gesyslog File Viewer version 2.1 is now available.

Effectivity: CT and PET systems

Details: The gesyslog File Viewer tool has been updated:

- Improved design includes:
 - o Better file management. Drop Down selection of gesyslog files of interest.
 - o Display of entire file contents upon startup/selection to re-display file contents.
 - o Expanded display window.
 - o Improved Help utility.
 - Better definition of color codes and their patterns.
 - Added detailed procedure that describes how to use Exceed FTP Host Explorer to retrieve gesyslog files from a scanner.
 - o Cut & Paste of specific messages.
 - o PSDB one button access. (SSO, GE Intranet/VPN required- Windows Version Only).

This tool can be loaded on any PC/Laptop, can be run directly from a CD media disk, and will soon be available directly from the Common Service Desktop under "Error Logs". (Pathfinder Release 2.0). It will continue to be available as a Multi-O/S CD-ROM and can be downloaded at your convenience. It will remain available for those CT and PET systems that have not yet been, or will not be included in the Pathfinder 2.0 software update distribution in 2007.

HOW TO GET VERSION 2.1:

This tool will not be delivered or published on a CD to all Field Service personnel. It is ONLY available as a Download .iso file at:

http://www.ct.med.ge.com/tech_pubs/fct/system/fct_home.htm

Scroll to the bottom of the Service Engineering Website:

1. Select: *Downloads*
2. Read the Proprietary warning statement, then Click on the blue lettering starting with:
"I certify that I am a GE Healthcare Employee...."
3. Select and download the file: **GFV_2.1M.zip** (Zip file size is 5,054KB) to your computer. (This zip file includes instructions on how to burn an .iso file to a blank CD media using Sonic RecordNow! Or Nero 6.0.)
4. Once downloaded unzip the file into a new (temp) folder.
5. Read the Readme.txt file for update information.

This software tool uses specifically, Sun Java 1.4.2 or greater. To check if your PC/Laptop has the correct version of Java:

1. Open Internet Explorer
2. Under "Tools" Select "Internet Options" tab.
3. Select "Advanced" tab
4. Look for version of Sun Java located right above Microsoft VM.
5. Check the box.

6. Reboot.

If you do not have Sun Java version 1.4.2 or greater installed and DO NOT use the VOLC tools:

1. Visit www.sun.com at: <http://java.sun.com/j2se/1.4.2/download.html>
2. download and install J2SE JRE (J2SEv1.4.2_12 JRE (1.4.2_12 or greater)

If you do not have Sun Java version 1.4.2 or greater installed and USE the VOLC tools:

1. Visit www.sun.com
2. From the main Window select Downloads.
3. Scroll down and download Java Runtime Environment (JRE) 5.0 Update 10 (Or greater).

INSTRUCTIONS FOR USE ON A SERVICE LAPTOP/PC:

If you do not have a previous version of gesyslog File Viewer installed:

1. Proceed to "Installation".

Save any log files you may have collected:

1. Using Windows Explorer:
 - a. Create a temporary backup folder.
 - b. Locate Folder c:\GFV\Log
 - c. COPY any gesyslog files you intend on saving into the backup folder. (No need to copy gesyslog_ct.log or gesys_PET.log files. These will be re-installed with version 2.1.)

Uninstall the previous version of the MI&CT gesyslog File viewer:

1. From Windows Select→ Start→Settings Control Panel→Add/Remove Programs→Locate MI&CT gesyslog File Viewer and select REMOVE.
2. Using Windows Explore, check to make sure that folder GFV has been removed:
 - d. If folder "GFV" exists DELETE it. (All contents and the GFV folder will be recreated upon install).

Installation:

You may install and run GFV two ways:

- Installing the software directly to your Windows based PC/Laptop from the temp folder.
- From aCD-ROM media disk you will create.
If you intend on using the software directly on the scanner you will need to burn a CD media disk. The enclosed .ISO file when burned to a CD creates a "Multi-OS" disk media that can run from your Service Laptop OR can be mounted and run directly from the SCSI tower on the CT or PET system.

Chose which way is logical for your use.

1. To install directly from the temp folder created:

- a. Open the temp folder, locate and execute the "SETUP.EXE" file.

2. To install from the Multi-OS CDROM media disk:

- a. Burn the "iso" file to a blank CD-ROM media disk. (If necessary) read the text file included for instructions on how to create a CD media disk from an .iso file.
- b. Insert the media disk you've created into your PC/Laptop. Autorun will take up to 20 seconds to begin the installation. (Note: If autorun is NOT enabled on your system Select Start→Run→ Browse→enter path to the CD Media (Drive E:) → Select "SETUP.EXE").

3. The software will load directly to your hard drive and become available off of your Start Menu.

4. Remove the software CD media disk and mark it "GFV Version 2.1" Multi-OS.

Running the Software From your PC/Laptop:

1. Goto → Start → Programs → MI&CT gesyslog File Viewer → MI&CT gesyslog File Viewer.
2. If this is the first time you are using this tool select "Help" and follow the Demonstration/Tutorial. (It takes about 5 minutes to complete).

Running the software directly on a PET or CT system using the media disk:

This software does not load on the system. It runs completely from the CD media disk you created. To properly run this software requires the following steps:

1. Insert CD media into DVD/CD player in SCSI Tower.
2. Open a Linux Shell.
3. Type: **su -**
4. Password: **#bigguy**
5. Type: **mount /mnt/cdrom1**
6. Type: **/mnt/cdrom1**
7. Type: **./StartGFV**
8. If you have never used this tool, it is highly recommended to perform the "Demonstration/Tutorial" by selecting the Help button in the tool. (It takes about 5 minutes to complete).
9. When you exit the software you must dismount the media disk. To dismount it:
 1. Open Linux Shell.
 2. Type: **su -**
 3. Password: **#bigguy**
 4. Type: **cd /mnt**
 5. Type: **umount /mnt/cdrom1** (Note: NOT **unmount** **u**mount)
 6. The CDROM reader green LED will blink a few times.
 7. Press the eject button on the front of the reader to eject the media disk.

Running the software on systems with Pathfinder 2.0 FMI installed, (2007):

1. Open the Service Desktop and from the Common Service Desktop:
2. Select: **CT** (PET systems only)
3. Select: **Error Logs**
4. Select: **StartGFV** from the menu. Note: If StartGFV selection does not exist as an option, then you must run the software from the media disk you created.

Presently, This tool is designed for use on CT and PET systems only: Updates will be announced periodically. You should check the Download Section on the MI&CT Service Website regularly for Service Documentation updates and Service Tool Downloads.

If you have problems running the software:

Locate the Readme.txt file on the CD media and read it for further instructions, OR point your browser to the "Help.htm or linuxhelp.htm" file found in folder/directory GFV. Refer to "Issues/Error messages" link.

Douglas R. Hofstetter
MI&CT Service Engineering

Basic Operation/Overview

This software tool uses a color-coding scheme to provide and describe all the messages involved with 4 key failure events. (Scan Aborts, Tube Spits, Resets, Shutdowns).

1. Select gesyslog file of interest from drop down list, your search type, then select the "Search" Button. (Perform the tutorial found in the Help feature for complete operation).
2. The tool detects key failure event messages and colorizes them. It does this by "Bracketing" the normal start or end of an exam, or fastcal (Green) or Start of exam or fastcal (Green) and the presence of one of these four events (Red).
 - a. Dark Blue (Pri) and Light Blue (Sec) messages fall between two green messages or between a Green Message and a Red Message. The Red message terminated the Exam.
 - b. If the message is marked PRI/Abort. This specific message caused the abort.
3. If the Red message is a reset, shutdown or scan abort, the Dark Blue (Pri) errors are the most likely cause of the event. Any Light Blue (Sec) messages may be associated with the PRIMARY message that caused the exam to end.
4. This tool counts Tube Spit **messages** NOT actual tube spits. The presence of tube spit message in the log file does not always cause a reset, shutdown or scan abort so they would normally be colored Dark Blue (Pri) and Light Blue (Sec) depending on severity, and may be bracketed between two green messages. If too may Tube Spit messages appear in a short amount of time in the log, this condition will be bracketed by a green start of exam or fastcal and a red reset, Shutdown or Scan Abort message. You must still make a judgment on the cause. Your system knowledge and training is still needed for an effective repair. The group of errors detected and colorized are meant to provide you with a quick analysis of the log file to make a better diagnosis for your pursuit of a repair.

