

MI&CT ESERVICE NOTE SN07090608a

Class M information is Property of GE. Disclosure to Third Parties prohibited.

This short-term informational note is invalid 6 months after the issue date.

Date: September 6, 2007

Subject: FCT Training Supplied "Z align calculator" and "Torque Values" Torque settings are incorrect.

Effectivity: LightSpeed, Brightspeed, RT and CT/i

Details: Excel spreadsheets containing torque value information for CT systems have been disseminated in training courses over the years. Some of the torque values listed in these spreadsheets have been found to be inaccurate or outdated.

Service on GE Healthcare equipment should always be based on information contained in the latest revision of Service Methods Documentation. Never depend solely on information disseminated through technical training courses.

Remember the following disclaimer that applies to all training documentation and materials:
For training purposes only.
This information may not be appropriate for your system.
Always use service documentation specific to the system you are servicing.

Resolution: If you have copies of either of the Excel spreadsheets titled "Torque Values" or "Z align calculator all systems" that were disseminated in CT Technical Training courses, please delete these files, and do not use the torque values listed therein for any service on systems.

PQR #: 13114478

Joseph L Getchel
FCT Technical Training Instructor

MI&CT Engineering Manager Reviewer: __Mark Nyman__