

Correlated Noise Data Sheet

1.5T System Correlated Noise

Image Data								Acceptance Guidelines		
Image R1=11	System Frequency	Mean (X)	STD DEV (SD)	Hot Pixel (P/)	Hot Pixel Location (L, P)	Zipper Peak Pixel (Pz)	Zipper Mean (Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
								$P \leq [(X) + 8(SD)]$	$(X) - 6(SD) \leq Pz \leq [(X) + 6(SD)]$ $[(X) - 3] \leq Xz \leq [(X) + 3]$	
1	63,764,000									
2	63,789,000									
3	63,814,000									
4	63,839,000									
5	63,864,000									
6	63,889,000									
7	63,914,000									
8	63,939,000									
9	63,964,000									
R1=7	Frequency	(X)	(SD)	(P/)	(L, P)	(Pz)	(Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
10	63,764,000									
11	63,789,000									
12	63,814,000									
13	63,839,000									
14	63,864,000									
15	63,889,000									
16	63,914,000									
17	63,939,000									
18	63,964,000									
R1	Freq (note)	(X)	(SD)	(P/)	(L, P)	(Pz)	(Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
11										
7										

Note: System frequency is set to magnet frequency for the matrix scans.

Correlated Noise Data Sheet

3.0T System Correlated Noise

Image Data								Acceptance Guidelines		
Image R1=11	System Frequency	Mean (X)	STD DEV (SD)	Hot Pixel (P/)	Hot Pixel Location (L, P)	Zipper Peak Pixel (Pz)	Zipper Mean (Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
								$P / \leq [(X) + 8(SD)]$	$(X) - 6(SD) \leq Pz \leq [(X) + 6(SD)]$ $[(X) - 3] \leq Xz \leq [(X) + 3]$	
1	127,653,000									
2	127,678,000									
3	127,703,000									
4	127,728,000									
5	127,753,000									
6	127,778,000									
7	127,803,000									
8	127,828,000									
9	127,853,000									
R1=7	Frequency	(X)	(SD)	(P/)	(L, P)	(Pz)	(Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
10	127,653,000									
11	127,678,000									
12	127,703,000									
13	127,728,000									
14	127,753,000									
15	127,778,000									
16	127,803,000									
17	127,828,000									
18	127,853,000									
R1	Freq (note)	(X)	(SD)	(P/)	(L, P)	(Pz)	(Xz)	Pixels (P/)	Zippers (Pz) or (Xz)	
11										
7										

Note: System frequency is set to magnet frequency for the matrix scans.