Patient Handling – Scan Room Sensors

Equipment Room

Main Disconnect Panel — CB3, TB1-15-16, CB2

Remote MRU

O_2 Monitor Control Unit — Meter, Ctrl Board, Alarm, Light

Customer-Supplied Outlet

Site E-Net

Facility Power Filters

Facility Utilities — Facility Coolant Supply, Facility Coolant Return

E-Off Switch

PGR

PDM L1/L2/L3 — J29 J25, J27 J28 J24

VRE 2 (32-ch) — J5, Net 3, Net 2, Net Mgmt, J1

VRE 1 — J1, Net Mgmt, Net 2, Net 3, J5

ETH 1 2 3 4 Term Server

Aux Board — J4, J6, J1 J2/J3 J9 J11 J10

XRFD LS1-A LS1-B LS2-A LS2-B LS3-A LS3-B LS4-A Term — PDU Top 1, PDU Top 1, PDU Front

XGD Fan x3 — J4

CB1 J1 J4 J3 — CB3 — RECON — RF/MNS J5 J2 PEN — PGR PDU — Host — J6/J7/J8 J11/J12/J13

CB2 J1 J2 — Service Panel — J21 J22 J11 J1 HSS Test

NB Amp — J4 J3, J14 J13 J10 J9, J8 J7 J5 J6 J19 J18, J3 J2

16 Port E-net Sw — 3 2 1 4 5 6 15 12 11, 14 13 7

J14 J15 J13 J12 J4 J2 J102 J19 J18 J17 J3 J6

Cronus Chassis

XPS CB — J13/J16, J12/J15, J11/J14

XGA CB — J13/J16, J12/J15, J11/J14

Cronus Backplane (Gold Fingers, Mezzanine only)

J10/J11 J6 J9 J8 J7 — GMC — J1 J2

P1 J3 J4 J5

Pwr XPS x3 — J2 J10, J1 J11, J8, J9 J3, J4/J5, J4/J5

XGA x3 — J8 J7 J16, J13 J15, J14 J5, J12 J6 J17 J18, J1/J2, J1/J2

Manifold Supply

SMA Coax 9-pin — LEM Icoil x3 — Grad Pulse In, Grad Pulse Out

Term Block x3

Manifold Return

J2 — CAM Chassis — ETH — SCP3 PMC — 10/100 Base — AGP2 — J3

UPM1 — Det J1 J2, Proc J3

UPM2 — Det J1 J2, Proc J3

J8 — J3 NB AIF J8 — STIF J4 J3 J18/J19

J23 I/O Panel J10 — IRF3 J8 — J14 J13 J11

J20 J22 J21 J11 J1 J5 J31 J16 J103

Door Switch

Coolant Removal Kit

PEN

J14 J1 J13 J11 J12 J10 — F/O Wvgde

J19 J15 J18 J17 J3 J9 J5 J6

J18 J7 — J23 J12 J22 — Exciter — J19 J20 J21 J29 J32 J30 — J15 J17 J1

J1 — Power Distribution — J4 J2 J3

J14 J19 J17 — J11 J12 J2 J7 J6 — Driver Module — J20 J1 — J25 J24 J23 J21 J22

J18 J19 J7/J8 J1/J2 J5/J6 J3/J4 — CFB — J17

J12 J9 J37/J38 — J2 J43 J39 J54 J41 J40 J42 J35 J36 — PHPS — J44 J8

J6 J5 J1 — J51 J52 J46 J48 J45 J50 J47 — SRPS — J3

J94 J96 J92 J95 J93 J76 J75 — DD Filter — J74 J72 J73

J39 J54 J41 J40 J42 J35 J36 J51 J52 J46 J48 J45 J50 J47 J29 J32 J30 J34 F11 F10 F9 F8 F7 F6 F5

Temp/Hum Sensor

Scan Room Blwr Box — Patient Comfort — Hub

Mag Mon — J5 J10 J8 J9 J7

HEC — HEC Power Box — Cryo, Power Supply, GC, PE, Blower — Signal Box — J24, J11 J12 — Grad HtXchr, Grad Res, Grad Pump, Blower HtXchgr, Flow Meter, Filter

JR Power **Cryo-Cooler Compressor** JC — Water Return, Water Supply, Gas Return, Gas Supply

PGR Pump, PGR Res, PGR HtXchr, Grad/RF Coil Blower

Operator Workspace

Temp/Hum Sensor

Server (Option)

Service Laptop

Optional Speaker

GOC

IRDC (Option) — J5, J4 J3, J2 J1

USB-Serial Converter

Ethernet 1 2 Switch

GOCAA — J12 J14 J17, J18, J15 J7, J6

Host PC — Ethernet B, Lwr Graphic Card Port 1, USB, RS232, USB, USB, USB, Upr Graphic Card Port 1, Ethernet A, Line In, Line Out

J10 J4 J6 J9 J2 GOC J5 J1 PDU J3

Global Modem (Option)

Key-board

Mouse

Primary Display

Pneumatic Patient Alert System

SCIM Spkr, Mic

Scan Room

SPW

J118 J17 — J100 Wvgde, J101 Wvgde, J111 Wvgde, J110 Wvgde, Patient Alert Wvgde

J115 J116 J117 J106 J16

Term Conn — F1 F2 F3 F4

X Y Z **Gradient Filter** X Y Z — J128

J121 J120 **Xmit Switch** J123 J122

J104 Wvgde

J105 Wvgde

Primary MRU

E-Off Switch

O_2 Monitor — Sensor, Alarm, Light

Mag Enclosure Chimney

Gradient Bus Bar — X Y Z

Mag I/O — J1 J3 J6, J4, J2 J5 J12 J15 J14 J11, J13 J8 J7 J18 J19 J9 J10 J17 J16

FO I/F — J28, J20 J21

Mag Bore

J8 J9 J6 J7 — Body Coil — J2 J3 J9

Bore-PAL L R

Left Arc Cover — Ambient Temp Snsr

Bore Temp Sensors

Front Mic

Airflow Sensors

X Y Z **XRMB Gradient** Temp Sensor

J101 MAG-Grad Coil, J100 MAG-Grad Coil

J104 Grad Coil Air

Patient Air

Magnet

J3 P403 Terminal Box

Pressure Transducer

Reed Switch

J2 RUO PreAmp J1

Coldhead Pwr Sleeve — Coldhead Supply Return

Airflow Sensor Module

J9/J10 J2 **PAC**

M3350 — M3344

SRI4 — J23 J10, J34 J3 J20 J32 J24, J1 J4, J5, J33 J11/J12 J13/J14 J21 J31 J2 J22 J17

Pre-Amp — Output Input

J1 J4 J2 J3 J6 **BHYB**

J11 J12 J2 J1 J9 J7 J10 **TDM**

J9 **RRx1** J7 J1 J4/J5 J8 J2

J7 **RRx2** (32-ch) J1 J4/J5 J2

MAG-PAC J4

PAC I/F

Squeeze Bulb

Speaker

Left E-Stop Button

Rt E-Stop Button

Patient Table

Coil Port P3

Coil Port P4

Touch N Go_R

Touch N Go_L

Table Hydraulics

Rear Pedestal

PED Antenna

Lights

RF Hub — J102 J101, J104 J103 — Hub Backplane (bias, power, status/control)

J1 J3 J2 J4 J5 J6 PDB

MCDB

J9 J13 J12 J16 J17 J10 J14/J15 J8 J1 J3 J7 J6 J2 J4 RFCB — Debug Port

J1–J6 (ch 17-32) RFSB (ch 1-16) J1–J6

J4 J6 J3 TIF J1 J2 J5

Long Drive J1 ASM (MTR) J2

POS Enc

J3 EOT Sensor, HOME Sensor, EGRESS Sensor LPCA

Load A

J4 J2 Port A J5 J1 — Port P2 J4, Port P1 J5 — Rear Mic

J2 PED I/O J1

J10 J7 P2A P2B P1A P1B **Reroute Box** P3A P3B P4A P4B J8 J9

Load B

Rear Pedestal Air

Dock

J2 Mag, J1 Dock, J3 Left

ODU

Key

TBD = To Be Determined (Future Enhancement)

Cabinet | Chassis/ Sub-system | Part

Data Cables, Coolant, Power, Pulse, Data Paths

© 2009, GE Healthcare