

3.0 T Coil Compatibility Matrix

					Platform Compatibility							Notes / Application Compatibility	# of elements	# of channels
Catalog	GE P/N on Label	Coil Description	Supplier	Connector Type	3T94 EXCITE	3T94 HD	3T94 HDx	G3 EXCITE	G3 HD	G3 HDx	MR750			
Head and NV	-	-	-	-	-	-	-	-	-	-	-	-	-	-
M3335LC	HNU: 5325974 NCU: 5325953-4 TLU: 5328289	HDx Head Neck and Spine (forward production systems)	Medrad	2 HD (1 16 ch and 1 8 ch)	no	no	no	no	no	HDXt only	no	Mcat currently not active - use F7002MR. Compatible with 16 ch systems only. Asset, Bravo, Propeller. T2 Map available.	29	16
M3335LD	HNU: 5325957 NCU: 5325953-4 TLU: 5328289	HDx Head Neck and Spine (upgrade systems)	Medrad	2 HD and 1 Legacy	no	no	no	no	no	HDXt only	no	Mcat currently not active - use F7003MR. Compatible with 16 ch systems only. Asset, Bravo, Propeller. T2 Map available.	29	16
M7000AP	HNU: 5339626 NCU: 5338938 TLU: 5336485	16-ch Head Neck and Spine (forward production systems, compatible with 16 ch or 32 ch systems)	GE Coils	1 HD and 1 P-connector	no	no	no	no	no	no	yes			
M3335HC	HNU: 5339626 NCU: 5338938 TLU: 5336485	16-ch Head Neck and Spine (forward production systems, compatible with 16 ch only)	GE Coils	2 HD (1 16 ch and 1 8 ch)	no	no	no	no	no	yes	no			
M3335HD	HNU: 5339626 NCU: 5338938 TLU: 5336485	16-ch Head Neck and Spine (upgrade systems, compatible with 16 ch only)	GE Coils	2 HD and 1 Legacy	no	no	yes-upgrades only	no	no	yes-upgrades only	no			
M3334TP	2414390	HD 8 Ch NV Array	GE Coils	HD	no	yes	yes	no	yes	yes	yes	Asset, Bravo, Propeller. T2 Map available.		
M1385AT	2414390	8 Ch NV Array	GE Coils	Legacy	yes	yes*	yes*-upgrades only	yes	yes*	yes*-upgrades only	no	* will work on HD and HDx upgrades but should order M3334TP instead. Asset, Bravo, Propeller. T2 Map available.		
M1385AN	2383478-2	4 Ch NV Array	GE Coils	Legacy	no	no	no	yes	yes	yes-upgrades only	no	T2 Map available.		
M1385AZ	2383477-2	4 Ch NV Array	GE Coils	Legacy	yes	yes	yes-upgrades only	no	no	no	no			
M3335LL	5147134-2	HD 8 ch HiRes Brain	Invivo	HD	no	no	yes	no	no	yes	yes	Asset, Bravo, Propeller. T2 Map available.		
M1385AS	2380637-2	8 Ch.Brain Array	Invivo	Legacy	yes	yes	yes-upgrades only	yes	yes	yes-upgrades only	no	Asset, Bravo, Propeller. T2 Map available.		
M3335LM	2415542	HD 8 ch CTL Spine Array	GE Coils	HD	no	no	yes	no	no	yes	yes	T2 Map available.	12	8
M1385AW	U1-100407 (USAi P/N) 2381682-2	8 Ch. CTL Spine Array	GE Coils	Legacy	yes	yes	yes-upgrades only	yes	yes	yes-upgrades only	no	T2 Map available.	12	8

Catalog	GE P/N on Label	Coil Description	Supplier	Connector Type	3T94 EXCITE	3T94 HD	3T94 HDx	G3 EXCITE	G3 HD	G3 HDx	MR750	Notes / Application Compatibility	# of elements	# of channels
M1385AB	2380634-2	4 Ch.Spine Array	GE Coils	Legacy	yes	yes	yes-upgrades only	no	no	no	no			
E8800DB	#N/A (E-cat coils do not have GE P/N's)	32-ch Head Coil	MR Instruments		no	no	no	no	no	no	yes			
Extremity	-	-	-	-	-	-	-	-	-	-	-	-	-	-
M3335LB	5160254-2	HD 8 Ch T/R Knee coil	Invivo	HD	no	yes*	yes	no	yes*	yes	yes	* HD compatible version available H2'06. Asset. T2 Map validated.	8	8
M3335LP	5147137-2	HD Quad Knee/Foot	Invivo	HD	no	no	yes	no	no	yes	yes	T2 Map validated		
M1385AM	2380635-2	Knee/Foot Extremity T/R	Invivo	Legacy	no	no	no	yes	yes	yes-upgrades only	no	T2 Map validated		
M1385AC	2357819-2	Knee/Foot Extremity T/R	Invivo	Legacy	yes	yes	yes-upgrades only	no	no	no	no	T2 Map validated		
M3334TR	2404877	HD 3 Ch Shoulder Coil	GE Coils	Legacy	yes	yes	yes-upgrades only	yes	yes	yes-upgrades only	no	If ordering shoulder coil new for HDx, then order M3335LR with HD connector. T2 Map available	3	3
M3335LR	2414331	HD 3 ChShoulder	GE Coils	HD	no	no	yes	no	no	yes	yes	T2 Map available	3	3
E8800R	#N/A (E-cat coils do not have GE P/N's)	Wrist	Mayo	HD	no	no	no	no	no	yes	yes		quadrature	single
E8800RB	#N/A (E-cat coils do not have GE P/N's)	Wrist (G3 Excite/HD)	Mayo	Legacy	no	no	no	yes	yes	yes-upgrades only	no	T2 Map available. Replaces NW2003AA.	quadrature	single
E8800RC	#N/A (E-cat coils do not have GE P/N's)	Wrist (3T94 Excite/HD)	Mayo	Legacy	yes	yes	yes-upgrades only	no	no	no	no	T2 Map available. Replaces NW2003A.	quadrature	single
E8800RC	#N/A (E-cat coils do not have GE P/N's)	Wrist (3T94 Excite/HD)	Mayo	Legacy	yes	yes	yes-upgrades only	no	no	no		T2 Map available. Replaces NW2003A.	quadrature	single
E8800RC	#N/A (E-cat coils do not have GE P/N's)	Wrist (3T94 Excite/HD)	Mayo	Legacy	yes	yes	yes-upgrades only	no	no	no		T2 Map available. Replaces NW2003A.	quadrature	single
M3340CA	5268024-2	8 Ch Wrist	Invivo	HD	no	yes	yes	no	yes	yes	yes		8	8
M3340CB	5273235-2	Foot Ankle	Invivo	HD	no	yes	yes	no	yes	yes	yes		8	8
M3340CC	5271265-2	Shoulder	Neo	HD	no	yes	yes	no	yes	yes	yes		8	8
Body and Breast	-	-	-	-	-	-	-	-	-	-	-	-	-	-
M7000TC	Anterior: 5339708-2 Posterior: 5339708-3	32-Ch Torso Array Coil	GE Coils	P-Connector	no	no	no	no	no	no	yes			
M3335LA	2415544	HD 8 ch VIBRANT Breast Coil	GE Coils	HD	no	no	no	no	yes*	yes	yes	Asset, Vibrant, Vibrant XV. * HD compatible version available start of July 2006 with sagittal-only Vibrant. Not compatible with 3T94 systems (HD or HDx). T2 Map available.	8	8
E8800BB	#N/A (E-cat coils do not have GE P/N's)	HD 7 ch Breast Coil	Invivo	HD	no	no	no	no	no	yes	yes	Vibrant on sagittal only		
M3335LN	2415410	HD 8 ch Torso Array	GE Coils	HD	no	no	no	no	no	yes	yes	Asset, Lava, Lava XV. T2 Map available.	8	8

Catalog	GE P/N on Label	Coil Description	Supplier	Connector Type	3T94 EXCITE	3T94 HD	3T94 HDx	G3 EXCITE	G3 HD	G3 HDx	MR750	Notes / Application Compatibility	# of elements	# of channels
M1385AY	100401	8 Ch.Torso Array	GE Coils	Legacy	no	no	no	yes	yes	yes-upgrades only	no	Assett, Lava, Lava XV. T2 Map available	8	8
M1385AE	2366942	4 Ch.Torso Array	GE Coils	Legacy	yes	yes	yes-upgrades only	no	no	no	no	Asset. T2 Map available		
E8801RH	#N/A	Endorectal Interface Device	Medrad	HD	no	no	no	no	no	yes	yes	Compatible with M3335LN HD 8 ch Torso. HDx only. Performs same function as Auto Tuning Device.		
E8801RK	#N/A	Disposable Endorectal probe	Medrad	HD	no	no	no	no	no	yes	yes			
General	-	-	-	-	-	-	-	-	-	-	-	-	-	-
M1385AR	2373367-2	GP Flex	GEMS	Legacy	no	no	no	yes	yes	yes-upgrades only	no			
M1385AF	2373367	GP Flex	GEMS	Legacy	yes	yes	yes	no	no	yes-forward production must get this one	yes	Coil has BNC connectors that connect to HDx single channel coil adapter. HDx coil adapter is included in all forward production systems and when LPCA upgrade is ordered.		
M0004SS	5260688	6 Ch Flex	GE Coils	Legacy	yes	yes	yes-upgrades only				no			
M0005SS	5260689	6 Ch Flex	GE Coils	HD	no	yes	yes				no			
M0050SS	5229646	6 Ch Phased Array Flex	NeoCoil	P-Connector	no	no	no	no	no	no	yes			
Cardiac	-	-	-	-	-	-	-	-	-	-	-	-	-	-
M3334TS	2411986 U1-100422 (USAi P/N)	HD 8 Ch. Cardiac Array	GE Coils	HD	no	yes	yes	no	yes	yes	yes	Asset, MR Echo, Lava. T2 Map available.		
M7000LA	Anterior: 5366627-2 Posterior: 5366627-3	3T 32-Channel Cardiac Coil	Invivo	P-Connector	no	no	no	no	no	no	yes		32	32