

HLA

SCHEDULE F (HIGH USAGE SCANNER)

SITE:

DATE:
SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
GANTRY			
Verify heater operation (LED)	_____	_____	_____
Check Mylar window and laser ports	_____	_____	_____
Check fans/ clean filters	_____	_____	_____
Inspect belts for wear	_____	_____	_____
Verify lamps	_____	_____	_____
Clean tube radiator/ check fans	_____	_____	_____
Inspect for leaky tube wells	_____	_____	_____
Clean covers (including top cover)	_____	_____	_____
Vacuum gantry base	_____	_____	_____
Check axial and tilt pots	_____	_____	_____
SYSTEM			
General Cleaning	_____	_____	_____
Update site log	_____	_____	_____
Check PM manual revision	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: