

HLA

SCHEDULE K (HIGH USAGE SCANNER)

SITE:

DATE:

SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
GENERATOR			
Check generator lamps	_____	_____	_____
Ground checks (CT 232)	_____	_____	_____
Clean generator filter/check fans	_____	_____	_____
Check PDU fans/ clean filters	_____	_____	_____
Check REM lamps	_____	_____	_____
SRU			
Check for proper fan operation of all fans in SRU	_____	_____	_____
For older style ST-CP AP, clean AP blower filter	_____	_____	_____
Clean dirt/dust/debris from bottom of SRU cabinet	_____	_____	_____
SYSTEM			
Update Site Log	_____	_____	_____
Check PM manual revision	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: