

HLA

SCHEDULE S (Normal USAGE SCANNER)

SITE:

DATE:

SYSTEM ID:

| <u>Procedure</u>                       | <u>Inspect</u> | <u>Clean</u> | <u>Split</u> |
|----------------------------------------|----------------|--------------|--------------|
| <b>GANTRY</b>                          |                |              |              |
| Inspect end stops                      | _____          | _____        | _____        |
| Verify heater operation (LED)          | _____          | _____        | _____        |
| Check Mylar window and laser ports     | _____          | _____        | _____        |
| Check DAS power supply fans and filter | _____          | _____        | _____        |
| Check fans/ clean filters              | _____          | _____        | _____        |
| Inspect belts for wear                 | _____          | _____        | _____        |
| Verify lamps                           | _____          | _____        | _____        |
| Clean tube radiator/ check fans        | _____          | _____        | _____        |
| Inspect for leaky tube wells           | _____          | _____        | _____        |
| Clean covers (including top cover)     | _____          | _____        | _____        |
| Vacuum gantry base                     | _____          | _____        | _____        |
| Check axial and tilt pots              | _____          | _____        | _____        |
| Inspect take-up drums                  | _____          | _____        | _____        |
| Inspect take-up for debris             | _____          | _____        | _____        |
| Inspect take-up bearings               | _____          | _____        | _____        |
| Inspect collimator filter age          | _____          | _____        | _____        |
| <b>SYSTEM</b>                          |                |              |              |
| Update site log                        | _____          | _____        | _____        |
| Check PM Manual Revision               | _____          | _____        | _____        |

## Comments:

---



---



---



---



---



---



---

Field Engineer:

Id#:

Dispatch#: