

HLA

SCHEDULE T (Normal USAGE SCANNER)

SITE:

DATE:
SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
TABLE			
Check I/S and cradle release	_____	_____	_____
Check I/S and cradle latch	_____	_____	_____
Check head holder and associated equipment	_____	_____	_____
Inspect Tableside Switch Panel	_____	_____	_____
Inspect cradle for adhesive tape/verify operation	_____	_____	_____
Inspect encoder cable	_____	_____	_____
Inspect and lube elevation screws	_____	_____	_____
Check I/S, cradle, elevation pots/belt	_____	_____	_____
Check limit switches	_____	_____	_____
Inspect TGGC fans	_____	_____	_____
Vacuum internal table	_____	_____	_____
Clean table covers and cradle	_____	_____	_____
SYSTEM			
Verify emergency off buttons (table/console)	_____	_____	_____
Check operation of final scan abort control	_____	_____	_____
Check X-Ray on lights	_____	_____	_____
Check Pm manual revision	_____	_____	_____
Update site log	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: