

IMAGE DATA SHEET

QA Phantom #3

QA#3 MEANS and STANDARD DEVIATION

QA#3/S/10mm/120KV/1 70mA/2 sec .

Scan at 50mm. PERISTALTIC OFF

Exam # # of Slice	Mean	Std Dev	Average Std Dev Average of 4 slices	Comments/ Artifacts	Scan Tech Aprov
SPECS	0.0 ± 1.5	---	See below *	---	---

Box Size : 51 x 51 pixels

* If tube has < 5000 scans, spec is 3.00 to 3.60.

* If tube has > 5000 scans, spec is 3.00 to 3.80.

Image Acceptance Date: _____ Image Reviewer: _____

QA#3 MEANS and STANDARD DEVIATION

QA#3/S/10mm/120KV/1 70mA/2 sec .

Do not scan, evaluate images from previous step.

Exam # # of Slice	AvXc	AvXo	AvXo - AvXc		Comments/Artifacts	Scan Tech Aprov
SPECS	---	---	---	---	---	---

Box Size: 31 x 31 pixels

Coordinates: Center: 255, 255

Outside: 255, 94 416, 255

255, 416 94, 255

Image Acceptance/Date: _____

Certified Image Reviewer: _____

Artifact Limits:

Rings	4.0/4.8 cts	Center Spot	N/A	Band	2.8 cts
Clump	N/A	Streaks	4.0 cts	Band radius	0-8.5 cm
Smudge	2.2 cts	Center Artifact	3.5 Std Dev		
1 mm	4.0 cts				

Image Acceptance Date: _____ Image Reviewer: _____