

9800

SCHEDULE I (HIGH USAGE SCANNER)

SITE:

DATE:
SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
COMPUTER			
Check fans and filter (MDD/ODD)	_____	_____	_____
Clean argus filter	_____	_____	_____
Check argus bad block log	_____	_____	_____
Check ACE fan	_____	_____	_____
Check ACE bad block log	_____	_____	_____
Check fans/ clean filter in the bays	_____	_____	_____
Clean magnetic tape unit	_____	_____	_____
Check fans in magnetic tape unit	_____	_____	_____
Check GaiDs fans and GaiDs Power Supply	_____	_____	_____
Check AP fans/ clean filters	_____	_____	_____
CONSOLE			
Verify countertop screws	_____	_____	_____
Clean display monitor/Console exterior	_____	_____	_____
Check/ adjust vertical and horizontal sizing	_____	_____	_____
Check 'ease of use' of keys and trackball	_____	_____	_____
Check for screen burns	_____	_____	_____
GANTRY			
Check laser accuracy	_____	_____	_____
Check Mylar window and laser ports	_____	_____	_____
Check fans/ clean filters	_____	_____	_____
SYSTEM			
General cleaning	_____	_____	_____
Update site log	_____	_____	_____
Check PM Manual revision	_____	_____	_____
Xe Blood Flow			
Calibrate Xenon sensor	_____	_____	_____
Replace Xenon sensor	_____	_____	_____
Check condition of hoses/bag	_____	_____	_____
Check fan	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: