

HSA/HSA-RP

SCHEDULE S

Gantry Revolutions:

DATE:

SITE:

SYSTEM ID:

<u>Procedure</u>	<u>Inspect</u>	<u>Clean</u>	<u>Split</u>
CONSOLE			
Check display monitor for raster burns	_____	_____	_____
Check for smooth rolling trackball / Clean	_____	_____	_____
Clean / vacuum console interior	_____	_____	_____
COMPUTER			
Clean/Vacuum all exhaust vent holes	_____	_____	_____
Check that all fans work	_____	_____	_____
GANTRY			
Check the number of gantry revolutions	_____	_____	_____
Remove and measure POWER brushes	_____	_____	_____
Remove and measure SIGNAL brushes	_____	_____	_____
Grease main bearing	_____	_____	_____
Check drive belt for wear	_____	_____	_____
Check scan and laser windows	_____	_____	_____
Clean ventilation holes and check fans	_____	_____	_____
Handling and removal of slip ring brush debris	_____	_____	_____
Gantry Slip Ring maintenance	_____	_____	_____
SYSTEM			
Check Web for Latest PM Manual Revision	_____	_____	_____
TABLE			
Upper Assembly PM	_____	_____	_____
Lower Assembly PM	_____	_____	_____
Check for wear or damage on table extensions	_____	_____	_____

Comments:

Field Engineer:

Id#:

Dispatch#: